

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032521\
 Data File : BF123455.D
 Acq On : 25 Mar 2021 11:22
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 25 12:43:14 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 25 12:42:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	93	0.00
2	1,4-Dioxane	40.000	39.450	1.4	91	0.00
3	Pyridine	40.000	39.902	0.2	94	0.00
4	n-Nitrosodimethylamine	40.000	37.684	5.8	86	0.00
5 S	2-Fluorophenol	80.000	78.400	2.0	93	0.00
6	Aniline	40.000	39.522	1.2	92	0.00
7 S	Phenol-d6	80.000	77.744	2.8	92	0.00
8	2-Chlorophenol	40.000	38.722	3.2	92	0.00
9	Benzaldehyde	40.000	39.788	0.5	90	0.00
10 C	Phenol	40.000	42.103	-5.3	100	0.00
11	bis(2-Chloroethyl)ether	40.000	38.258	4.4	90	0.00
12	1,3-Dichlorobenzene	40.000	38.749	3.1	92	0.00
13 C	1,4-Dichlorobenzene	40.000	39.232	1.9	93	0.00
14	1,2-Dichlorobenzene	40.000	39.073	2.3	93	0.00
15	Benzyl Alcohol	40.000	38.331	4.2	87	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.860	10.4	83	0.00
17	2-Methylphenol	40.000	38.170	4.6	89	0.00
18	Hexachloroethane	40.000	38.136	4.7	87	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.196	9.5	84	0.00
20	3+4-Methylphenols	40.000	38.234	4.4	90	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	91	0.00
22	Acetophenone	40.000	37.776	5.6	90	0.00
23 S	Nitrobenzene-d5	80.000	78.967	1.3	87	0.00
24	Nitrobenzene	40.000	39.525	1.2	89	0.00
25	Isophorone	40.000	37.273	6.8	85	0.00
26 C	2-Nitrophenol	40.000	44.411	-11.0	91	0.00
27	2,4-Dimethylphenol	40.000	37.277	6.8	85	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.302	6.7	85	0.00
29 C	2,4-Dichlorophenol	40.000	39.234	1.9	89	0.00
30	1,2,4-Trichlorobenzene	40.000	39.214	2.0	90	0.00
31	Naphthalene	40.000	38.507	3.7	90	0.00
32	Benzoic acid	40.000	42.880	-7.2	93	0.00
33	4-Chloroaniline	40.000	39.127	2.2	90	0.00
34 C	Hexachlorobutadiene	40.000	37.645	5.9	86	0.00
35	Caprolactam	40.000	40.756	-1.9	89	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.157	2.1	88	0.00
37	2-Methylnaphthalene	40.000	37.506	6.2	88	0.00
38	1-Methylnaphthalene	40.000	38.663	3.3	90	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	88	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.758	5.6	88	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.694	5.8	82	0.00
42 S	2,4,6-Tribromophenol	80.000	80.502	-0.6	87	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.651	3.4	87	0.00
44	2,4,5-Trichlorophenol	40.000	39.842	0.4	89	0.00
45 S	2-Fluorobiphenyl	80.000	69.889	12.6	87	0.00
46	1,1'-Biphenyl	40.000	37.165	7.1	88	0.00
47	2-Chloronaphthalene	40.000	37.592	6.0	89	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032521\
 Data File : BF123455.D
 Acq On : 25 Mar 2021 11:22
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 25 12:43:14 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 25 12:42:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.498	-3.7	84	0.00
49	Acenaphthylene	40.000	37.354	6.6	87	0.00
50	Dimethylphthalate	40.000	36.621	8.4	83	0.00
51	2,6-Dinitrotoluene	40.000	39.169	2.1	83	0.00
52 C	Acenaphthene	40.000	38.157	4.6	89	0.00
53	3-Nitroaniline	40.000	40.357	-0.9	85	0.00
54 P	2,4-Dinitrophenol	40.000	40.701	-1.8	96	0.00
55	Dibenzofuran	40.000	37.452	6.4	87	0.00
56 P	4-Nitrophenol	40.000	41.623	-4.1	87	0.00
57	2,4-Dinitrotoluene	40.000	40.721	-1.8	83	0.00
58	Fluorene	40.000	36.774	8.1	87	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.149	-0.4	87	0.00
60	Diethylphthalate	40.000	36.578	8.6	81	0.00
61	4-Chlorophenyl-phenylether	40.000	36.137	9.7	84	0.00
62	4-Nitroaniline	40.000	41.772	-4.4	89	0.00
63	Azobenzene	40.000	36.019	10.0	82	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	87	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.019	-2.5	87	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.442	3.9	87	0.00
67	4-Bromophenyl-phenylether	40.000	38.630	3.4	85	0.00
68	Hexachlorobenzene	40.000	38.138	4.7	85	0.00
69	Atrazine	40.000	39.258	1.9	83	0.00
70 C	Pentachlorophenol	40.000	40.541	-1.4	81	0.00
71	Phenanthrene	40.000	38.772	3.1	87	0.00
72	Anthracene	40.000	38.304	4.2	86	0.00
73	Carbazole	40.000	39.195	2.0	86	0.00
74	Di-n-butylphthalate	40.000	36.038	9.9	77	0.00
75 C	Fluoranthene	40.000	38.765	3.1	85	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	84	0.00
77	Benzydine	40.000	45.452	-13.6	94	0.00
78	Pyrene	40.000	40.303	-0.8	86	0.00
79 S	Terphenyl-d14	80.000	76.835	4.0	82	0.00
80	Butylbenzylphthalate	40.000	40.196	-0.5	75	0.00
81	Benzo(a)anthracene	40.000	40.643	-1.6	86	0.00
82	3,3'-Dichlorobenzidine	40.000	39.938	0.2	78	0.00
83	Chrysene	40.000	38.491	3.8	81	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	36.683	8.3	71	0.00
85 c	Di-n-octyl phthalate	40.000	35.318	11.7	66	0.00
86 I	Perylene-d12	20.000	20.000	0.0	81	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.526	-3.8	91	0.00
88	Benzo(b)fluoranthene	40.000	38.725	3.2	75	0.00
89	Benzo(k)fluoranthene	40.000	39.530	1.2	80	0.00
90 C	Benzo(a)pyrene	40.000	39.662	0.8	80	0.00
91	Dibenzo(a,h)anthracene	40.000	40.055	-0.1	86	0.00
92	Benzo(g,h,i)perylene	40.000	42.034	-5.1	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032521\
Data File : BF123455.D
Acq On : 25 Mar 2021 11:22
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 25 12:43:14 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 25 12:42:42 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0