




Quant Time: Mar 25 11:19:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration



TIC: BF142067.D\data.ms

(70) Pentachlorophenol (C)




11.198min (-0.006) 96884.65 ng

response **208846**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.80	62.08
264.00	62.70	62.07
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032525\
Data File : BF142067.D
Acq On : 25 Mar 2025 09:39
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 25 11:19:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration



TIC: BF142067.D\data.ms

(77) Benzidine

12.727min (-0.006) 165543.78 ng

response 584341

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	14.16
183.00	12.10	12.03
0.00	0.00	0.00