


Data Path : Z:\HPCHEM1\BNA F\Data\BF032618\
 Data File : BF103928.D
 Acq On : 26 Mar 2018 16:45
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031518.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Mar 23 13:26:17 2018



AutoFind: Scans 1641, 1642, 1643; Background Corrected with Scan 1636

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.1	24955	PASS
68	69	0.00	2	1.8	473	PASS
69	198	0.00	100	34.1	26544	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	48.7	37863	PASS
197	198	0.00	2	0.7	567	PASS
198	198	100	100	100.0	77824	PASS
199	198	5	9	6.7	5184	PASS
275	198	10	60	27.0	21017	PASS
365	198	1	100	3.4	2612	PASS
441	443	0.01	100	82.1	11598	PASS
442	198	50	100	95.9	74624	PASS
443	442	15	24	18.9	14126	PASS