

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF032625\  
 Data File : BF142093.D  
 Acq On : 26 Mar 2025 13:07  
 Operator : RC/JU  
 Sample : PB167291BL  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 PB167291BL

Quant Time: Mar 26 13:41:17 2025  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF031025.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Mar 10 15:46:22 2025  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-----------------------------|--------|------|----------|---------|-------|----------|
| -----                       |        |      |          |         |       |          |
| Internal Standards          |        |      |          |         |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 6.869  | 152  | 128672   | 20.000  | ng    | -0.01    |
| 21) Naphthalene-d8          | 8.151  | 136  | 510860   | 20.000  | ng    | -0.01    |
| 39) Acenaphthene-d10        | 9.904  | 164  | 278564   | 20.000  | ng    | -0.01    |
| 64) Phenanthrene-d10        | 11.392 | 188  | 505504   | 20.000  | ng    | -0.01    |
| 76) Chrysene-d12            | 14.033 | 240  | 368173   | 20.000  | ng    | 0.00     |
| 86) Perylene-d12            | 15.504 | 264  | 300702   | 20.000  | ng    | 0.00     |
| System Monitoring Compounds |        |      |          |         |       |          |
| 5) 2-Fluorophenol           | 5.492  | 112  | 1062589  | 137.825 | ng    | 0.00     |
| 7) Phenol-d6                | 6.492  | 99   | 1271968  | 129.579 | ng    | -0.01    |
| 23) Nitrobenzene-d5         | 7.428  | 82   | 851094   | 93.756  | ng    | -0.02    |
| 42) 2,4,6-Tribromophenol    | 10.698 | 330  | 537502   | 152.097 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 9.228  | 172  | 1770589  | 96.665  | ng    | -0.01    |
| 79) Terphenyl-d14           | 12.980 | 244  | 2477065  | 99.470  | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF032625\  
Data File : BF142093.D  
Acq On : 26 Mar 2025 13:07  
Operator : RC/JU  
Sample : PB167291BL  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
PB167291BL

Quant Time: Mar 26 13:41:17 2025  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF031025.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Mar 10 15:46:22 2025  
Response via : Initial Calibration

