



#1
 1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 5.95 min Scan# 691
 Delta R.T. 0.00 min
 Lab File: BF093939.D
 Acq: 27 Mar 2017 12:54

Instrument :
 BNA_F
 ClientSampleId :
 E2-BB10-NSW

Tot Ion	Ratio	Lower	Upper
152	100		
150	157.9	126.2	189.2
115	58.1	52.2	78.2

