

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032717\
 Data File : BF093978.D
 Acq On : 28 Mar 2017 7:57
 Operator : SJ/MA
 Sample : I2387-01
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 032317-PRRT-E-D-M

Quant Time: Mar 28 08:18:09 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 27 16:10:49 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.94	152	166974	20.00	ng	-0.01
21) Naphthalene-d8	7.44	136	661929	20.00	ng	-0.02
38) Acenaphthene-d10	9.58	164	294180	20.00	ng	-0.02
63) Phenanthrene-d10	11.41	188	472911	20.00	ng	-0.01
75) Chrysene-d12	14.66	240	301492	20.00	ng	-0.02
86) Perylene-d12	16.29	264	276302	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	4.47	112	600827	60.78	ng	0.00
7) Phenol-d6	5.53	99	489249	40.00	ng	-0.02
23) Nitrobenzene-d5	6.59	82	1048387	90.39	ng	-0.02
41) 2,4,6-Tribromophenol	10.56	330	228455	98.27	ng	-0.02
44) 2-Fluorobiphenyl	8.77	172	1795541	93.10	ng	-0.01
78) Terphenyl-d14	13.39	244	1242498	92.38	ng	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032717\
Data File : BF093978.D
Acq On : 28 Mar 2017 7:57
Operator : SJ/MA
Sample : I2387-01
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
032317-PRRT-E-D-M

Quant Time: Mar 28 08:18:09 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 27 16:10:49 2017
Response via : Initial Calibration

