

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032718\
 Data File : BF103980.D
 Acq On : 27 Mar 2018 19:55
 Operator : JU/SJ
 Sample : J2066-01
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Mar 28 03:55:52 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF031518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 23 13:26:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.97	152	141058	20.00	ng	0.00
21) Naphthalene-d8	8.25	136	552239	20.00	ng	0.00
38) Acenaphthene-d10	10.02	164	209075	20.00	ng	0.00
63) Phenanthrene-d10	11.53	188	185671	20.00	ng	0.02
75) Chrysene-d12	14.17	240	252223	20.00	ng	0.00
86) Perylene-d12	15.70	264	243718	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.60	112	873920	94.23	ng	0.01
7) Phenol-d6	6.60	99	1067500	94.08	ng	0.00
23) Nitrobenzene-d5	7.53	82	611020	77.16	ng	0.00
41) 2,4,6-Tribromophenol	10.82	330	159239	88.58	ng	0.01
44) 2-Fluorobiphenyl	9.33	172	1004984	76.68	ng	0.00
78) Terphenyl-d14	13.11	244	574297	52.85	ng	0.01

Target Compounds


						Qvalue
9) Benzaldehyde	6.75	77	16998	2.280	ng	87
19) n-Nitroso-di-n-propylamine	7.53	70	80267	11.039	ng	# 72
25) Isophorone	7.53	82	611020	33.331	ng	# 64
26) 2-Nitrophenol	7.99	139	241	9.501	ng	# 1
32) Benzoic acid	8.06	122	114	9.921	ng	# 1
47) 2-Nitroaniline	9.55	65	1577	7.188	ng	# 55
49) Dimethylphthalate	9.72	163	121504	7.617	ng	# 88
50) 2,6-Dinitrotoluene	9.80	165	8294	6.449	ng	# 37
52) 3-Nitroaniline	9.96	138	9342	6.140	ng	# 28
53) 2,4-Dinitrophenol	10.20	184	1312	9.487	ng	# 1
55) 4-Nitrophenol	10.12	139	5067	6.597	ng	# 1
56) 2,4-Dinitrotoluene	10.16	165	3061	6.895	ng	# 43
61) 4-Nitroaniline	10.43	138	11601	6.754	ng	90
64) 4,6-Dinitro-2-methylphenol	10.63	198	2426	11.937	ng	# 1
65) n-Nitrosodiphenylamine	10.68	169	50108	7.929	ng	# 52
70) Phenanthrene	11.55	178	85022	8.371	ng	# 70
73) Di-n-butylphthalate	12.07	149	37977	3.157	ng	# 47
74) Fluoranthene	12.75	202	27889	2.665	ng	# 64
77) Pvrene	12.99	202	39086	2.110	ng	# 65
83) Bis(2-ethylhexyl)phthalate	14.12	149	42599	3.634	ng	# 95
87) Benzo(b)fluoranthene	15.24	252	37971	2.466	ng	# 81
88) Benzo(k)fluoranthene	15.24	252	37971	2.565	ng	# 84


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032718\
Data File : BF103980.D
Acq On : 27 Mar 2018 19:55
Operator : JU/SJ
Sample : J2066-01
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :

Quant Time: Mar 28 03:55:52 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF031518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 23 13:26:17 2018
Response via : Initial Calibration






#1



1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.97 min Scan# 830
Delta R.T. -0.01 min
Lab File: BF103980.D
Acq: 27 Mar 2018 19:55

Instrument :
BNA_F
ClientSampleId :

Tot Ion:152 Resp: 141058
Ion Ratio Lower Upper
152 100
150 156.1 124.6 186.8
115 64.0 50.1 75.1




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5

2-Fluorophenol
Concen: 94.233 ng
RT: 5.60 min Scan# 598
Delta R.T. 0.01 min
Lab File: BF103980.D
Acq: 27 Mar 2018 19:55

Tgt Ion:112 Resp: 873920
Ion Ratio Lower Upper
112 100
64 57.0 47.9 71.9
63 28.0 23.7 35.5



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF1
Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 94.085 ng

RT: 6.60 min Scan# 768

Delta R.T. 0.01 min

Lab File: BF103980.D

Acq: 27 Mar 2018 19:55

Instrument :

BNA_F

ClientSampled :


Tot Ion: 99 Resp: 1067500

Ion Ratio Lower Upper




99 100

42 15.7 14.5 21.7

71 32.7 25.4 38.0



Abundance Ion 99.00 (98.70 to 99.70): BF1
Ion 42.00 (41.70 to 42.70): BF1
Ion 71.00 (70.70 to 71.70): BF1



#9

Benzaldehyde

Concen: 2.280 ng

RT: 6.75 min Scan# 792

Delta R.T. 0.22 min

Lab File: BF103980.D

Acq: 27 Mar 2018 19:55


Tgt Ion: 77 Resp: 16998

Ion Ratio Lower Upper



77 100




105 85.0 81.1 121.1

106 85.2 74.5 114.5



Abundance Ion 77.00 (76.70 to 77.70): BF1
Ion 105.00 (104.70 to 105.70): BF1
Ion 106.00 (105.70 to 106.70): BF1





#19

n-Nitroso-di-n-propylamine

Concen: 11.039 ng

RT: 7.53 min Scan# 926

Delta R.T. 0.15 min

Lab File: BF103980.D

Acq: 27 Mar 2018 19:55

Instrument :

BNA_F

ClientSampleId :

Tot Ion: 70 Resp: 80267





Ion Ratio Lower Upper

70 100

42 41.3 47.5 71.3#

101 0.0 7.4 11.2#

130 2.3 16.6 25.0#



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.25 min Scan# 1048

Delta R.T. -0.01 min

Lab File: BF103980.D

Acq: 27 Mar 2018 19:55

Tgt Ion: 136 Resp: 552239


Ion Ratio Lower Upper


136 100

137 11.1 8.9 13.3

54 7.2 6.6 9.8

68 6.2 5.0 7.4





#23

Nitrobenzene-d5

Concen: 77.159 ng

RT: 7.53 min Scan# 926

Delta R.T. -0.01 min

Lab File: BF103980.D

Acq: 27 Mar 2018 19:55

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 82 Resp: 611020

Ion Ratio Lower Upper

82 100

128 46.8 36.1 54.1




54 44.8 41.4 62.2



Abundance Ion 82.00 (81.70 to 82.70): BF1

Ion 128.00 (127.70 to 128.70): BF1

Ion 54.00 (53.70 to 54.70): BF1



#25

Isophorone

Concen: 33.331 ng

RT: 7.53 min Scan# 926

Delta R.T. -0.26 min

Lab File: BF103980.D

Acq: 27 Mar 2018 19:55


Tgt Ion: 82 Resp: 611020

Ion Ratio Lower Upper

82 100

95 0.2 5.2 7.8#



138 0.0 14.9 22.3#




Abundance Ion 82.00 (81.70 to 82.70): BF1

Ion 95.00 (94.70 to 95.70): BF1

Ion 138.00 (137.70 to 138.70): BF1









#26
2-Nitrophenol
Concen: 9.501 ng
RT: 7.99 min Scan# 1003
Delta R.T. 0.11 min
Lab File: BF103980.D
Acq: 27 Mar 2018 19:55




Instrument :
BNA_F
ClientSampled :


Tot Ion:139 Resp: 241
Ion Ratio Lower Upper
139 100
109 339.3 22.7 34.1#
65 160.7 40.5 60.7#



#32
Benzoic acid
Concen: 9.921 ng
RT: 8.06 min Scan# 1016
Delta R.T. 0.05 min
Lab File: BF103980.D
Acq: 27 Mar 2018 19:55

Tgt Ion:122 Resp: 114
Ion Ratio Lower Upper
122 100
105 529.2 101.1 141.1#
77 222.7 70.7 110.7#









#38
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.02 min Scan# 1348
Delta R.T. -0.01 min
Lab File: BF103980.D
Acq: 27 Mar 2018 19:55

Instrument :
BNA_F
ClientSampleId :

Tot Ion:164 Resp: 209075
Ion Ratio Lower Upper
164 100
162 100.4 86.1 129.1
160 45.2 38.3 57.5




Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E






#41
2,4,6-Tribromophenol
Concen: 88.582 ng
RT: 10.82 min Scan# 1485
Delta R.T. 0.01 min
Lab File: BF103980.D
Acq: 27 Mar 2018 19:55

Tgt Ion:330 Resp: 159239
Ion Ratio Lower Upper
330 100
332 98.3 77.0 115.6
141 24.9 29.9 44.9#



Abundance Ion 329.80 (329.50 to 330.50): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E









#44
 2-Fluorobiphenyl
 Concen: 76.676 ng
 RT: 9.33 min Scan# 1231
 Delta R.T. -0.01 min
 Lab File: BF103980.D
 Acq: 27 Mar 2018 19:55

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:172 Resp: 1004984
 Ion Ratio Lower Upper
 172 100
 171 36.1 29.7 44.5
 170 24.1 19.6 29.4




Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E






#47
 2-Nitroaniline
 Concen: 7.188 ng
 RT: 9.55 min Scan# 1268
 Delta R.T. -0.01 min
 Lab File: BF103980.D
 Acq: 27 Mar 2018 19:55

Tgt Ion: 65 Resp: 1577
 Ion Ratio Lower Upper
 65 100
 92 100.1 55.8 83.6#
 138 60.3 91.2 136.8#



Abundance Ion 65.00 (64.70 to 65.70): BF1
 Ion 92.00 (91.70 to 92.70): BF1
 Ion 138.00 (137.70 to 138.70): E





#49

Dimethylphthalate

Concen: 7.617 ng

RT: 9.72 min Scan# 1298

Delta R.T. -0.01 min

Lab File: BF103980.D

Acq: 27 Mar 2018 19:55

Instrument :

BNA_F

ClientSampled :


Tot Ion:163 Resp: 121504

Ion Ratio Lower Upper

163 100

194 11.9 2.7 4.1#




164 12.9 8.2 12.2#



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#50

2,6-Dinitrotoluene

Concen: 6.449 ng

RT: 9.80 min Scan# 1312

Delta R.T. 0.01 min

Lab File: BF103980.D

Acq: 27 Mar 2018 19:55


Tgt Ion:165 Resp: 8294

Ion Ratio Lower Upper

165 100

63 12.7 44.7 67.1#



89 7.1 44.0 66.0#




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF1

Ion 89.00 (88.70 to 89.70): BF1





#52

3-Nitroaniline

Concen: 6.140 ng

RT: 9.96 min Scan# 1339

Delta R.T. -0.01 min

Lab File: BF103980.D

Acq: 27 Mar 2018 19:55

Instrument :

BNA_F

ClientSampled :





Tot Ion:138 Resp: 9342

Ion Ratio Lower Upper

138 100

108 89.2 9.4 14.0#

92 49.0 89.4 134.2#



#53

2,4-Dinitrophenol

Concen: 9.487 ng

RT: 10.20 min Scan# 1380

Delta R.T. 0.13 min

Lab File: BF103980.D

Acq: 27 Mar 2018 19:55




Tgt Ion:184 Resp: 1312


Ion Ratio Lower Upper

184 100

63 279.7 54.2 81.2#

154 341.3 184.0 276.0#









#55
4-Nitrophenol
Concen: 6.597 ng
RT: 10.12 min Scan# 1365
Delta R.T. -0.01 min
Lab File: BF103980.D
Acq: 27 Mar 2018 19:55




Instrument :
BNA_F
ClientSampled :


Tot Ion:139 Resp: 5067
Ion Ratio Lower Upper
139 100
109 695.7 48.4 88.4#
65 109.3 72.6 112.6



#56
2,4-Dinitrotoluene
Concen: 6.895 ng
RT: 10.16 min Scan# 1373
Delta R.T. -0.04 min
Lab File: BF103980.D
Acq: 27 Mar 2018 19:55

Tgt Ion:165 Resp: 3061
Ion Ratio Lower Upper
165 100
63 19.9 36.4 54.6#
89 14.0 57.8 86.6#
182 11.1 1.6 2.4#





#61

4-Nitroaniline

Concen: 6.754 ng

RT: 10.43 min Scan# 1418

Delta R.T. -0.16 min

Lab File: BF103980.D

Acq: 27 Mar 2018 19:55

Instrument :

BNA_F

ClientSampled :


Tot Ion:138 Resp: 11601

Ion Ratio Lower Upper

138 100

92 51.5 30.2 70.2




108 85.4 52.5 92.5



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BF1

Ion 108.00 (107.70 to 108.70): E



#63

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.53 min Scan# 1605

Delta R.T. 0.02 min

Lab File: BF103980.D

Acq: 27 Mar 2018 19:55


Tgt Ion:188 Resp: 185671

Ion Ratio Lower Upper

188 100

94 19.6 8.5 12.7#



80 19.9 9.2 13.8#




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1





#64

4,6-Dinitro-2-methylphenol

Concen: 11.937 ng

RT: 10.63 min Scan# 1453

Delta R.T. 0.02 min

Lab File: BF103980.D


Acq: 27 Mar 2018 19:55

Instrument :




BNA_F

ClientSampleId :

Tot Ion:198 Resp: 2426
 Ion Ratio Lower Upper
 198 100
 51 178.0 37.7 77.7#
 105 1068.7 36.3 76.3#



Abundance Ion 198.00 (197.70 to 198.70): E
 Ion 51.00 (50.70 to 51.70): BF1
 Ion 105.00 (104.70 to 105.70): E



#65

n-Nitrosodiphenylamine

Concen: 7.929 ng


RT: 10.68 min Scan# 1461

Delta R.T. 0.01 min



Lab File: BF103980.D


Acq: 27 Mar 2018 19:55

Tgt Ion:169 Resp: 50108
 Ion Ratio Lower Upper
 169 100
 168 15.9 54.4 81.6#
 167 26.5 29.8 44.6#



Abundance Ion 169.00 (168.70 to 169.70): E
 Ion 168.00 (167.70 to 168.70): E
 Ion 167.00 (166.70 to 167.70): E





#70

Phenanthrene

Concen: 8.371 ng

RT: 11.55 min Scan# 1609

Delta R.T. 0.01 min

Lab File: BF103980.D

Acq: 27 Mar 2018 19:55

Instrument :

BNA_F

ClientSampled :


Tot Ion:178 Resp: 85022

Ion Ratio Lower Upper

178 100

176 23.2 15.8 23.8




179 41.1 13.0 19.6#



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#73

Di-n-butylphthalate

Concen: 3.157 ng

RT: 12.07 min Scan# 1698

Delta R.T. 0.01 min

Lab File: BF103980.D

Acq: 27 Mar 2018 19:55


Tgt Ion:149 Resp: 37977

Ion Ratio Lower Upper

149 100

150 33.8 7.8 11.6#



104 14.1 3.8 5.8#




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E





#74

Fluoranthene

Concen: 2.665 ng

RT: 12.75 min Scan# 1813

Delta R.T. 0.02 min

Lab File: BF103980.D

Acq: 27 Mar 2018 19:55

Instrument :

BNA_F

ClientSampleId :


Tot Ion:202 Resp: 27889

Ion Ratio Lower Upper

202 100

101 16.5 0.0 34.1




203 44.2 0.0 38.1#



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#75

Chrysene-d12

Concen: 20.000 ng

RT: 14.17 min Scan# 2054

Delta R.T. -0.00 min

Lab File: BF103980.D

Acq: 27 Mar 2018 19:55


Tgt Ion:240 Resp: 252223

Ion Ratio Lower Upper

240 100

120 12.8 9.5 14.3



236 25.4 20.7 31.1




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#77

Pvrene

Concen: 2.110 ng

RT: 12.99 min Scan# 1853

Delta R.T. 0.02 min

Lab File: BF103980.D

Acq: 27 Mar 2018 19:55

Instrument :

BNA_F

ClientSampled :


Tot Ion:202 Resp: 39086

Ion Ratio Lower Upper

202 100

200 23.5 17.4 26.2

203 50.2 14.3 21.5#




Abundance Ion 202.00 (201.70 to 202.70): E


Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E

Time-->



Time-->



#78

Terphenyl-d14

Concen: 52.851 ng

RT: 13.11 min Scan# 1874

Delta R.T. 0.01 min

Lab File: BF103980.D

Acq: 27 Mar 2018 19:55


Tgt Ion:244 Resp: 574297

Ion Ratio Lower Upper

244 100

212 7.1 5.4 8.0

122 13.5 11.0 16.4




Abundance Ion 244.00 (243.70 to 244.70): E


Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E

Time-->







Time-->



#83
 Bis(2-ethylhexyl)phthalate
 Concen: 3.634 ng
 RT: 14.12 min Scan# 2045
 Delta R.T. -0.01 min
 Lab File: BF103980.D
 Acq: 27 Mar 2018 19:55




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:149 Resp: 42599
 Ion Ratio Lower Upper
 149 100
 167 29.1 21.3 31.9
 279 4.8 3.0 4.4#



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.70 min Scan# 2315
 Delta R.T. -0.01 min
 Lab File: BF103980.D
 Acq: 27 Mar 2018 19:55

Tgt Ion:264 Resp: 243718
 Ion Ratio Lower Upper
 264 100
 260 22.7 19.2 28.8
 265 20.7 17.5 26.3









#87
Benzo(b)fluoranthene
Concen: 2.466 ng
RT: 15.24 min Scan# 2237
Delta R.T. -0.01 min
Lab File: BF103980.D
Acq: 27 Mar 2018 19:55

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
252	100		
253	26.9	17.0	25.6#
125	23.4	9.0	13.6#



#88
Benzo(k)fluoranthene
Concen: 2.565 ng
RT: 15.24 min Scan# 2237
Delta R.T. -0.04 min
Lab File: BF103980.D
Acq: 27 Mar 2018 19:55

Tgt Ion	Ratio	Lower	Upper
252	100		
253	26.9	17.9	26.9#
125	23.4	10.2	15.2#

