


Data Path : Z:\HPCHEM1\BNA F\Data\BF032718\
 Data File : BF103994.D
 Acq On : 28 Mar 2018 2:55
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031518.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Mar 23 13:26:17 2018



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1635

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.8	26602	PASS
68	69	0.00	2	1.9	522	PASS
69	198	0.00	100	35.1	27621	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.6	39085	PASS
197	198	0.00	2	0.7	518	PASS
198	198	100	100	100.0	78768	PASS
199	198	5	9	6.7	5309	PASS
275	198	10	60	25.8	20353	PASS
365	198	1	100	3.0	2329	PASS
441	443	0.01	100	82.9	9986	PASS
442	198	50	100	82.6	65034	PASS
443	442	15	24	18.5	12047	PASS