

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032816\
 Data File : BF085808.D
 Acq On : 28 Mar 2016 18:22
 Operator : UM/SJ
 Sample : PB89183BL
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB89183BL

Quant Time: Mar 29 01:09:08 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 28 15:27:05 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.81	152	107120	20.00	ng	-0.01
21) Naphthalene-d8	8.10	136	417076	20.00	ng	-0.01
38) Acenaphthene-d10	9.85	164	188640	20.00	ng	-0.01
63) Phenanthrene-d10	11.34	188	338546	20.00	ng	-0.01
75) Chrysene-d12	13.97	240	190879	20.00	ng	-0.01
86) Perylene-d12	15.39	264	173154	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	675583	105.03	ng	0.01
7) Phenol-d6	6.45	99	756299	92.48	ng	-0.01
23) Nitrobenzene-d5	7.38	82	534359	72.15	ng	-0.01
41) 2,4,6-Tribromophenol	10.64	330	176548	98.50	ng	-0.01
44) 2-Fluorobiphenyl	9.18	172	869147	76.98	ng	-0.01
78) Terphenyl-d14	12.93	244	789794	85.09	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032816\
Data File : BF085808.D
Acq On : 28 Mar 2016 18:22
Operator : UM/SJ
Sample : PB89183BL
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB89183BL

Quant Time: Mar 29 01:09:08 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032416.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 28 15:27:05 2016
Response via : Initial Calibration

