

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032816\
 Data File : BF085841.D
 Acq On : 29 Mar 2016 9:51
 Operator : UM/SJ
 Sample : H1834-03
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 5

Manual Integrations
 APPROVED

Sohil
 3/29/2016 5:04:37 PM

Quant Time: Mar 29 14:45:51 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 28 15:27:05 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	79039m	20.00	ng	0.00
21) Naphthalene-d8	8.14	136	86945m	20.00	ng	0.02
38) Acenaphthene-d10	9.94	164	81815m	20.00	ng	0.08
63) Phenanthrene-d10	11.42	188	101522m	20.00	ng	0.07
75) Chrysene-d12	13.99	240	181115m	20.00	ng	0.01
86) Perylene-d12	15.40	264	203114m	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.41	112	682241	143.75	ng	0.00
7) Phenol-d6	6.46	99	738116m	122.32	ng	0.00
23) Nitrobenzene-d5	7.41	82	322774	209.06	ng	0.01
41) 2,4,6-Tribromophenol	10.74	330	79777	102.63	ng	0.09
44) 2-Fluorobiphenyl	9.25	172	329264	67.24	ng	0.06
78) Terphenyl-d14	12.97	244	382377m	43.42	ng	0.05
Target Compounds						
31) Naphthalene	8.21	128	644673	147.15	ng	# 1
37) 2-Methylnaphthalene	9.01	142	2655879	Below Cal		# 86
45) 1,1'-Biphenyl	9.34	154	459897	66.62	ng	94
51) Acenaphthene	9.99	154	333513m	66.01	ng	
57) Fluorene	10.50	166	730999m	134.47	ng	
70) Phenanthrene	11.45	178	2102315m	403.30	ng	
77) Pyrene	12.85	202	300364m	20.18	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032816\
Data File : BF085841.D
Acq On : 29 Mar 2016 9:51
Operator : UM/SJ
Sample : H1834-03
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
5

Manual Integrations
APPROVED

Sohil
3/29/2016 5:04:37 PM

Quant Time: Mar 29 14:45:51 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032416.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 28 15:27:05 2016
Response via : Initial Calibration

