

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032818\
 Data File : BF104008.D
 Acq On : 28 Mar 2018 14:26
 Operator : JU/SJ
 Sample : SSTDC0040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Mar 28 16:49:24 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 28 13:41:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.97	152	137520	20.00	ng	0.00
21) Naphthalene-d8	8.25	136	568249	20.00	ng	0.00
38) Acenaphthene-d10	10.00	164	267125	20.00	ng	0.00
63) Phenanthrene-d10	11.50	188	490047	20.00	ng	0.00
75) Chrysene-d12	14.16	240	387591	20.00	ng	0.00
86) Perylene-d12	15.70	264	387675	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.60	112	668840	77.56	ng	0.00
7) Phenol-d6	6.60	99	799070	75.32	ng	0.00
23) Nitrobenzene-d5	7.53	82	730010	78.56	ng	0.00
41) 2,4,6-Tribromophenol	10.80	330	231250	77.75	ng	0.00
44) 2-Fluorobiphenyl	9.32	172	1317909	77.80	ng	0.00
78) Terphenyl-d14	13.09	244	1249721	74.45	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	2.82	88	134603	36.192	ng	91
3) Pyridine	3.62	79	347701	36.929	ng	89
4) n-Nitrosodimethylamine	3.59	42	86030	30.740	ng	79
6) Aniline	6.64	93	506741	39.899	ng	93
8) 2-Chlorophenol	6.76	128	367516	38.852	ng	96
9) Benzaldehyde	6.53	77	228245	42.461	ng	95
10) Phenol	6.62	94	436323	37.118	ng	98
11) bis(2-Chloroethyl)ether	6.70	93	404572	39.940	ng	92
12) 1,3-Dichlorobenzene	6.90	146	410250	38.582	ng	99
13) 1,4-Dichlorobenzene	6.99	146	437803	38.426	ng	99
14) 1,2-Dichlorobenzene	7.13	146	393199	38.396	ng	100
15) Benzyl Alcohol	7.11	79	273378	39.632	ng	98
16) 2,2'-oxybis(1-Chloropropan	7.23	45	424157	38.435	ng	62
17) 2-Methylphenol	7.22	107	301955	39.220	ng	# 91
18) Hexachloroethane	7.47	117	147706	38.720	ng	94
19) n-Nitroso-di-n-propylamine	7.37	70	244680	38.862	ng	91
20) 3+4-Methylphenols	7.37	107	396272	40.330	ng	# 64
22) Acetophenone	7.38	105	520328	37.846	ng	98
24) Nitrobenzene	7.55	77	384431	39.322	ng	98
25) Isophorone	7.78	82	645234	37.680	ng	100
26) 2-Nitrophenol	7.86	139	197674	38.735	ng	96
27) 2,4-Dimethylphenol	7.90	122	278463	37.835	ng	99
28) bis(2-Chloroethoxy)methane	7.99	93	412841	38.467	ng	99
29) 2,4-Dichlorophenol	8.11	162	300937	39.146	ng	98
30) 1,2,4-Trichlorobenzene	8.19	180	339928	38.972	ng	97
31) Naphthalene	8.27	128	1060645	38.206	ng	99
32) Benzoic acid	8.04	122	207246	38.438	ng	97
33) 4-Chloroaniline	8.32	127	463700	39.620	ng	98
34) Hexachlorobutadiene	8.37	225	182410	38.430	ng	98
35) Caprolactam	8.70	113	93351	38.297	ng	# 76
36) 4-Chloro-3-methylphenol	8.80	107	319887	39.340	ng	95
37) 2-Methylnaphthalene	8.96	142	700593	38.312	ng	99
39) 1,2,4,5-Tetrachlorobenzene	9.12	216	315069	38.463	ng	98
40) Hexachlorocyclopentadiene	9.10	237	141546	39.625	ng	98


Data Path : Z:\HPCHEM1\BNA F\DATA\BF032818\
 Data File : BF104008.D
 Acq On : 28 Mar 2018 14:26
 Operator : JU/SJ
 Sample : SSTDC0040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Mar 28 16:49:24 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 28 13:41:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	9.24	196	204316	39.390	ng	99
43) 2,4,5-Trichlorophenol	9.29	196	242130	38.661	ng	95
45) 1,1'-Biphenyl	9.42	154	878642	38.324	ng	99
46) 2-Chloronaphthalene	9.45	162	695000	38.430	ng	99
47) 2-Nitroaniline	9.56	65	197493	38.297	ng	95
48) Acenaphthylene	9.87	152	1033110	37.708	ng	100
49) Dimethylphthalate	9.72	163	798042	37.839	ng	100
50) 2,6-Dinitrotoluene	9.79	165	177086	39.028	ng	90
51) Acenaphthene	10.04	154	593386	37.439	ng	100
52) 3-Nitroaniline	9.97	138	198775	38.109	ng	99
53) 2,4-Dinitrophenol	10.07	184	72002	39.389	ng	# 30
54) Dibenzofuran	10.22	168	868859	37.540	ng	99
55) 4-Nitrophenol	10.14	139	120828	38.634	ng	92
56) 2,4-Dinitrotoluene	10.20	165	226757	39.164	ng	93
57) Fluorene	10.56	166	711815	37.284	ng	99
58) 2,3,4,6-Tetrachlorophenol	10.33	232	180615	38.492	ng	89
59) Diethylphthalate	10.42	149	767207	37.972	ng	99
60) 4-Chlorophenyl-phenylether	10.55	204	331307	37.783	ng	92
61) 4-Nitroaniline	10.59	138	183328	37.563	ng	95
62) Azobenzene	10.70	77	700329	38.333	ng	95
64) 4,6-Dinitro-2-methylphenol	10.61	198	121215	40.863	ng	81
65) n-Nitrosodiphenylamine	10.67	169	626400	37.743	ng	100
66) 4-Bromophenyl-phenylether	11.03	248	201665	38.465	ng	94
67) Hexachlorobenzene	11.10	284	228732	37.927	ng	# 91
68) Atrazine	11.19	200	193874	38.131	ng	98
69) Pentachlorophenol	11.30	266	131881	39.740	ng	96
70) Phenanthrene	11.53	178	987075	37.204	ng	99
71) Anthracene	11.58	178	1039963	37.526	ng	99
72) Carbazole	11.74	167	994579	37.602	ng	99
73) Di-n-butylphthalate	12.04	149	1178682	37.411	ng	99
74) Fluoranthene	12.72	202	1044332	37.031	ng	100
76) Benzidine	12.84	184	465929	42.183	ng	98
77) Pyrene	12.95	202	1074350	38.559	ng	99
79) Butylbenzylphthalate	13.55	149	504110	38.404	ng	99
80) Benzo(a)anthracene	14.14	228	925375	38.156	ng	99
81) 3,3'-Dichlorobenzidine	14.10	252	300117	37.386	ng	97
82) Chrysene	14.19	228	896416	37.738	ng	99
83) Bis(2-ethylhexyl)phthalate	14.10	149	638398	37.640	ng	99
84) Di-n-octyl phthalate	14.73	149	1110832	38.075	ng	# 95
85) Indeno(1,2,3-cd)pyrene	17.29	276	874041	35.508	ng	96
87) Benzo(b)fluoranthene	15.24	252	854296	36.209	ng	98
88) Benzo(k)fluoranthene	15.27	252	880999	39.640	ng	98
89) Benzo(a)pyrene	15.63	252	824705	37.720	ng	98
90) Dibenzo(a,h)anthracene	17.30	278	719768	35.606	ng	98
91) Benzo(g,h,i)perylene	17.77	276	707207	34.928	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed




#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.97 min Scan# 830
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26

Instrument :
BNA_F
ClientSampleId :

Tot Ion:152 Resp: 137520
Ion Ratio Lower Upper
152 100
150 157.7 124.6 186.8
115 60.6 50.1 75.1




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 36.192 ng
RT: 2.82 min Scan# 125
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26

Tgt Ion: 88 Resp: 134603
Ion Ratio Lower Upper
88 100
58 71.3 62.6 93.8
43 24.7 24.6 37.0



Abundance Ion 88.00 (87.70 to 88.70): BF1
Ion 58.00 (57.70 to 58.70): BF1
Ion 43.00 (42.70 to 43.70): BF1









#3
 Pvridine
 Concen: 36.929 ng
 RT: 3.62 min Scan# 261
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
79	100		
52	64.9	59.8	89.6
51	31.9	29.8	44.8



#4
 n-Nitrosodimethylamine
 Concen: 30.740 ng
 RT: 3.59 min Scan# 255
 Delta R.T. 0.01 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26

Tgt Ion	Ratio	Lower	Upper
42	100		
74	156.9	104.9	157.3
44	8.1	6.9	10.3





#5

2-Fluorophenol

Concen: 77.562 ng

RT: 5.60 min Scan# 597

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26

Instrument :

BNA_F

ClientSampled :


Tot Ion:112 Resp: 668840

Ion Ratio Lower Upper

112 100

64 55.6 47.9 71.9




63 27.4 23.7 35.5



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF1

Ion 63.00 (62.70 to 63.70): BF1



#6

Aniline

Concen: 39.899 ng

RT: 6.64 min Scan# 774

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26


Tgt Ion: 93 Resp: 506741

Ion Ratio Lower Upper

93 100

66 44.3 32.2 48.2



65 23.9 16.4 24.6




Abundance Ion 93.00 (92.70 to 93.70): BF1

Ion 66.00 (65.70 to 66.70): BF1

Ion 65.00 (64.70 to 65.70): BF1





#7

Phenol-d6

Concen: 75.318 ng

RT: 6.60 min Scan# 768

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 99 Resp: 799070

Ion Ratio Lower Upper




99 100

42 15.1 14.5 21.7

71 31.8 25.4 38.0



Abundance Ion 99.00 (98.70 to 99.70): BF1
Ion 42.00 (41.70 to 42.70): BF1
Ion 71.00 (70.70 to 71.70): BF1



#8

2-Chlorophenol

Concen: 38.852 ng

RT: 6.76 min Scan# 794

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26


Tgt Ion: 128 Resp: 367516

Ion Ratio Lower Upper



128 100


130 33.2 13.0 53.0

64 44.6 29.4 69.4



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 130.00 (129.70 to 130.70): E
Ion 64.00 (63.70 to 64.70): BF1





#9

Benzaldehyde

Concen: 42.461 ng

RT: 6.53 min Scan# 755

Delta R.T. 0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26

Instrument :

BNA_F

ClientSampled :


Tot Ion: 77 Resp: 228245

Ion Ratio Lower Upper

77 100

105 105.2 81.1 121.1




106 101.2 74.5 114.5



Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 105.00 (104.70 to 105.70): BF1

Ion 106.00 (105.70 to 106.70): BF1



#10

Phenol

Concen: 37.118 ng

RT: 6.62 min Scan# 771

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26


Tgt Ion: 94 Resp: 436323

Ion Ratio Lower Upper

94 100

65 26.8 7.9 47.9



66 37.8 16.2 56.2



Abundance Ion 94.00 (93.70 to 94.70): BF1

Ion 65.00 (64.70 to 65.70): BF1

Ion 66.00 (65.70 to 66.70): BF1









#11
 bis(2-Chloroethyl)ether
 Concen: 39.940 ng
 RT: 6.70 min Scan# 785
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion: 93 Resp: 404572
 Ion Ratio Lower Upper
 93 100
 63 69.3 58.6 98.6
 95 31.6 11.8 51.8



#12
 1,3-Dichlorobenzene
 Concen: 38.582 ng
 RT: 6.90 min Scan# 819
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26

Tgt Ion: 146 Resp: 410250
 Ion Ratio Lower Upper
 146 100
 148 63.8 51.8 77.8
 75 30.5 24.2 36.4









#13
 1,4-Dichlorobenzene
 Concen: 38.426 ng
 RT: 6.99 min Scan# 833
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26




Instrument :
 BNA_F
 ClientSampled :


Tot Ion:146 Resp: 437803
 Ion Ratio Lower Upper
 146 100
 148 63.0 50.8 76.2
 111 41.5 34.2 51.2



#14
 1,2-Dichlorobenzene
 Concen: 38.396 ng
 RT: 7.13 min Scan# 858
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26

Tgt Ion:146 Resp: 393199
 Ion Ratio Lower Upper
 146 100
 148 63.3 50.6 75.8
 111 45.5 36.0 54.0









#15
Benzyl Alcohol
Concen: 39.632 ng
RT: 7.11 min Scan# 854
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26

Instrument :
BNA_F
ClientSampled :

Tot Ion: 79 Resp: 273378
Ion Ratio Lower Upper
79 100
108 79.4 65.1 97.7
77 61.5 50.6 75.8




Abundance Ion 79.00 (78.70 to 79.70): BF1
Ion 108.00 (107.70 to 108.70): BF1
Ion 77.00 (76.70 to 77.70): BF1






#16
2,2'-oxybis(1-chloropropane)
Concen: 38.435 ng
RT: 7.23 min Scan# 875
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26

Tgt Ion: 45 Resp: 424157
Ion Ratio Lower Upper
45 100
77 27.9 0.0 32.9
79 23.7 0.0 29.6



Abundance Ion 45.00 (44.70 to 45.70): BF1
Ion 77.00 (76.70 to 77.70): BF1
Ion 79.00 (78.70 to 79.70): BF1









#17
2-Methylphenol
Concen: 39.220 ng
RT: 7.22 min Scan# 873
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26




Instrument :
BNA_F
ClientSampled :


Tot Ion:107 Resp: 301955
Ion Ratio Lower Upper
107 100
108 110.5 91.7 137.5
77 54.2 33.8 50.8#
79 51.3 33.8 50.8#



#18
Hexachloroethane
Concen: 38.720 ng
RT: 7.47 min Scan# 915
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26

Tgt Ion:117 Resp: 147706
Ion Ratio Lower Upper
117 100
119 97.4 75.8 113.8
201 85.2 75.7 113.5





#19

n-Nitroso-di-n-propylamine

Concen: 38.862 ng

RT: 7.37 min Scan# 899

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26

Instrument :

BNA_F

ClientSampleId :

Tot Ion: 70 Resp: 244680


Ion	Ratio	Lower	Upper
70	100		
42	51.2	47.5	71.3
101	9.9	7.4	11.2
130	23.2	16.6	25.0

70 100

42 51.2 47.5 71.3

101 9.9 7.4 11.2

130 23.2 16.6 25.0






Abundance

Ion 70.00 (69.70 to 70.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 101.00 (100.70 to 101.70): BF1

Ion 130.00 (129.70 to 130.70): BF1



#20

3+4-Methylphenols

Concen: 40.330 ng

RT: 7.37 min Scan# 899

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26

Tgt Ion: 107 Resp: 396272


Ion	Ratio	Lower	Upper
107	100		
108	88.1	68.2	108.2
77	128.9	26.7	66.7#
79	27.5	6.3	46.3

107 100

108 88.1 68.2 108.2

77 128.9 26.7 66.7#

79 27.5 6.3 46.3





Abundance


Ion 107.00 (106.70 to 107.70): BF1

Ion 108.00 (107.70 to 108.70): BF1

Ion 77.00 (76.70 to 77.70): BF1

Ion 79.00 (78.70 to 79.70): BF1





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.25 min Scan# 1048

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26

Instrument :

BNA_F

ClientSampleId :

Tot Ion:136 Resp: 568249


Ion Ratio Lower Upper

136 100

137 10.9 8.9 13.3

54 6.3 6.6 9.8#

68 5.5 5.0 7.4






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF1

Ion 68.00 (67.70 to 68.70): BF1



#22

Acetophenone

Concen: 37.846 ng

RT: 7.38 min Scan# 900

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26

Tgt Ion:105 Resp: 520328


Ion Ratio Lower Upper

105 100

71 4.9 4.4 6.6

51 26.6 22.3 33.5

120 22.3 16.9 25.3






Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF1

Ion 51.00 (50.70 to 51.70): BF1

Ion 120.00 (119.70 to 120.70): E









#23
Nitrobenzene-d5
Concen: 78.564 ng
RT: 7.53 min Scan# 926
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26

Instrument :
BNA_F
ClientSampleId :

Tot Ion: 82 Resp: 730010
Ion Ratio Lower Upper
82 100
128 46.6 36.1 54.1
54 43.7 41.4 62.2




Abundance Ion 82.00 (81.70 to 82.70): BF1
Ion 128.00 (127.70 to 128.70): BF1
Ion 54.00 (53.70 to 54.70): BF1






#24
Nitrobenzene
Concen: 39.322 ng
RT: 7.55 min Scan# 929
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26

Tgt Ion: 77 Resp: 384431
Ion Ratio Lower Upper
77 100
123 48.2 37.9 56.9
65 12.6 11.0 16.4



Abundance Ion 77.00 (76.70 to 77.70): BF1
Ion 123.00 (122.70 to 123.70): BF1
Ion 65.00 (64.70 to 65.70): BF1





#25

Isophorone

Concen: 37.680 ng

RT: 7.78 min Scan# 968

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26

Instrument :

BNA_F

ClientSampleId :





Tot Ion: 82 Resp: 645234

Ion Ratio Lower Upper

82 100

95 6.7 5.2 7.8

138 18.5 14.9 22.3



#26

2-Nitrophenol

Concen: 38.735 ng

RT: 7.86 min Scan# 982

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26




Tgt Ion: 139 Resp: 197674


Ion Ratio Lower Upper

139 100

109 27.3 22.7 34.1

65 46.8 40.5 60.7









#27
 2,4-Dimethylphenol
 Concen: 37.835 ng
 RT: 7.90 min Scan# 988
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26




Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:122 Resp: 278463
 Ion Ratio Lower Upper
 122 100
 107 113.0 91.2 136.8
 121 57.7 47.2 70.8



#28
 bis(2-Chloroethoxy)methane
 Concen: 38.467 ng
 RT: 7.99 min Scan# 1004
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26

Tgt Ion: 93 Resp: 412841
 Ion Ratio Lower Upper
 93 100
 95 32.5 26.3 39.5
 123 11.6 9.8 14.6









#29
 2,4-Dichlorophenol
 Concen: 39.146 ng
 RT: 8.11 min Scan# 1024
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26

Instrument :
 BNA_F
 ClientSampled :

Tot Ion:162 Resp: 300937
 Ion Ratio Lower Upper
 162 100
 164 64.8 42.7 82.7
 98 37.1 18.1 58.1




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 98.00 (97.70 to 98.70): BF1






#30
 1,2,4-Trichlorobenzene
 Concen: 38.972 ng
 RT: 8.19 min Scan# 1037
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26

Tgt Ion:180 Resp: 339928
 Ion Ratio Lower Upper
 180 100
 182 93.6 76.8 115.2
 145 31.3 26.5 39.7



Abundance Ion 180.00 (179.70 to 180.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 145.00 (144.70 to 145.70): E









#31
Naphthalene
Concen: 38.206 ng
RT: 8.27 min Scan# 1051
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26




Instrument :
BNA_F
ClientSampleId :


Tot Ion:128 Resp: 1060645
Ion Ratio Lower Upper
128 100
129 11.4 8.8 13.2
127 13.7 10.9 16.3



#32
Benzoic acid
Concen: 38.438 ng
RT: 8.04 min Scan# 1012
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26

Tgt Ion:122 Resp: 207246
Ion Ratio Lower Upper
122 100
105 119.4 101.1 141.1
77 86.6 70.7 110.7





#33

4-Chloroaniline

Concen: 39.620 ng

RT: 8.32 min Scan# 1060

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26

Instrument :

BNA_F

ClientSampled :

Tot Ion:127 Resp: 463700


Ion Ratio Lower Upper

127 100

129 32.8 25.9 38.9

65 28.4 25.0 37.4

92 19.6 15.2 22.8






Abundance Ion 127.00 (126.70 to 127.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 65.00 (64.70 to 65.70): BF1

Ion 92.00 (91.70 to 92.70): BF1



#34

Hexachlorobutadiene

Concen: 38.430 ng

RT: 8.37 min Scan# 1069

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26


Tgt Ion:225 Resp: 182410

Ion Ratio Lower Upper

225 100

223 62.4 50.6 76.0



227 62.1 51.9 77.9




Abundance Ion 225.00 (224.70 to 225.70): E

Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E









#35
 Caprolactam
 Concen: 38.297 ng
 RT: 8.70 min Scan# 1125
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:113 Resp: 93351
 Ion Ratio Lower Upper
 113 100
 55 148.4 163.3 203.3#
 56 108.4 115.7 155.7#



#36
 4-Chloro-3-methylphenol
 Concen: 39.340 ng
 RT: 8.80 min Scan# 1142
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26

Tgt Ion:107 Resp: 319887
 Ion Ratio Lower Upper
 107 100
 144 27.0 20.5 30.7
 142 82.7 62.0 93.0









#37
 2-Methylnaphthalene
 Concen: 38.312 ng
 RT: 8.96 min Scan# 1168
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
142	100		
141	89.2	70.8	106.2
115	32.8	26.1	39.1



#38
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 10.00 min Scan# 1346
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26

Tgt Ion	Ratio	Lower	Upper
164	100		
162	103.9	86.1	129.1
160	46.0	38.3	57.5









#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 38.463 ng
 RT: 9.12 min Scan# 1196
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:216	Resp:	315069
Ion Ratio	Lower	Upper
216	100	
214	80.3	63.0 94.4
179	22.2	18.1 27.1
108	19.7	17.2 25.8




Abundance Ion 216.00 (215.70 to 216.70): E
 Ion 214.00 (213.70 to 214.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 108.00 (107.70 to 108.70): E






#40
 Hexachlorocyclopentadiene
 Concen: 39.625 ng
 RT: 9.10 min Scan# 1193
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26

Tgt Ion:237	Resp:	141546
Ion Ratio	Lower	Upper
237	100	
235	63.4	44.8 84.8
272	13.8	0.0 33.3



Abundance Ion 236.80 (236.50 to 237.50): E
 Ion 234.90 (234.60 to 235.60): E
 Ion 271.75 (271.45 to 272.45): E









#41
 2,4,6-Tribromophenol
 Concen: 77.746 ng
 RT: 10.80 min Scan# 1482
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26




Instrument :
 BNA_F
 ClientSampled :


Tot Ion	Ratio	Lower	Upper
330	100		
332	95.5	77.0	115.6
141	31.3	29.9	44.9



#42
 2,4,6-Trichlorophenol
 Concen: 39.390 ng
 RT: 9.24 min Scan# 1216
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26

Tgt Ion	Ratio	Lower	Upper
196	100		
198	94.3	75.7	113.5
200	29.8	24.5	36.7









#43
 2,4,5-Trichlorophenol
 Concen: 38.661 ng
 RT: 9.29 min Scan# 1224
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:196 Resp: 242130
 Ion Ratio Lower Upper
 196 100
 198 95.9 74.6 112.0
 97 47.3 42.9 64.3
 132 29.3 26.3 39.5




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF1
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 77.800 ng
 RT: 9.32 min Scan# 1230
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26

Tgt Ion:172 Resp: 1317909
 Ion Ratio Lower Upper
 172 100
 171 35.9 29.7 44.5
 170 23.9 19.6 29.4



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#45

1,1'-Biphenyl

Concen: 38.324 ng

RT: 9.42 min Scan# 1247

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26

Instrument :

BNA_F

ClientSampleId :


Tot Ion:154 Resp: 878642

Ion Ratio Lower Upper

154 100

153 40.6 21.3 61.3




76 13.5 0.0 34.0



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 76.00 (75.70 to 76.70): BF1



#46

2-Chloronaphthalene

Concen: 38.430 ng

RT: 9.45 min Scan# 1252

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26


Tgt Ion:162 Resp: 695000

Ion Ratio Lower Upper

162 100

127 42.7 33.9 50.9



164 33.4 26.5 39.7




Abundance Ion 162.00 (161.70 to 162.70): E

Ion 127.00 (126.70 to 127.70): E

Ion 164.00 (163.70 to 164.70): E









#47
2-Nitroaniline
Concen: 38.297 ng
RT: 9.56 min Scan# 1270
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26

Instrument :
BNA_F
ClientSampleID :

Tot Ion: 65 Resp: 197493
Ion Ratio Lower Upper
65 100
92 73.1 55.8 83.6
138 119.3 91.2 136.8




Abundance Ion 65.00 (64.70 to 65.70): BF1
Ion 92.00 (91.70 to 92.70): BF1
Ion 138.00 (137.70 to 138.70): BF1






#48
Acenaphthylene
Concen: 37.708 ng
RT: 9.87 min Scan# 1323
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26

Tgt Ion: 152 Resp: 1033110
Ion Ratio Lower Upper
152 100
151 20.4 16.3 24.5
153 13.3 10.7 16.1



Abundance Ion 152.00 (151.70 to 152.70): BF1
Ion 151.00 (150.70 to 151.70): BF1
Ion 153.00 (152.70 to 153.70): BF1









#49
Dimethylphthalate
Concen: 37.839 ng
RT: 9.72 min Scan# 1298
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26

Instrument :
BNA_F
ClientSampled :

Tot Ion:163 Resp: 798042
Ion Ratio Lower Upper
163 100
194 3.6 2.7 4.1
164 10.4 8.2 12.2




Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E






#50
2,6-Dinitrotoluene
Concen: 39.028 ng
RT: 9.79 min Scan# 1310
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26

Tgt Ion:165 Resp: 177086
Ion Ratio Lower Upper
165 100
63 48.2 44.7 67.1
89 48.6 44.0 66.0



Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BF1
Ion 89.00 (88.70 to 89.70): BF1





#51

Acenaphthene

Concen: 37.439 ng

RT: 10.04 min Scan# 1352

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26

Instrument :

BNA_F

ClientSampleId :


Tot Ion:154 Resp: 593386

Ion Ratio Lower Upper

154 100

153 113.9 91.2 136.8




152 54.8 43.8 65.8



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 38.109 ng

RT: 9.97 min Scan# 1341

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26


Tgt Ion:138 Resp: 198775

Ion Ratio Lower Upper

138 100

108 11.1 9.4 14.0



92 110.8 89.4 134.2




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF1









#53
 2,4-Dinitrophenol
 Concen: 39.389 ng
 RT: 10.07 min Scan# 1358
 Delta R.T. 0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:184 Resp: 72002
 Ion Ratio Lower Upper
 184 100
 63 64.4 54.2 81.2
 154 81.6 184.0 276.0#



#54
 Dibenzofuran
 Concen: 37.540 ng
 RT: 10.22 min Scan# 1382
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26

Tgt Ion:168 Resp: 868859
 Ion Ratio Lower Upper
 168 100
 139 38.4 30.1 45.1
 169 13.6 10.9 16.3









#55
4-Nitrophenol
Concen: 38.634 ng
RT: 10.14 min Scan# 1369
Delta R.T. -0.01 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26




Instrument :
BNA_F
ClientSampleId :


Tot Ion:139 Resp: 120828
Ion Ratio Lower Upper
139 100
109 61.0 48.4 88.4
65 85.6 72.6 112.6



#56
2,4-Dinitrotoluene
Concen: 39.164 ng
RT: 10.20 min Scan# 1379
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26

Tgt Ion:165 Resp: 226757
Ion Ratio Lower Upper
165 100
63 39.7 36.4 54.6
89 66.7 57.8 86.6
182 2.4 1.6 2.4









#57
 Fluorene
 Concen: 37.284 ng
 RT: 10.56 min Scan# 1440
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26




Instrument :
 BNA_F
 ClientSampled :


Tot Ion:166 Resp: 711815
 Ion Ratio Lower Upper
 166 100
 165 98.4 79.8 119.8
 167 13.2 10.6 16.0



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 38.492 ng
 RT: 10.33 min Scan# 1402
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26

Tgt Ion:232 Resp: 180615
 Ion Ratio Lower Upper
 232 100
 131 44.0 43.8 65.8
 130 2.3 2.1 3.1
 166 31.1 27.8 41.6









#59
Diethylphthalate
Concen: 37.972 ng
RT: 10.42 min Scan# 1416
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26

Instrument :
BNA_F
ClientSampled :

Tot Ion:149 Resp: 767207
Ion Ratio Lower Upper
149 100
177 20.9 16.1 24.1
150 13.0 10.1 15.1




Abundance Ion 149.00 (148.70 to 149.70): E
Ion 177.00 (176.70 to 177.70): E
Ion 150.00 (149.70 to 150.70): E






#60
4-Chlorophenyl-phenylether
Concen: 37.783 ng
RT: 10.55 min Scan# 1438
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26

Tgt Ion:204 Resp: 331307
Ion Ratio Lower Upper
204 100
206 32.3 26.5 39.7
141 59.4 54.6 81.8



Abundance Ion 204.00 (203.70 to 204.70): E
Ion 206.00 (205.70 to 206.70): E
Ion 141.00 (140.70 to 141.70): E









#61
4-Nitroaniline
Concen: 37.563 ng
RT: 10.59 min Scan# 1446
Delta R.T. -0.01 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26




Instrument :
BNA_F
ClientSampleId :


Tot Ion:138 Resp: 183328
Ion Ratio Lower Upper
138 100
92 49.7 30.2 70.2
108 66.0 52.5 92.5



#62
Azobenzene
Concen: 38.333 ng
RT: 10.70 min Scan# 1465
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26

Tgt Ion: 77 Resp: 700329
Ion Ratio Lower Upper
77 100
182 24.3 1.7 41.7
105 23.9 3.0 43.0
51 25.8 9.9 49.9









#63
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.50 min Scan# 1600
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26




Instrument :
BNA_F
ClientSampleId :


Tot Ion:188 Resp: 490047
Ion Ratio Lower Upper
188 100
94 10.9 8.5 12.7
80 11.6 9.2 13.8



#64
4,6-Dinitro-2-methylphenol
Concen: 40.863 ng
RT: 10.61 min Scan# 1449
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26

Tgt Ion:198 Resp: 121215
Ion Ratio Lower Upper
198 100
51 40.6 37.7 77.7
105 45.0 36.3 76.3





#65

n-Nitrosodiphenylamine

Concen: 37.743 ng

RT: 10.67 min Scan# 1459

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26

Instrument :

BNA_F

ClientSampleId :


Tot Ion:169 Resp: 626400

Ion Ratio Lower Upper

169 100

168 68.3 54.4 81.6




167 37.0 29.8 44.6



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#66

4-Bromophenyl-phenylether

Concen: 38.465 ng

RT: 11.03 min Scan# 1521

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26


Tgt Ion:248 Resp: 201665

Ion Ratio Lower Upper

248 100

250 98.1 76.9 115.3



141 84.1 74.4 111.6




Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E





#67

Hexachlorobenzene

Concen: 37.927 ng

RT: 11.10 min Scan# 1533

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26

Instrument :

BNA_F

ClientSampleId :


Tot Ion:284 Resp: 228732

Ion Ratio Lower Upper

284 100

142 33.6 34.6 52.0#




249 30.4 24.8 37.2



Abundance Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E



#68

Atrazine

Concen: 38.131 ng

RT: 11.19 min Scan# 1548

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26


Tgt Ion:200 Resp: 193874

Ion Ratio Lower Upper

200 100

173 28.5 8.6 48.6



215 46.8 25.1 65.1




Abundance Ion 200.00 (199.70 to 200.70): E

Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E





#69

Pentachlorophenol

Concen: 39.740 ng

RT: 11.30 min Scan# 1567

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26

Instrument :

BNA_F

ClientSampleId :


Tot Ion:266 Resp: 131881

Ion Ratio Lower Upper

266 100

268 68.5 51.1 76.7




264 63.7 49.5 74.3



Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E



#70

Phenanthrene

Concen: 37.204 ng

RT: 11.53 min Scan# 1605

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26


Tgt Ion:178 Resp: 987075

Ion Ratio Lower Upper

178 100

176 19.6 15.8 23.8



179 15.8 13.0 19.6




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E









#71
 Anthracene
 Concen: 37.526 ng
 RT: 11.58 min Scan# 1614
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:178 Resp: 1039963
 Ion Ratio Lower Upper
 178 100
 176 18.8 15.4 23.2
 179 15.4 12.6 19.0



#72
 Carbazole
 Concen: 37.602 ng
 RT: 11.74 min Scan# 1641
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26

Tgt Ion:167 Resp: 994579
 Ion Ratio Lower Upper
 167 100
 166 22.0 17.6 26.4
 139 12.9 10.9 16.3





#73

Di-n-butylphthalate

Concen: 37.411 ng

RT: 12.04 min Scan# 1693

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26

Instrument :

BNA_F

ClientSampleId :


Tot Ion:149 Resp: 1178682

Ion Ratio Lower Upper

149 100

150 9.2 7.8 11.6




104 4.5 3.8 5.8



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 37.031 ng

RT: 12.72 min Scan# 1807

Delta R.T. -0.01 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26


Tgt Ion:202 Resp: 1044332

Ion Ratio Lower Upper

202 100

101 13.8 0.0 34.1



203 18.1 0.0 38.1




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.000 ng

RT: 14.16 min Scan# 2052

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26

Instrument :

BNA_F

ClientSampleId :


Tot Ion:240 Resp: 387591

Ion Ratio Lower Upper

240 100

120 12.4 9.5 14.3




236 25.8 20.7 31.1



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 42.183 ng

RT: 12.84 min Scan# 1829

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26


Tgt Ion:184 Resp: 465929

Ion Ratio Lower Upper

184 100

185 17.0 12.6 18.8



183 12.0 9.3 13.9




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E









#77
Pvrene
Concen: 38.559 ng
RT: 12.95 min Scan# 1847
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26




Instrument :
BNA_F
ClientSampleId :

Tot Ion:202 Resp: 1074350
Ion Ratio Lower Upper
202 100
200 21.0 17.4 26.2
203 17.6 14.3 21.5



#78
Terphenyl-d14
Concen: 74.448 ng
RT: 13.09 min Scan# 1870
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26

Tgt Ion:244 Resp: 1249721
Ion Ratio Lower Upper
244 100
212 6.7 5.4 8.0
122 11.4 11.0 16.4









#79
 Butylbenzylphthalate
 Concen: 38.404 ng
 RT: 13.55 min Scan# 1949
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:149 Resp: 504110
 Ion Ratio Lower Upper
 149 100
 91 70.9 56.8 85.2
 206 19.4 14.1 21.1



#80
 Benzo(a)anthracene
 Concen: 38.156 ng
 RT: 14.14 min Scan# 2050
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26

Tgt Ion:228 Resp: 925375
 Ion Ratio Lower Upper
 228 100
 226 27.7 22.6 33.8
 229 19.9 15.8 23.6









#81
 3,3'-Dichlorobenzidine
 Concen: 37.386 ng
 RT: 14.10 min Scan# 2043
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:252 Resp: 300117
 Ion Ratio Lower Upper
 252 100
 254 65.0 50.4 75.6
 126 12.5 11.3 16.9



#82
 Chrysene
 Concen: 37.738 ng
 RT: 14.19 min Scan# 2057
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26

Tgt Ion:228 Resp: 896416
 Ion Ratio Lower Upper
 228 100
 226 30.7 25.0 37.6
 229 19.5 16.2 24.4









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 37.640 ng
 RT: 14.10 min Scan# 2043
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26




Instrument :
 BNA_F
 ClientSampled :


Tot Ion:149 Resp: 638398
 Ion Ratio Lower Upper
 149 100
 167 26.2 21.3 31.9
 279 3.8 3.0 4.4



#84
 Di-n-octyl phthalate
 Concen: 38.075 ng
 RT: 14.73 min Scan# 2150
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26

Tgt Ion:149 Resp: 1110832
 Ion Ratio Lower Upper
 149 100
 167 1.5 1.2 1.8
 43 8.3 8.4 12.6#









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 35.508 ng
 RT: 17.29 min Scan# 2584
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
276	100		
138	27.6	25.0	37.6
277	25.3	20.1	30.1




Abundance Ion 276.00 (275.70 to 276.70): E
 Ion 138.00 (137.70 to 138.70): E
 Ion 277.00 (276.70 to 277.70): E






#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.70 min Scan# 2314
 Delta R.T. -0.00 min
 Lab File: BF104008.D
 Acq: 28 Mar 2018 14:26

Tgt Ion	Ratio	Lower	Upper
264	100		
260	23.5	19.2	28.8
265	22.3	17.5	26.3



Abundance Ion 264.00 (263.70 to 264.70): E
 Ion 260.00 (259.70 to 260.70): E
 Ion 265.00 (264.70 to 265.70): E









#87
Benzo(b)fluoranthene
Concen: 36.209 ng
RT: 15.24 min Scan# 2236
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26




Instrument :
BNA_F
ClientSampleId :

Tot Ion:252 Resp: 854296
Ion Ratio Lower Upper
252 100
253 22.4 17.0 25.6
125 11.6 9.0 13.6



#88
Benzo(k)fluoranthene
Concen: 39.640 ng
RT: 15.27 min Scan# 2242
Delta R.T. 0.01 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26

Tgt Ion:252 Resp: 880999
Ion Ratio Lower Upper
252 100
253 22.1 17.9 26.9
125 11.3 10.2 15.2









#89
Benzo(a)pyrene
Concen: 37.720 ng
RT: 15.63 min Scan# 2303
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26




Instrument :
BNA_F
ClientSampleId :




Tot Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.1	25.7
125	13.2	9.8	14.6



#90
Dibenzo(a,h)anthracene
Concen: 35.606 ng
RT: 17.30 min Scan# 2587
Delta R.T. -0.00 min
Lab File: BF104008.D
Acq: 28 Mar 2018 14:26

Tgt Ion	Ratio	Lower	Upper
278	100		
139	18.0	15.9	23.9
279	23.2	19.0	28.4





#91

Benzo(a,h,i)perylene

Concen: 34.928 ng

RT: 17.77 min Scan# 2666

Delta R.T. -0.00 min

Lab File: BF104008.D

Acq: 28 Mar 2018 14:26

Instrument :

BNA_F

ClientSampleId :

Tot Ion: 276 Resp: 707207

Ion Ratio Lower Upper

276 100

277 23.9 17.9 26.9

138 24.0 21.8 32.8

