

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032818\
 Data File : BF104038.D
 Acq On : 29 Mar 2018 7:19
 Operator : JU/SJ
 Sample : J2048-01DL 20X
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-108(4-5)DL

Manual Integrations
 APPROVED

Sohil
 3/29/2018 5:37:53 PM

Quant Time: Mar 29 14:03:37 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 28 13:41:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.96	152	98572	20.00	ng	0.00
21) Naphthalene-d8	8.25	136	408186	20.00	ng	0.00
38) Acenaphthene-d10	10.00	164	171946	20.00	ng	0.00
63) Phenanthrene-d10	11.50	188	248635	20.00	ng	0.00
75) Chrysene-d12	14.15	240	216878	20.00	ng	0.00
86) Perylene-d12	15.70	264	152045	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.70	112	21596m	3.49	ng	0.11
7) Phenol-d6	6.63	99	33809m	4.45	ng	0.03
23) Nitrobenzene-d5	7.54	82	18724	2.81	ng	0.00
41) 2,4,6-Tribromophenol	10.80	330	5907	3.09	ng	0.00
44) 2-Fluorobiphenyl	9.32	172	49077	4.50	ng	0.00
78) Terphenyl-d14	13.08	244	36601	3.90	ng	0.00
Target Compounds						
31) Naphthalene	8.27	128	852120	42.731	ng	100
37) 2-Methylnaphthalene	8.96	142	267087	20.333	ng	98
70) Phenanthrene	11.53	178	58179	4.322	ng	99
74) Fluoranthene	12.72	202	51865	3.625	ng	97
77) Pyrene	12.95	202	75200	4.823	ng	98
80) Benzo(a)anthracene	14.14	228	32891	2.424	ng	94
82) Chrysene	14.18	228	30445	2.291	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032818\
Data File : BF104038.D
Acq On : 29 Mar 2018 7:19
Operator : JU/SJ
Sample : J2048-01DL 20X
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-108(4-5)DL

Manual Integrations
APPROVED

Sohil
3/29/2018 5:37:53 PM

Quant Time: Mar 29 14:03:37 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 28 13:41:47 2018
Response via : Initial Calibration

