


Benzaldehyde

Response Ratio




$R = -1.26e-001 A^2 + 1.04e+000 A + 2.48e-002$
Coef of Det (r^2) = 0.998 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF032818.M
Calibration Table Last Updated: Wed Apr 04 18:06:58 2018


Hexachlorocyclopentadiene

Response Ratio



Resp Ratio = $3.10 \times 10^{-1} \times \text{Amt} - 8.34 \times 10^{-2}$
Coef of Det (r^2) = 0.999 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF032818.M
Calibration Table Last Updated: Wed Apr 04 18:06:58 2018



Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF032818.M
Calibration Table Last Updated: Wed Apr 04 18:06:58 2018