

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032818\
 Data File : BF104023.D
 Acq On : 28 Mar 2018 21:09
 Operator : JU/SJ
 Sample : J2067-01 2X
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RO-CV-RB47185RT

Quant Time: Mar 29 09:40:11 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 28 13:41:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.97	152	120160	20.00	ng	0.00
21) Naphthalene-d8	8.25	136	432087	20.00	ng	0.00
38) Acenaphthene-d10	10.04	164	22608	20.00	ng	0.04
63) Phenanthrene-d10	11.63	188	41488	20.00	ng	0.13
75) Chrysene-d12	14.21	240	149958	20.00	ng	0.05
86) Perylene-d12	15.72	264	236295	20.00	ng	0.02


System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.65	112	8231	1.09	ng	0.05
7) Phenol-d6	6.65	99	17266	1.86	ng	0.04
23) Nitrobenzene-d5	7.53	82	34479	4.88	ng	0.00
41) 2,4,6-Tribromophenol	10.89	330	3186	12.66	ng	0.08
44) 2-Fluorobiphenyl	9.35	172	82232	57.36	ng	0.03
78) Terphenyl-d14	13.17	244	18859	2.90	ng	0.08


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) Benzaldehyde	6.52	77	1180	Below Cal		# 1
27) 2,4-Dimethylphenol	7.90	122	11432	2.043	ng	# 17
35) Caprolactam	8.69	113	17379	9.376	ng	# 1
36) 4-Chloro-3-methylphenol	8.71	107	86984	14.069	ng	# 20
45) 1,1'-Biphenyl	9.46	154	19577	10.089	ng	# 24
47) 2-Nitroaniline	9.55	65	6078	13.926	ng	# 72
48) Acenaphthylene	9.91	152	8410	3.627	ng	# 1
49) Dimethylphthalate	9.76	163	4796	2.687	ng	# 3
50) 2,6-Dinitrotoluene	9.80	165	5733	14.929	ng	# 38
52) 3-Nitroaniline	9.96	138	4405	9.979	ng	# 28
54) Dibenzofuran	10.25	168	5458	2.786	ng	# 1
55) 4-Nitrophenol	10.26	139	1812	6.846	ng	# 1
57) Fluorene	10.60	166	15547	9.622	ng	# 1
60) 4-Chlorophenyl-phenylether	10.80	204	4279	5.766	ng	# 1
61) 4-Nitroaniline	10.67	138	1631	3.949	ng	# 12
62) Azobenzene	10.79	77	10909	7.055	ng	# 34
64) 4,6-Dinitro-2-methylphenol	10.78	198	6142	24.457	ng	# 1
65) n-Nitrosodiphenylamine	10.75	169	69434	49.417	ng	# 60
66) 4-Bromophenyl-phenylether	11.20	248	1190	2.681	ng	# 1
70) Phenanthrene	11.65	178	94798	42.204	ng	# 15
71) Anthracene	11.74	178	10745	4.580	ng	# 1
73) Di-n-butylphthalate	12.19	149	12377	4.640	ng	# 17
74) Fluoranthene	12.87	202	8589	3.597	ng	# 1
83) Bis(2-ethylhexyl)phthalate	14.13	149	27828	4.241	ng	# 51

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
BNA_F
ClientSampleId :
RO-CV-RB47185RT

Quant Time: Mar 29 09:40:11 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 28 13:41:47 2018
Response via : Initial Calibration









#1

1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.97 min Scan# 830
 Delta R.T. -0.00 min
 Lab File: BF104023.D
 Acq: 28 Mar 2018 21:09

Instrument :
 BNA_F
 ClientSampleId :
 RO-CV-RB47185RT




Tot Ion:152 Resp: 120160
 Ion Ratio Lower Upper
 152 100
 150 155.5 124.6 186.8
 115 59.4 50.1 75.1




#5

2-Fluorophenol
 Concen: 1.092 ng
 RT: 5.65 min Scan# 606
 Delta R.T. 0.05 min
 Lab File: BF104023.D
 Acq: 28 Mar 2018 21:09

Tgt Ion:112 Resp: 8231
 Ion Ratio Lower Upper
 112 100
 64 56.9 47.9 71.9
 63 23.9 23.7 35.5





#7

Phenol-d6

Concen: 1.863 ng

RT: 6.65 min Scan# 775

Delta R.T. 0.04 min

Lab File: BF104023.D

Acq: 28 Mar 2018 21:09

Instrument :

BNA_F

ClientSampleId :

RO-CV-RB47185RT


Tot Ion: 99 Resp: 17266

Ion Ratio Lower Upper

99 100

42 27.9 14.5 21.7#




71 50.4 25.4 38.0#



Abundance Ion 99.00 (98.70 to 99.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 71.00 (70.70 to 71.70): BF1



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.25 min Scan# 1048

Delta R.T. -0.00 min

Lab File: BF104023.D

Acq: 28 Mar 2018 21:09

Tgt Ion: 136 Resp: 432087


Ion Ratio Lower Upper

136 100

137 12.1 8.9 13.3

54 7.7 6.6 9.8

68 7.9 5.0 7.4#






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF1

Ion 68.00 (67.70 to 68.70): BF1









#23
 Nitrobenzene-d5
 Concen: 4.880 ng
 RT: 7.53 min Scan# 926
 Delta R.T. -0.00 min
 Lab File: BF104023.D
 Acq: 28 Mar 2018 21:09




Instrument :
 BNA_F
 ClientSampleId :
 RO-CV-RB47185RT


Tot Ion: 82 Resp: 34479
 Ion Ratio Lower Upper
 82 100
 128 40.6 36.1 54.1
 54 44.3 41.4 62.2



#27
 2,4-Dimethylphenol
 Concen: 2.043 ng
 RT: 7.90 min Scan# 988
 Delta R.T. -0.00 min
 Lab File: BF104023.D
 Acq: 28 Mar 2018 21:09

Tgt Ion: 122 Resp: 11432
 Ion Ratio Lower Upper
 122 100
 107 211.2 91.2 136.8#
 121 111.2 47.2 70.8#





#35

Caprolactam

Concen: 9.376 ng

RT: 8.69 min Scan# 1123

Delta R.T. -0.01 min

Lab File: BF104023.D

Acq: 28 Mar 2018 21:09

Instrument :

BNA_F

ClientSampleId :

RO-CV-RB47185RT


Tot Ion:113 Resp: 17379

Ion Ratio Lower Upper

113 100

55 1248.9 163.3 203.3#




56 427.6 115.7 155.7#



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BF1

Ion 56.00 (55.70 to 56.70): BF1



#36

4-Chloro-3-methylphenol

Concen: 14.069 ng

RT: 8.71 min Scan# 1126

Delta R.T. -0.09 min

Lab File: BF104023.D

Acq: 28 Mar 2018 21:09


Tgt Ion:107 Resp: 86984

Ion Ratio Lower Upper

107 100

144 0.0 20.5 30.7#



142 0.0 62.0 93.0#




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E





#38

Acenaphthene-d10

Concen: 20.000 ng

RT: 10.04 min Scan# 1352

Delta R.T. 0.04 min

Lab File: BF104023.D

Acq: 28 Mar 2018 21:09

Instrument :

BNA_F

ClientSampleId :

RO-CV-RB47185RT





Tot Ion:164 Resp: 22608

Ion Ratio Lower Upper

164 100

162 80.1 86.1 129.1#

160 43.3 38.3 57.5



#41

2,4,6-Tribromophenol

Concen: 12.656 ng

RT: 10.89 min Scan# 1496

Delta R.T. 0.08 min

Lab File: BF104023.D

Acq: 28 Mar 2018 21:09




Tgt Ion:330 Resp: 3186


Ion Ratio Lower Upper

330 100

332 66.0 77.0 115.6#

141 126.5 29.9 44.9#









#44
2-Fluorobiphenyl
Concen: 57.357 ng
RT: 9.35 min Scan# 1235
Delta R.T. 0.03 min
Lab File: BF104023.D
Acq: 28 Mar 2018 21:09

Instrument :
BNA_F
ClientSampleId :
RO-CV-RB47185RT

Tot Ion:172 Resp: 82232
Ion Ratio Lower Upper
172 100
171 36.5 29.7 44.5
170 22.2 19.6 29.4




Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E






#45
1,1'-Biphenyl
Concen: 10.089 ng
RT: 9.46 min Scan# 1254
Delta R.T. 0.04 min
Lab File: BF104023.D
Acq: 28 Mar 2018 21:09

Tgt Ion:154 Resp: 19577
Ion Ratio Lower Upper
154 100
153 98.1 21.3 61.3#
76 27.4 0.0 34.0



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 76.00 (75.70 to 76.70): BF1





#47

2-Nitroaniline

Concen: 13.926 ng

RT: 9.55 min Scan# 1269

Delta R.T. -0.01 min

Lab File: BF104023.D

Acq: 28 Mar 2018 21:09

Instrument :

BNA_F

ClientSampleId :

RO-CV-RB47185RT





Tot Ion: 65 Resp: 6078

Ion Ratio Lower Upper

65 100

92 81.6 55.8 83.6

138 75.2 91.2 136.8#



#48

Acenaphthylene

Concen: 3.627 ng

RT: 9.91 min Scan# 1330

Delta R.T. 0.04 min

Lab File: BF104023.D

Acq: 28 Mar 2018 21:09




Tgt Ion: 152 Resp: 8410


Ion Ratio Lower Upper

152 100

151 168.2 16.3 24.5#

153 66.2 10.7 16.1#





#49

Dimethylphthalate

Concen: 2.687 ng

RT: 9.76 min Scan# 1305

Delta R.T. 0.04 min

Lab File: BF104023.D

Acq: 28 Mar 2018 21:09

Instrument :

BNA_F

ClientSampleId :

RO-CV-RB47185RT





Tot Ion:163 Resp: 4796

Ion Ratio Lower Upper

163 100

194 71.0 2.7 4.1#

164 33.0 8.2 12.2#



#50

2,6-Dinitrotoluene

Concen: 14.929 ng

RT: 9.80 min Scan# 1312

Delta R.T. 0.01 min

Lab File: BF104023.D

Acq: 28 Mar 2018 21:09




Tgt Ion:165 Resp: 5733


Ion Ratio Lower Upper

165 100

63 13.2 44.7 67.1#

89 7.5 44.0 66.0#





#52

3-Nitroaniline

Concen: 9.979 ng

RT: 9.96 min Scan# 1338

Delta R.T. -0.02 min

Lab File: BF104023.D

Acq: 28 Mar 2018 21:09

Instrument :

BNA_F

ClientSampleId :

RO-CV-RB47185RT





Tot Ion:138 Resp: 4405

Ion Ratio Lower Upper

138 100

108 100.7 9.4 14.0#

92 52.6 89.4 134.2#



#54

Dibenzofuran

Concen: 2.786 ng

RT: 10.25 min Scan# 1388

Delta R.T. 0.04 min

Lab File: BF104023.D

Acq: 28 Mar 2018 21:09




Tgt Ion:168 Resp: 5458


Ion Ratio Lower Upper

168 100

139 378.6 30.1 45.1#

169 169.6 10.9 16.3#





#55

4-Nitrophenol

Concen: 6.846 ng

RT: 10.26 min Scan# 1389

Delta R.T. 0.11 min

Lab File: BF104023.D

Acq: 28 Mar 2018 21:09

Instrument :

BNA_F

ClientSampleId :

RO-CV-RB47185RT





Tot Ion:139 Resp: 1812

Ion Ratio Lower Upper

139 100

109 656.6 48.4 88.4#

65 125.0 72.6 112.6#



#57

Fluorene

Concen: 9.622 ng

RT: 10.60 min Scan# 1448

Delta R.T. 0.05 min

Lab File: BF104023.D

Acq: 28 Mar 2018 21:09




Tgt Ion:166 Resp: 15547


Ion Ratio Lower Upper

166 100

165 237.6 79.8 119.8#

167 93.8 10.6 16.0#









#60
4-Chlorophenyl-phenylether
Concen: 5.766 ng
RT: 10.80 min Scan# 1481
Delta R.T. 0.25 min
Lab File: BF104023.D
Acq: 28 Mar 2018 21:09




Instrument :
BNA_F
ClientSampleId :
RO-CV-RB47185RT


Tot Ion:204 Resp: 4279
Ion Ratio Lower Upper
204 100
206 68.1 26.5 39.7#
141 358.6 54.6 81.8#



#61
4-Nitroaniline
Concen: 3.949 ng
RT: 10.67 min Scan# 1460
Delta R.T. 0.08 min
Lab File: BF104023.D
Acq: 28 Mar 2018 21:09

Tgt Ion:138 Resp: 1631
Ion Ratio Lower Upper
138 100
92 103.2 30.2 70.2#
108 153.0 52.5 92.5#





#62

Azobenzene

Concen: 7.055 ng

RT: 10.79 min Scan# 1479

Delta R.T. 0.08 min

Lab File: BF104023.D

Acq: 28 Mar 2018 21:09

Instrument :

BNA_F

ClientSampleId :

RO-CV-RB47185RT

Tot Ion: 77 Resp: 10909


Ion Ratio Lower Upper

77 100

182 13.4 1.7 41.7

105 104.7 3.0 43.0#

51 17.9 9.9 49.9






Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 182.00 (181.70 to 182.70): BF1

Ion 105.00 (104.70 to 105.70): BF1

Ion 51.00 (50.70 to 51.70): BF1



#63

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.63 min Scan# 1622

Delta R.T. 0.13 min

Lab File: BF104023.D

Acq: 28 Mar 2018 21:09


Tgt Ion: 188 Resp: 41488

Ion Ratio Lower Upper

188 100

94 199.5 8.5 12.7#



80 214.0 9.2 13.8#




Abundance Ion 188.00 (187.70 to 188.70): BF1

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1









#64
 4,6-Dinitro-2-methylphenol
 Concen: 24.457 ng
 RT: 10.78 min Scan# 1478
 Delta R.T. 0.17 min
 Lab File: BF104023.D
 Acq: 28 Mar 2018 21:09




Instrument :
 BNA_F
 ClientSampleId :
 RO-CV-RB47185RT


Tot Ion:198 Resp: 6142
 Ion Ratio Lower Upper
 198 100
 51 201.6 37.7 77.7#
 105 1147.7 36.3 76.3#



#65
 n-Nitrosodiphenylamine
 Concen: 49.417 ng
 RT: 10.75 min Scan# 1473
 Delta R.T. 0.08 min
 Lab File: BF104023.D
 Acq: 28 Mar 2018 21:09

Tgt Ion:169 Resp: 69434
 Ion Ratio Lower Upper
 169 100
 168 23.8 54.4 81.6#
 167 45.0 29.8 44.6#









#66
4-Bromophenyl-phenylether
Concen: 2.681 ng
RT: 11.20 min Scan# 1550
Delta R.T. 0.17 min
Lab File: BF104023.D
Acq: 28 Mar 2018 21:09




Instrument :
BNA_F
ClientSampleId :
RO-CV-RB47185RT


Tot Ion:248 Resp: 1190
Ion Ratio Lower Upper
248 100
250 108.0 76.9 115.3
141 701.7 74.4 111.6#



#70
Phenanthrene
Concen: 42.204 ng
RT: 11.65 min Scan# 1626
Delta R.T. 0.12 min
Lab File: BF104023.D
Acq: 28 Mar 2018 21:09

Tgt Ion:178 Resp: 94798
Ion Ratio Lower Upper
178 100
176 32.5 15.8 23.8#
179 82.7 13.0 19.6#





#71

Anthracene

Concen: 4.580 ng

RT: 11.74 min Scan# 1642

Delta R.T. 0.16 min

Lab File: BF104023.D

Acq: 28 Mar 2018 21:09

Instrument :

BNA_F

ClientSampleId :

RO-CV-RB47185RT





Tot Ion:178 Resp: 10745

Ion Ratio Lower Upper

178 100

176 40.8 15.4 23.2#

179 204.5 12.6 19.0#



#73

Di-n-butylphthalate

Concen: 4.640 ng

RT: 12.19 min Scan# 1717

Delta R.T. 0.14 min

Lab File: BF104023.D

Acq: 28 Mar 2018 21:09




Tgt Ion:149 Resp: 12377


Ion Ratio Lower Upper

149 100

150 46.6 7.8 11.6#

104 21.1 3.8 5.8#





#74

Fluoranthene

Concen: 3.597 ng

RT: 12.87 min Scan# 1833

Delta R.T. 0.15 min

Lab File: BF104023.D

Acq: 28 Mar 2018 21:09


Instrument :

BNA_F

ClientSampleId :

RO-CV-RB47185RT




Tot Ion:	202	Resp:	8589
Ion	Ratio	Lower	Upper
202	100		
101	23.3	0.0	34.1
203	174.9	0.0	38.1#



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#75

Chrysene-d12

Concen: 20.000 ng


RT: 14.21 min Scan# 2061

Delta R.T. 0.05 min

Lab File: BF104023.D

Acq: 28 Mar 2018 21:09



Tgt Ion:	240	Resp:	149958
Ion	Ratio	Lower	Upper
240	100		
120	21.4	9.5	14.3#
236	28.8	20.7	31.1




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#78

Terphenyl-d14

Concen: 2.904 ng

RT: 13.17 min Scan# 1884

Delta R.T. 0.08 min

Lab File: BF104023.D

Acq: 28 Mar 2018 21:09

Instrument :

BNA_F

ClientSampleId :

RO-CV-RB47185RT





Tot Ion:244 Resp: 18859

Ion Ratio Lower Upper

244 100

212 27.3 5.4 8.0#

122 99.1 11.0 16.4#



#83

Bis(2-ethylhexyl)phthalate

Concen: 4.241 ng

RT: 14.13 min Scan# 2048

Delta R.T. 0.03 min

Lab File: BF104023.D

Acq: 28 Mar 2018 21:09




Tgt Ion:149 Resp: 27828




Ion Ratio Lower Upper

149 100

167 53.7 21.3 31.9#

279 11.2 3.0 4.4#





#86
Pervlene-d12
Concen: 20.000 ng
RT: 15.72 min Scan# 2318
Delta R.T. 0.02 min
Lab File: BF104023.D
Acq: 28 Mar 2018 21:09

Instrument :
BNA_F
ClientSampleId :
RO-CV-RB47185RT

Tot Ion	Ratio	Lower	Upper
264	100		
260	23.8	19.2	28.8
265	22.2	17.5	26.3

