

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032822\
 Data File : BF127538.D
 Acq On : 28 Mar 2022 13:07
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 29 07:15:15 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 29 07:13:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2	1,4-Dioxane	0.550	0.695	-26.4#	144	0.00
3	Pyridine	1.553	1.709	-10.0	116	0.00
4	n-Nitrosodimethylamine	0.853	0.957	-12.2	120	0.00
5 S	2-Fluorophenol	1.253	1.351	-7.8	115	0.00
6	Aniline	2.044	2.206	-7.9	113	0.00
7 S	Phenol-d6	1.713	1.820	-6.2	114	0.00
8	2-Chlorophenol	1.280	1.378	-7.7	114	0.00
9	Benzaldehyde	0.965	0.937	2.9	109	0.00
10 C	Phenol	1.865	2.022	-8.4	114	0.00
11	bis(2-Chloroethyl)ether	1.365	1.502	-10.0	117	0.00
12	1,3-Dichlorobenzene	1.440	1.510	-4.9	113	0.00
13 C	1,4-Dichlorobenzene	1.449	1.525	-5.2	112	0.00
14	1,2-Dichlorobenzene	1.411	1.417	-0.4	113	0.00
15	Benzyl Alcohol	1.245	1.302	-4.6	109	0.00
16	2,2'-oxybis(1-Chloropropane	2.456	2.684	-9.3	116	0.00
17	2-Methylphenol	1.081	1.156	-6.9	114	0.00
18	Hexachloroethane	0.530	0.569	-7.4	115	0.00
19 P	n-Nitroso-di-n-propylamine	1.027	1.079	-5.1	115	0.00
20	3+4-Methylphenols	1.333	1.406	-5.5	113	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
22	Acetophenone	0.511	0.536	-4.9	113	0.00
23 S	Nitrobenzene-d5	0.468	0.499	-6.6	113	0.00
24	Nitrobenzene	0.446	0.488	-9.4	114	0.00
25	Isophorone	0.745	0.836	-12.2	115	0.00
26 C	2-Nitrophenol	0.187	0.206	-10.2	114	0.00
27	2,4-Dimethylphenol	0.282	0.308	-9.2	112	0.00
28	bis(2-Chloroethoxy)methane	0.484	0.536	-10.7	116	0.00
29 C	2,4-Dichlorophenol	0.320	0.353	-10.3	115	0.00
30	1,2,4-Trichlorobenzene	0.372	0.394	-5.9	112	0.00
31	Naphthalene	1.062	1.136	-7.0	113	0.00
32	Benzoic acid	0.140	0.157	-12.1	105	0.00
33	4-Chloroaniline	0.411	0.452	-10.0	115	0.00
34 C	Hexachlorobutadiene	0.236	0.255	-8.1	113	0.00
35	Caprolactam	0.099	0.108	-9.1	110	0.00
36 C	4-Chloro-3-methylphenol	0.347	0.378	-8.9	114	0.00
37	2-Methylnaphthalene	0.683	0.734	-7.5	113	0.00
38	1-Methylnaphthalene	0.675	0.711	-5.3	112	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	0.636	0.675	-6.1	111	0.00
41 P	Hexachlorocyclopentadiene	0.147	0.142	3.4	96	0.00
42 S	2,4,6-Tribromophenol	0.213	0.214	-0.5	105	0.00
43 C	2,4,6-Trichlorophenol	0.384	0.424	-10.4	112	0.00
44	2,4,5-Trichlorophenol	0.396	0.436	-10.1	107	0.00
45 S	2-Fluorobiphenyl	1.370	1.429	-4.3	112	0.00
46	1,1'-Biphenyl	1.511	1.590	-5.2	111	0.00
47	2-Chloronaphthalene	1.189	1.262	-6.1	111	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032822\
 Data File : BF127538.D
 Acq On : 28 Mar 2022 13:07
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 29 07:15:15 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 29 07:13:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.456	0.498	-9.2	112	0.00
49	Acenaphthylene	1.798	1.876	-4.3	108	0.00
50	Dimethylphthalate	1.391	1.497	-7.6	112	0.00
51	2,6-Dinitrotoluene	0.292	0.314	-7.5	111	0.00
52 C	Acenaphthene	1.049	1.117	-6.5	109	0.00
53	3-Nitroaniline	0.320	0.343	-7.2	110	0.00
54 P	2,4-Dinitrophenol	0.114	0.120	-5.3	95	0.00
55	Dibenzofuran	1.710	1.689	1.2	108	0.00
56 P	4-Nitrophenol	0.163	0.150	8.0	87	0.00
57	2,4-Dinitrotoluene	0.376	0.385	-2.4	106	0.00
58	Fluorene	1.353	1.338	1.1	110	0.00
59	2,3,4,6-Tetrachlorophenol	0.350	0.356	-1.7	107	0.00
60	Diethylphthalate	1.271	1.343	-5.7	112	0.00
61	4-Chlorophenyl-phenylether	0.719	0.732	-1.8	113	0.00
62	4-Nitroaniline	0.300	0.312	-4.0	108	0.00
63	Azobenzene	1.500	1.615	-7.7	113	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
65	4,6-Dinitro-2-methylphenol	0.123	0.140	-13.8	103	0.00
66 c	n-Nitrosodiphenylamine	0.609	0.680	-11.7	110	0.00
67	4-Bromophenyl-phenylether	0.234	0.264	-12.8	109	0.00
68	Hexachlorobenzene	0.255	0.282	-10.6	108	0.00
69	Atrazine	0.203	0.221	-8.9	102	0.00
70 C	Pentachlorophenol	0.082	0.084	-2.4	89	0.00
71	Phenanthrene	1.043	1.122	-7.6	106	0.00
72	Anthracene	1.035	1.120	-8.2	106	0.00
73	Carbazole	0.990	1.045	-5.6	103	0.00
74	Di-n-butylphthalate	1.024	1.184	-15.6	112	0.00
75 C	Fluoranthene	1.147	1.198	-4.4	102	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	85	0.00
77	Benzidine	0.477	0.458	4.0	91	0.00
78	Pyrene	1.716	1.920	-11.9	100	0.00
79 S	Terphenyl-d14	1.222	1.365	-11.7	100	0.00
80	Butylbenzylphthalate	0.565	0.676	-19.6	101	0.00
81	Benzo(a)anthracene	1.337	1.443	-7.9	95	0.00
82	3,3'-Dichlorobenzidine	0.393	0.429	-9.2	96	0.00
83	Chrysene	1.267	1.376	-8.6	96	0.00
84	Bis(2-ethylhexyl)phthalate	0.653	0.797	-22.1	103	0.00
85 c	Di-n-octyl phthalate	0.869	1.215	-39.8#	119	0.00
86 I	Perylene-d12	1.000	1.000	0.0	106	0.00
87	Indeno(1,2,3-cd)pyrene	1.353	1.455	-7.5	108	0.00
88	Benzo(b)fluoranthene	1.261	1.348	-6.9	113	0.00
89	Benzo(k)fluoranthene	1.307	1.317	-0.8	116	0.00
90 C	Benzo(a)pyrene	1.035	1.136	-9.8	118	0.00
91	Dibenzo(a,h)anthracene	1.117	1.220	-9.2	109	0.00
92	Benzo(g,h,i)perylene	1.155	1.220	-5.6	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032822\
Data File : BF127538.D
Acq On : 28 Mar 2022 13:07
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 29 07:15:15 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 29 07:13:31 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1