

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032917\
 Data File : BF094026.D
 Acq On : 29 Mar 2017 18:49
 Operator : SJ/MA
 Sample : I2443-15
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 E2-BB-CC-7-8-BASE

Quant Time: Mar 30 03:53:04 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 27 16:10:49 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.92	152	159658	20.00	ng	-0.04
21) Naphthalene-d8	7.42	136	666440	20.00	ng	-0.04
38) Acenaphthene-d10	9.56	164	313807	20.00	ng	-0.04
63) Phenanthrene-d10	11.39	188	551065	20.00	ng	-0.04
75) Chrysene-d12	14.64	240	452576	20.00	ng	-0.04
86) Perylene-d12	16.27	264	328610	20.00	ng	-0.04
System Monitoring Compounds						
5) 2-Fluorophenol	4.46	112	1257998	133.08	ng	-0.02
7) Phenol-d6	5.52	99	1564146	133.76	ng	-0.03
23) Nitrobenzene-d5	6.57	82	985232	84.37	ng	-0.04
41) 2,4,6-Tribromophenol	10.54	330	338986	136.70	ng	-0.04
44) 2-Fluorobiphenyl	8.75	172	1485933	72.22	ng	-0.04
78) Terphenyl-d14	13.37	244	1224765	60.66	ng	-0.04
Target Compounds						
49) Dimethylphthalate	9.24	163	120758	5.41	ng	98
70) Phenanthrene	11.41	178	103998	3.39	ng	96
74) Fluoranthene	12.89	202	104894	3.34	ng	96
77) Pyrene	13.15	202	87313	2.66	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032917\
Data File : BF094026.D
Acq On : 29 Mar 2017 18:49
Operator : SJ/MA
Sample : I2443-15
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
E2-BB-CC-7-8-BASE

Quant Time: Mar 30 03:53:04 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 27 16:10:49 2017
Response via : Initial Calibration

