

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032917\
 Data File : BF094035.D
 Acq On : 29 Mar 2017 22:56
 Operator : SJ/MA
 Sample : I2382-01
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PX-11

Quant Time: Mar 30 04:22:27 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 27 16:10:49 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.92	152	148081	20.00	ng	-0.04
21) Naphthalene-d8	7.42	136	555289	20.00	ng	-0.04
38) Acenaphthene-d10	9.57	164	204522	20.00	ng	-0.03
63) Phenanthrene-d10	11.41	188	228934	20.00	ng	-0.01
75) Chrysene-d12	14.73	240	88385	20.00	ng	0.05
86) Perylene-d12	16.39	264	77923	20.00	ng	0.08
System Monitoring Compounds						
5) 2-Fluorophenol	4.45	112	1218822	139.02	ng	-0.02
7) Phenol-d6	5.52	99	1551080	143.01	ng	-0.03
23) Nitrobenzene-d5	6.57	82	934314	96.02	ng	-0.04
41) 2,4,6-Tribromophenol	10.55	330	226729	140.29	ng	-0.03
44) 2-Fluorobiphenyl	8.75	172	1449562	108.10	ng	-0.03
78) Terphenyl-d14	13.43	244	393319	99.75	ng	0.03
Target Compounds						
49) Dimethylphthalate	9.25	163	84235	5.79	ng	# 98
51) Acenaphthene	9.60	154	152446	12.17	ng	# 99
57) Fluorene	10.25	166	59512	4.21	ng	# 81
70) Phenanthrene	11.44	178	197048	15.47	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032917\
Data File : BF094035.D
Acq On : 29 Mar 2017 22:56
Operator : SJ/MA
Sample : I2382-01
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PX-11

Quant Time: Mar 30 04:22:27 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 27 16:10:49 2017
Response via : Initial Calibration

