

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032918\
 Data File : BF104084.D
 Acq On : 30 Mar 2018 7:02
 Operator : JU/SJ
 Sample : J2113-01
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 032618-TIDO-E-D-B

Manual Integrations
 APPROVED

Sohil
 3/30/2018 6:03:24 PM

Quant Time: Mar 30 10:53:19 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 28 13:41:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.97	152	105568	20.00	ng	0.00
21) Naphthalene-d8	8.25	136	435429	20.00	ng	0.00
38) Acenaphthene-d10	10.00	164	174832	20.00	ng	0.00
63) Phenanthrene-d10	11.50	188	250634	20.00	ng	0.00
75) Chrysene-d12	14.16	240	238838	20.00	ng	0.00
86) Perylene-d12	15.69	264	218915	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	224950	33.98	ng	0.00
7) Phenol-d6	6.60	99	174222	21.39	ng	0.00
23) Nitrobenzene-d5	7.53	82	676136	94.96	ng	0.00
41) 2,4,6-Tribromophenol	10.80	330	141939	72.91	ng	0.00
44) 2-Fluorobiphenyl	9.32	172	1236958	111.57	ng	0.00
78) Terphenyl-d14	13.09	244	810946	78.40	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.71	163	28491	2.064	ng	Qvalue 98
70) Phenanthrene	11.53	178	40010m	2.949	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032918\
Data File : BF104084.D
Acq On : 30 Mar 2018 7:02
Operator : JU/SJ
Sample : J2113-01
Misc :
ALS Vial : 40 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
032618-TIDO-E-D-B

Manual Integrations
APPROVED

Sohil
3/30/2018 6:03:24 PM

Quant Time: Mar 30 10:53:19 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 28 13:41:47 2018
Response via : Initial Calibration

