

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032922\
 Data File : BF127557.D
 Acq On : 29 Mar 2022 10:41
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 29 15:52:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 29 07:13:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	125	0.00
2	1,4-Dioxane	0.550	0.635	-15.5	156#	0.00
3	Pyridine	1.553	1.603	-3.2	130	0.00
4	n-Nitrosodimethylamine	0.853	0.869	-1.9	129	0.00
5 S	2-Fluorophenol	1.253	1.227	2.1	125	0.00
6	Aniline	2.044	2.032	0.6	124	0.00
7 S	Phenol-d6	1.713	1.687	1.5	125	0.00
8	2-Chlorophenol	1.280	1.277	0.2	125	0.00
9	Benzaldehyde	0.965	0.870	9.8	121	0.00
10 C	Phenol	1.865	1.858	0.4	125	0.00
11	bis(2-Chloroethyl)ether	1.365	1.403	-2.8	130	0.00
12	1,3-Dichlorobenzene	1.440	1.385	3.8	123	0.00
13 C	1,4-Dichlorobenzene	1.449	1.412	2.6	124	0.00
14	1,2-Dichlorobenzene	1.411	1.298	8.0	124	0.00
15	Benzyl Alcohol	1.245	1.230	1.2	123	0.00
16	2,2'-oxybis(1-Chloropropane	2.456	2.493	-1.5	128	0.00
17	2-Methylphenol	1.081	1.076	0.5	127	0.00
18	Hexachloroethane	0.530	0.525	0.9	127	0.00
19 P	n-Nitroso-di-n-propylamine	1.027	1.020	0.7	129	0.00
20	3+4-Methylphenols	1.333	1.299	2.6	125	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	122	0.00
22	Acetophenone	0.511	0.508	0.6	127	0.00
23 S	Nitrobenzene-d5	0.468	0.475	-1.5	127	0.00
24	Nitrobenzene	0.446	0.461	-3.4	128	0.00
25	Isophorone	0.745	0.814	-9.3	133	0.00
26 C	2-Nitrophenol	0.187	0.195	-4.3	128	0.00
27	2,4-Dimethylphenol	0.282	0.295	-4.6	127	0.00
28	bis(2-Chloroethoxy)methane	0.484	0.522	-7.9	134	0.00
29 C	2,4-Dichlorophenol	0.320	0.331	-3.4	128	0.00
30	1,2,4-Trichlorobenzene	0.372	0.374	-0.5	126	0.00
31	Naphthalene	1.062	1.080	-1.7	127	0.00
32	Benzoic acid	0.140	0.158	-12.9	125	0.01
33	4-Chloroaniline	0.411	0.422	-2.7	127	0.00
34 C	Hexachlorobutadiene	0.236	0.238	-0.8	126	0.00
35	Caprolactam	0.099	0.109	-10.1	131	0.00
36 C	4-Chloro-3-methylphenol	0.347	0.358	-3.2	128	0.00
37	2-Methylnaphthalene	0.683	0.704	-3.1	128	0.00
38	1-Methylnaphthalene	0.675	0.680	-0.7	128	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	122	0.00
40	1,2,4,5-Tetrachlorobenzene	0.636	0.635	0.2	127	0.00
41 P	Hexachlorocyclopentadiene	0.147	0.128	12.9	106	0.00
42 S	2,4,6-Tribromophenol	0.213	0.206	3.3	124	0.00
43 C	2,4,6-Trichlorophenol	0.384	0.400	-4.2	129	0.00
44	2,4,5-Trichlorophenol	0.396	0.406	-2.5	122	0.00
45 S	2-Fluorobiphenyl	1.370	1.334	2.6	127	0.00
46	1,1'-Biphenyl	1.511	1.490	1.4	128	0.00
47	2-Chloronaphthalene	1.189	1.184	0.4	127	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032922\
 Data File : BF127557.D
 Acq On : 29 Mar 2022 10:41
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 29 15:52:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 29 07:13:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.456	0.468	-2.6	129	0.00
49	Acenaphthylene	1.798	1.730	3.8	121	0.00
50	Dimethylphthalate	1.391	1.440	-3.5	132	0.00
51	2,6-Dinitrotoluene	0.292	0.299	-2.4	129	0.00
52 C	Acenaphthene	1.049	1.046	0.3	125	0.00
53	3-Nitroaniline	0.320	0.319	0.3	125	0.00
54 P	2,4-Dinitrophenol	0.114	0.122	-7.0	118	0.00
55	Dibenzofuran	1.710	1.556	9.0	122	0.00
56 P	4-Nitrophenol	0.163	0.145	11.0	103	0.00
57	2,4-Dinitrotoluene	0.376	0.359	4.5	120	0.00
58	Fluorene	1.353	1.243	8.1	125	0.00
59	2,3,4,6-Tetrachlorophenol	0.350	0.347	0.9	127	0.00
60	Diethylphthalate	1.271	1.297	-2.0	132	0.00
61	4-Chlorophenyl-phenylether	0.719	0.694	3.5	131	0.00
62	4-Nitroaniline	0.300	0.290	3.3	123	0.00
63	Azobenzene	1.500	1.510	-0.7	129	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
65	4,6-Dinitro-2-methylphenol	0.123	0.136	-10.6	124	0.00
66 c	n-Nitrosodiphenylamine	0.609	0.637	-4.6	127	0.00
67	4-Bromophenyl-phenylether	0.234	0.245	-4.7	126	0.00
68	Hexachlorobenzene	0.255	0.266	-4.3	125	0.00
69	Atrazine	0.203	0.210	-3.4	120	0.00
70 C	Pentachlorophenol	0.082	0.088	-7.3	115	0.00
71	Phenanthrene	1.043	1.057	-1.3	123	0.00
72	Anthracene	1.035	1.045	-1.0	123	0.00
73	Carbazole	0.990	0.980	1.0	119	0.00
74	Di-n-butylphthalate	1.024	1.146	-11.9	133	0.00
75 C	Fluoranthene	1.147	1.150	-0.3	121	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	114	0.00
77	Benzydine	0.477	0.436	8.6	116	0.00
78	Pyrene	1.716	1.712	0.2	120	0.00
79 S	Terphenyl-d14	1.222	1.235	-1.1	122	0.00
80	Butylbenzylphthalate	0.565	0.665	-17.7	134	0.00
81	Benzo(a)anthracene	1.337	1.356	-1.4	120	0.00
82	3,3'-Dichlorobenzidine	0.393	0.385	2.0	116	0.00
83	Chrysene	1.267	1.294	-2.1	121	0.00
84	Bis(2-ethylhexyl)phthalate	0.653	0.833	-27.6#	145	0.00
85 c	Di-n-octyl phthalate	0.869	1.267	-45.8#	167#	0.00
86 I	Perylene-d12	1.000	1.000	0.0	125	0.00
87	Indeno(1,2,3-cd)pyrene	1.353	1.271	6.1	111	0.00
88	Benzo(b)fluoranthene	1.261	1.289	-2.2	127	0.00
89	Benzo(k)fluoranthene	1.307	1.356	-3.7	140	0.00
90 C	Benzo(a)pyrene	1.035	1.079	-4.3	132	0.00
91	Dibenzo(a,h)anthracene	1.117	1.058	5.3	111	0.00
92	Benzo(g,h,i)perylene	1.155	1.053	8.8	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032922\
Data File : BF127557.D
Acq On : 29 Mar 2022 10:41
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 29 15:52:59 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 29 07:13:31 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1