


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032922\
 Data File : BF127556.D
 Acq On : 29 Mar 2022 09:35
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Mar 29 07:13:31 2022






AutoFind: Scans 1583, 1584, 1585; Background Corrected with Scan 1577

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	45.1	72859	PASS
68	69	0.00	2	1.8	1084	PASS
69	198	0.00	100	38.1	61491	PASS
70	69	0.00	2	0.5	302	PASS
127	198	10	80	44.7	72197	PASS
197	198	0.00	2	0.5	781	PASS
198	198	100	100	100.0	161504	PASS
199	198	5	9	6.5	10564	PASS
275	198	10	60	28.2	45475	PASS
365	198	1	100	3.8	6072	PASS
441	198	0.01	100	15.7	25308	PASS
442	442	50	100	100.0	162645	PASS
443	442	15	24	18.3	29749	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032922\
Data File : BF127556.D
Acq On : 29 Mar 2022 09:35
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Mar 29 16:02:11 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 29 07:13:31 2022
Response via : Initial Calibration



TIC: BF127556.D\data.ms

(70) Pentachlorophenol (C)

11.175min (+ 0.000) 50906.66 ng




response 211962

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.00	65.39
264.00	60.60	63.59
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032922\
Data File : BF127556.D
Acq On : 29 Mar 2022 09:35
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Mar 29 16:02:11 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 29 07:13:31 2022
Response via : Initial Calibration



TIC: BF127556.D\data.ms

(77) Benzidine

12.710min (+ 0.000) 13926.91 ng m

response 916188

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.20	14.47
183.00	11.70	12.30
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/29/2022	BNA_F	<u>BF127556.D</u>
Compound Name	Response	Retention Time
DDT	433609	13.52
DDD	1188	13.22
DDE	778	12.88
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
1966	435575	0.45

Instrument :
BNA_F
ClientSampleId :
DFTPP