

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF033015\
 Data File : BF078092.D
 Acq On : 30 Mar 2015 20:16
 Operator : TP/IZ
 Sample : G1708-01MSD
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FK-GF-01-013-AMSD

Quant Time: Mar 31 00:43:02 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 31 01:02:18 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.06	152	30648	20.00	ng	0.00
21) Naphthalene-d8	8.64	136	125601	20.00	ng	0.00
38) Acenaphthene-d10	10.81	164	62405	20.00	ng	0.00
63) Phenanthrene-d10	12.65	188	119212	20.00	ng	0.00
75) Chrysene-d12	15.95	240	129924	20.00	ng	0.01
86) Perylene-d12	17.79	264	125192	20.00	ng	0.09

System Monitoring Compounds

5) 2-Fluorophenol	5.34	112	222302	126.61	ng	0.00
7) Phenol-d6	6.65	99	279083	128.65	ng	0.00
23) Nitrobenzene-d5	7.76	82	161304	84.18	ng	0.00
41) 2,4,6-Tribromophenol	11.79	330	72621	121.63	ng	0.00
44) 2-Fluorobiphenyl	9.98	172	320348	75.44	ng	0.00
78) Terphenyl-d14	14.65	244	371336	69.81	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	1.92	88	26737	33.54	ng	# 100
3) Pyridine	2.58	79	67480	31.50	ng	95
4) n-Nitrosodimethylamine	2.51	42	28847	30.93	ng	# 97
6) Aniline	6.65	93	84247	27.15	ng	# 77
8) 2-Chlorophenol	6.80	128	92853	44.43	ng	98
9) Benzaldehyde	6.51	77	2835	3.19	ng	92
10) Phenol	6.67	94	110735	43.18	ng	# 69
11) bis(2-Chloroethyl)ether	6.76	93	84918	44.21	ng	98
12) 1,3-Dichlorobenzene	6.98	146	93944	40.41	ng	# 96
13) 1,4-Dichlorobenzene	7.08	146	95703	39.62	ng	96
14) 1,2-Dichlorobenzene	7.26	146	92042	41.57	ng	96
15) Benzyl Alcohol	7.25	79	70610	41.70	ng	95
16) 2,2'-oxybis(1-Chloropropan	7.44	45	109729	42.39	ng	99
17) 2-Methylphenol	7.41	107	73707	43.32	ng	97
18) Hexachloroethane	7.68	117	32172	40.61	ng	# 90
19) n-Nitroso-di-n-propylamine	7.58	70	60799	40.51	ng	99
20) 3+4-Methylphenols	7.61	107	93612	41.93	ng	96
22) Acetophenone	7.57	105	123187	44.30	ng	97
24) Nitrobenzene	7.78	77	86265	41.79	ng	96
25) Isophorone	8.08	82	160048	41.36	ng	# 89
26) 2-Nitrophenol	8.18	139	42827	42.38	ng	98
27) 2,4-Dimethylphenol	8.26	122	72192	39.21	ng	99
28) bis(2-Chloroethoxy)methane	8.37	93	102202	43.51	ng	97
29) 2,4-Dichlorophenol	8.48	162	75955	43.28	ng	92
30) 1,2,4-Trichlorobenzene	8.57	180	77775	39.61	ng	# 95
31) Naphthalene	8.66	128	264344	40.98	ng	100
32) Benzoic acid	8.36	122	32200	32.43	ng	98
33) 4-Chloroaniline	8.75	127	78647	30.04	ng	99
34) Hexachlorobutadiene	8.82	225	43725	39.29	ng	98
35) Caprolactam	9.16	113	22157	43.20	ng	97
36) 4-Chloro-3-methylphenol	9.37	107	78228	44.30	ng	94
37) 2-Methylnaphthalene	9.53	142	178845	41.27	ng	98
39) 1,2,4,5-Tetrachlorobenzene	9.73	216	74834	40.21	ng	# 100
40) Hexachlorocyclopentadiene	9.71	237	67223	72.46	ng	99


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF033015\
 Data File : BF078092.D
 Acq On : 30 Mar 2015 20:16
 Operator : TP/IZ
 Sample : G1708-01MSD
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FK-GF-01-013-AMSD

Quant Time: Mar 31 00:43:02 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 31 01:02:18 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	9.88	196	54517	42.28	ng	98
43) 2,4,5-Trichlorophenol	9.93	196	57067	40.97	ng #	79
45) 1,1'-Biphenyl	10.11	154	207603	41.62	ng	97
46) 2-Chloronaphthalene	10.12	162	171667	42.34	ng	94
47) 2-Nitroaniline	10.26	65	44814	44.51	ng	89
48) Acenaphthylene	10.64	152	271064	40.21	ng	98
49) Dimethylphthalate	10.50	163	218678	47.25	ng	99
50) 2,6-Dinitrotoluene	10.57	165	41877	43.65	ng	91
51) Acenaphthene	10.84	154	150824	39.02	ng	98
52) 3-Nitroaniline	10.77	138	39454	35.41	ng	85
53) 2,4-Dinitrophenol	10.91	184	24041	69.40	ng #	83
54) Dibenzofuran	11.06	168	240283	41.91	ng	94
55) 4-Nitrophenol	11.01	139	68059	82.46	ng	88
56) 2,4-Dinitrotoluene	11.07	165	58970	47.14	ng #	65
57) Fluorene	11.48	166	194966	40.96	ng	100
58) 2,3,4,6-Tetrachlorophenol	11.22	232	43385	40.45	ng #	100
59) Diethylphthalate	11.38	149	200348	44.43	ng	99
60) 4-Chlorophenyl-phenylether	11.50	204	88797	40.87	ng	94
61) 4-Nitroaniline	11.53	138	46962	42.29	ng	83
62) Azobenzene	11.70	77	179958	41.75	ng	93
64) 4,6-Dinitro-2-methylphenol	11.57	198	20041	36.72	ng #	70
65) n-Nitrosodiphenylamine	11.65	169	172516	43.46	ng	100
66) 4-Bromophenyl-phenylether	12.10	248	52266	41.08	ng #	77
67) Hexachlorobenzene	12.16	284	56596	40.49	ng #	83
68) Atrazine	12.33	200	53018	47.66	ng	96
69) Pentachlorophenol	12.42	266	47644	65.65	ng	99
70) Phenanthrene	12.67	178	286581	41.88	ng	99
71) Anthracene	12.74	178	293860	43.46	ng	99
72) Carbazole	12.96	167	284926	44.30	ng	98
73) Di-n-butylphthalate	13.41	149	311429	43.18	ng #	98
74) Fluoranthene	14.15	202	312302	41.37	ng	96
76) Benzidine	14.34	184	137554	42.84	ng	99
77) Pyrene	14.43	202	334925	40.99	ng	99
79) Butylbenzylphthalate	15.28	149	136411	42.35	ng	97
80) Benzo(a)anthracene	15.94	228	311541	40.45	ng	100
81) 3,3'-Dichlorobenzidine	15.93	252	84673	33.33	ng	98
82) Chrysene	15.99	228	296141	42.76	ng	100
83) Bis(2-ethylhexyl)phthalate	16.02	149	201717	41.34	ng #	95
84) Di-n-octyl phthalate	16.87	149	330654	39.63	ng	100
85) Indeno(1,2,3-cd)pyrene	19.19	276	341243	39.48	ng #	100
87) Benzo(b)fluoranthene	17.32	252	315518	38.61	ng #	95
88) Benzo(k)fluoranthene	17.36	252	296695	46.48	ng	98
89) Benzo(a)pyrene	17.72	252	296491	43.48	ng #	96
90) Dibenzo(a,h)anthracene	19.21	278	267721	39.58	ng	97
91) Benzo(g,h,i)perylene	19.57	276	284113	41.26	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed




#1




1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 7.06 min Scan# 540
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Instrument :
 BNA_F
 ClientSampleId :
 FK-GF-01-013-AMSD

Tgt Ion:152 Resp: 30648
 Ion Ratio Lower Upper
 152 100
 150 156.5 119.5 179.3
 115 63.1 43.0 64.4




Abundance Ion 152.00 (151.70 to 152.70): E
 Ion 150.00 (149.70 to 150.70): E
 Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
 Concen: 33.54 ng
 RT: 1.92 min Scan# 90
 Delta R.T. 0.01 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Tgt Ion: 88 Resp: 26737
 Ion Ratio Lower Upper
 88 100
 58 73.0 0.0 0.0#
 43 26.3 0.0 0.0#



Abundance Ion 88.00 (87.70 to 88.70): BF0
 Ion 58.00 (57.70 to 58.70): BF0
 Ion 43.00 (42.70 to 43.70): BF0





#3

Pyridine

Concen: 31.50 ng

RT: 2.58 min Scan# 148

Delta R.T. 0.01 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16

Instrument :

BNA_F

ClientSampleId :

FK-GF-01-013-AMSD





Tgt Ion: 79 Resp: 67480

Ion Ratio Lower Upper

79 100

52 69.4 51.1 76.7

51 33.8 26.1 39.1



#4

n-Nitrosodimethylamine

Concen: 30.93 ng

RT: 2.51 min Scan# 142

Delta R.T. 0.01 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16




Tgt Ion: 42 Resp: 28847


Ion Ratio Lower Upper

42 100

74 145.6 119.3 178.9

44 8.7 5.4 8.2





#5

2-Fluorophenol

Concen: 126.61 ng

RT: 5.34 min Scan# 390

Delta R.T. -0.00 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16

Instrument :

BNA_F

ClientSampleId :

FK-GF-01-013-AMSD


Tgt Ion: 112 Resp: 222302

Ion Ratio Lower Upper

112 100

64 57.6 38.6 57.8




63 29.5 19.3 28.9#



Abundance Ion 112.00 (111.70 to 112.70): B

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0



#6

Aniline

Concen: 27.15 ng

RT: 6.65 min Scan# 504

Delta R.T. -0.00 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16


Tgt Ion: 93 Resp: 84247

Ion Ratio Lower Upper

93 100

66 57.7 30.8 46.2#



65 23.2 16.0 24.0




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 66.00 (65.70 to 66.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#7

Phenol-d6

Concen: 128.65 ng

RT: 6.65 min Scan# 504

Delta R.T. -0.00 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16

Instrument :

BNA_F

ClientSampleId :

FK-GF-01-013-AMSD


Tgt Ion: 99 Resp: 279083

Ion Ratio Lower Upper




99 100

42 18.2 11.0 16.4#

71 33.5 23.4 35.2



Abundance Ion 99.00 (98.70 to 99.70): BF0
Ion 42.00 (41.70 to 42.70): BF0
Ion 71.00 (70.70 to 71.70): BF0



#8

2-Chlorophenol

Concen: 44.43 ng

RT: 6.80 min Scan# 517

Delta R.T. -0.00 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16


Tgt Ion: 128 Resp: 92853

Ion Ratio Lower Upper



128 100


130 32.1 12.0 52.0

64 42.6 24.9 64.9



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 130.00 (129.70 to 130.70): E
Ion 64.00 (63.70 to 64.70): BF0





#9

Benzaldehyde

Concen: 3.19 ng

RT: 6.51 min Scan# 492

Delta R.T. -0.00 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16

Instrument :

BNA_F

ClientSampleId :

FK-GF-01-013-AMSD





Tgt Ion: 77 Resp: 2835

Ion Ratio Lower Upper

77 100

105 109.3 78.8 118.8

106 98.9 73.7 113.7



#10

Phenol

Concen: 43.18 ng

RT: 6.67 min Scan# 506

Delta R.T. -0.00 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16




Tgt Ion: 94 Resp: 110735


Ion Ratio Lower Upper

94 100

65 24.4 18.8 58.8

66 32.4 39.9 79.9#








#11
bis(2-Chloroethyl)ether
Concen: 44.21 ng
RT: 6.76 min Scan# 514
Delta R.T. -0.00 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16




Instrument :
BNA_F
ClientSampleId :
FK-GF-01-013-AMSD


Tgt Ion: 93 Resp: 84918
Ion Ratio Lower Upper
93 100
63 71.4 49.5 89.5
95 32.4 13.5 53.5



#12
1,3-Dichlorobenzene
Concen: 40.41 ng
RT: 6.98 min Scan# 533
Delta R.T. -0.00 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16

Tgt Ion: 146 Resp: 93944
Ion Ratio Lower Upper
146 100
148 62.9 50.9 76.3
75 32.8 21.4 32.2#









#13
 1,4-Dichlorobenzene
 Concen: 39.62 ng
 RT: 7.08 min Scan# 542
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Instrument :
 BNA_F
 ClientSampleId :
 FK-GF-01-013-AMSD

Tgt Ion:146 Resp: 95703
 Ion Ratio Lower Upper
 146 100
 148 63.8 51.9 77.9
 111 41.9 29.2 43.8




Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E





#14
 1,2-Dichlorobenzene
 Concen: 41.57 ng
 RT: 7.26 min Scan# 558
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Tgt Ion:146 Resp: 92042
 Ion Ratio Lower Upper
 146 100
 148 64.7 52.3 78.5
 111 43.5 30.6 45.8



Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E





#15

Benzyl Alcohol

Concen: 41.70 ng

RT: 7.25 min Scan# 557

Delta R.T. -0.00 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16

Instrument :

BNA_F

ClientSampleId :

FK-GF-01-013-AMSD


Tgt Ion: 79 Resp: 70610

Ion Ratio Lower Upper

79 100

108 75.2 65.0 97.4




77 63.6 53.0 79.6



Abundance Ion 79.00 (78.70 to 79.70): BF0

Ion 108.00 (107.70 to 108.70): BF0

Ion 77.00 (76.70 to 77.70): BF0



#16

2,2'-oxybis(1-Chloropropane)

Concen: 42.39 ng

RT: 7.44 min Scan# 573

Delta R.T. -0.00 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16


Tgt Ion: 45 Resp: 109729

Ion Ratio Lower Upper

45 100

77 15.4 0.0 35.2



79 12.3 0.0 32.0




Abundance Ion 45.00 (44.70 to 45.70): BF0

Ion 77.00 (76.70 to 77.70): BF0

Ion 79.00 (78.70 to 79.70): BF0






#17
2-Methylphenol
Concen: 43.32 ng
RT: 7.41 min Scan# 571
Delta R.T. -0.00 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16

Instrument :
BNA_F
ClientSampleId :
FK-GF-01-013-AMSD

Tgt Ion:	107	Resp:	73707
Ion	Ratio	Lower	Upper
107	100		
108	110.5	89.8	134.8
77	44.3	32.6	48.8
79	42.4	32.8	49.2






Abundance

Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E


Ion 77.00 (76.70 to 77.70): BF0

Ion 79.00 (78.70 to 79.70): BF0



#18
Hexachloroethane
Concen: 40.61 ng
RT: 7.68 min Scan# 594
Delta R.T. -0.00 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16

Tgt Ion:	117	Resp:	32172
Ion	Ratio	Lower	Upper
117	100		
119	98.1	79.2	118.8
201	68.9	69.8	104.6#








Abundance

Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E

Ion 201.00 (200.70 to 201.70): E





#19

n-Nitroso-di-n-propylamine

Concen: 40.51 ng

RT: 7.58 min Scan# 586

Delta R.T. -0.00 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16

Instrument :

BNA_F

ClientSampleId :

FK-GF-01-013-AMSD

Tgt Ion: 70 Resp: 60799





Ion Ratio Lower Upper

70 100

42 56.6 44.5 66.7

101 9.1 7.4 11.2

130 19.5 15.3 22.9



#20

3+4-Methylphenols

Concen: 41.93 ng

RT: 7.61 min Scan# 588

Delta R.T. -0.00 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16

Tgt Ion: 107 Resp: 93612


Ion Ratio Lower Upper


107 100

108 89.2 73.2 113.2

77 31.5 10.7 50.7

79 26.7 9.6 49.6









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.64 min Scan# 678
Delta R.T. -0.00 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16

Instrument :
BNA_F
ClientSampleId :
FK-GF-01-013-AMSD

Tgt Ion:136 Resp: 125601
Ion Ratio Lower Upper
136 100
137 10.9 8.7 13.1
54 7.5 4.6 6.8#
68 6.1 3.7 5.5#




Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF0
Ion 68.00 (67.70 to 68.70): BF0





#22
Acetophenone
Concen: 44.30 ng
RT: 7.57 min Scan# 585
Delta R.T. -0.00 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16

Tgt Ion:105 Resp: 123187
Ion Ratio Lower Upper
105 100
71 0.6 0.6 1.0
51 25.3 21.6 32.4
120 21.1 16.1 24.1



Abundance Ion 105.00 (104.70 to 105.70): E
Ion 71.00 (70.70 to 71.70): BF0
Ion 51.00 (50.70 to 51.70): BF0
Ion 120.00 (119.70 to 120.70): E









#23
Nitrobenzene-d5
Concen: 84.18 ng
RT: 7.76 min Scan# 601
Delta R.T. -0.00 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16

Instrument :
BNA_F
ClientSampleId :
FK-GF-01-013-AMSD

Tgt Ion: 82 Resp: 161304
Ion Ratio Lower Upper
82 100
128 41.5 32.8 49.2
54 49.0 40.6 60.8




Abundance Ion 82.00 (81.70 to 82.70): BF0
Ion 128.00 (127.70 to 128.70): B
Ion 54.00 (53.70 to 54.70): BF0





#24
Nitrobenzene
Concen: 41.79 ng
RT: 7.78 min Scan# 603
Delta R.T. -0.00 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16

Tgt Ion: 77 Resp: 86265
Ion Ratio Lower Upper
77 100
123 43.6 32.6 49.0
65 13.2 10.9 16.3



Abundance Ion 77.00 (76.70 to 77.70): BF0
Ion 123.00 (122.70 to 123.70): B
Ion 65.00 (64.70 to 65.70): BF0





#25

Isophorone

Concen: 41.36 ng

RT: 8.08 min Scan# 629

Delta R.T. -0.01 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16

Instrument :

BNA_F

ClientSampleId :

FK-GF-01-013-AMSD





Tgt Ion: 82 Resp: 160048

Ion Ratio Lower Upper

82 100

95 6.3 5.8 8.6

138 13.9 16.3 24.5#



#26

2-Nitrophenol

Concen: 42.38 ng

RT: 8.18 min Scan# 638

Delta R.T. -0.00 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16




Tgt Ion: 139 Resp: 42827


Ion Ratio Lower Upper

139 100

109 24.9 20.1 30.1

65 37.7 31.5 47.3









#27
 2,4-Dimethylphenol
 Concen: 39.21 ng
 RT: 8.26 min Scan# 645
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Instrument :
 BNA_F
 ClientSampleId :
 FK-GF-01-013-AMSD

Tgt Ion: 122 Resp: 72192
 Ion Ratio Lower Upper
 122 100
 107 105.0 84.7 127.1
 121 57.1 46.6 69.8




Abundance Ion 122.00 (121.70 to 122.70): E
 Ion 107.00 (106.70 to 107.70): E
 Ion 121.00 (120.70 to 121.70): E






#28
 bis(2-Chloroethoxy)methane
 Concen: 43.51 ng
 RT: 8.37 min Scan# 655
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Tgt Ion: 93 Resp: 102202
 Ion Ratio Lower Upper
 93 100
 95 33.1 25.5 38.3
 123 12.9 8.6 13.0



Abundance Ion 93.00 (92.70 to 93.70): BF0
 Ion 95.00 (94.70 to 95.70): BF0
 Ion 123.00 (122.70 to 123.70): E









#29
2,4-Dichlorophenol
Concen: 43.28 ng
RT: 8.48 min Scan# 664
Delta R.T. -0.01 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16

Instrument :
BNA_F
ClientSampleId :
FK-GF-01-013-AMSD

Tgt Ion:162 Resp: 75955
Ion Ratio Lower Upper
162 100
164 60.1 42.8 82.8
98 43.6 14.5 54.5




Abundance Ion 162.00 (161.70 to 162.70): E
Ion 164.00 (163.70 to 164.70): E
Ion 98.00 (97.70 to 98.70): BF0






#30
1,2,4-Trichlorobenzene
Concen: 39.61 ng
RT: 8.57 min Scan# 672
Delta R.T. -0.01 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16

Tgt Ion:180 Resp: 77775
Ion Ratio Lower Upper
180 100
182 93.7 77.6 116.4
145 35.3 23.4 35.2#



Abundance Ion 180.00 (179.70 to 180.70): E
Ion 182.00 (181.70 to 182.70): E
Ion 145.00 (144.70 to 145.70): E





#31

Naphthalene

Concen: 40.98 ng

RT: 8.66 min Scan# 680

Delta R.T. -0.01 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16

Instrument :

BNA_F

ClientSampleId :

FK-GF-01-013-AMSD


Tgt Ion:128 Resp: 264344

Ion Ratio Lower Upper

128 100

129 11.4 8.9 13.3




127 13.0 10.4 15.6



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 32.43 ng

RT: 8.36 min Scan# 654

Delta R.T. -0.01 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16


Tgt Ion:122 Resp: 32200

Ion Ratio Lower Upper

122 100

105 122.6 99.9 139.9



77 91.7 73.7 113.7



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF0








#33
4-Chloroaniline
Concen: 30.04 ng
RT: 8.75 min Scan# 688
Delta R.T. -0.00 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16

Instrument :
BNA_F
ClientSampleId :
FK-GF-01-013-AMSD

Tgt Ion:	127	Resp:	78647
Ion	Ratio	Lower	Upper
127	100		
129	31.1	25.3	37.9
65	22.7	17.6	26.4
92	17.1	13.1	19.7




Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BF0
Ion 92.00 (91.70 to 92.70): BF0






#34
Hexachlorobutadiene
Concen: 39.29 ng
RT: 8.82 min Scan# 694
Delta R.T. -0.00 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16

Tgt Ion:	225	Resp:	43725
Ion	Ratio	Lower	Upper
225	100		
223	62.7	47.9	71.9
227	62.7	50.1	75.1



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E








#35
 Caprolactam
 Concen: 43.20 ng
 RT: 9.16 min Scan# 724
 Delta R.T. -0.01 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16




Instrument :
 BNA_F
 ClientSampleId :
 FK-GF-01-013-AMSD


Tgt Ion:113 Resp: 22157
 Ion Ratio Lower Upper
 113 100
 55 169.2 145.9 185.9
 56 117.2 101.0 141.0



#36
 4-Chloro-3-methylphenol
 Concen: 44.30 ng
 RT: 9.37 min Scan# 742
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Tgt Ion:107 Resp: 78228
 Ion Ratio Lower Upper
 107 100
 144 27.2 23.3 34.9
 142 85.0 72.6 109.0







#37
 2-Methylnaphthalene
 Concen: 41.27 ng
 RT: 9.53 min Scan# 756
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Instrument :
 BNA_F
 ClientSampleId :
 FK-GF-01-013-AMSD



Tgt Ion:142 Resp: 178845
 Ion Ratio Lower Upper
 142 100
 141 88.6 71.9 107.9
 115 29.6 25.4 38.0



Abundance Ion 142.00 (141.70 to 142.70): E
 Ion 141.00 (140.70 to 141.70): E
 Ion 115.00 (114.70 to 115.70): E




Time--> 9.45 9.50 9.55 9.60




#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 10.81 min Scan# 868
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16


Tgt Ion:164 Resp: 62405
 Ion Ratio Lower Upper
 164 100
 162 104.9 82.4 123.6
 160 44.3 35.8 53.6




Abundance Ion 164.00 (163.70 to 164.70): E
 Ion 162.00 (161.70 to 162.70): E
 Ion 160.00 (159.70 to 160.70): E



Time--> 10.75 10.80 10.85









#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 40.21 ng
 RT: 9.73 min Scan# 774
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16




Instrument :
 BNA_F
 ClientSampleId :
 FK-GF-01-013-AMSD


Tgt Ion:	216	Resp:	74834
Ion	Ratio	Lower	Upper
216	100		
214	78.6	0.0	0.0#
179	21.0	0.0	0.0#
108	20.3	0.0	0.0#



#40
 Hexachlorocyclopentadiene
 Concen: 72.46 ng
 RT: 9.71 min Scan# 772
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Tgt Ion:	237	Resp:	67223
Ion	Ratio	Lower	Upper
237	100		
235	62.5	42.0	82.0
272	11.8	0.0	33.4









#41
 2,4,6-Tribromophenol
 Concen: 121.63 ng
 RT: 11.79 min Scan# 954
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Instrument :
 BNA_F
 ClientSampleId :
 FK-GF-01-013-AMSD

Tgt Ion:330 Resp: 72621
 Ion Ratio Lower Upper
 330 100
 332 94.4 0.0 0.0#
 141 36.0 0.0 0.0#




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#42
 2,4,6-Trichlorophenol
 Concen: 42.28 ng
 RT: 9.88 min Scan# 787
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Tgt Ion:196 Resp: 54517
 Ion Ratio Lower Upper
 196 100
 198 96.7 76.3 114.5
 200 30.3 25.6 38.4



Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E









#43
 2,4,5-Trichlorophenol
 Concen: 40.97 ng
 RT: 9.93 min Scan# 791
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Instrument :
 BNA_F
 ClientSampleId :
 FK-GF-01-013-AMSD

Tgt Ion:196 Resp: 57067
 Ion Ratio Lower Upper
 196 100
 198 90.6 79.4 119.0
 97 63.5 26.8 40.2#
 132 35.3 18.6 28.0#




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF0
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 75.44 ng
 RT: 9.98 min Scan# 796
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Tgt Ion:172 Resp: 320348
 Ion Ratio Lower Upper
 172 100
 171 34.8 28.5 42.7
 170 23.7 18.8 28.2



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#45
 1,1'-Biphenyl
 Concen: 41.62 ng
 RT: 10.11 min Scan# 807
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16




Instrument :
 BNA_F
 ClientSampleId :
 FK-GF-01-013-AMSD

Tgt Ion:154 Resp: 207603
 Ion Ratio Lower Upper
 154 100
 153 38.7 20.7 60.7
 76 9.3 0.0 30.9



#46
 2-Chloronaphthalene
 Concen: 42.34 ng
 RT: 10.12 min Scan# 808
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Tgt Ion:162 Resp: 171667
 Ion Ratio Lower Upper
 162 100
 127 39.9 26.9 40.3
 164 33.2 26.8 40.2









#47
 2-Nitroaniline
 Concen: 44.51 ng
 RT: 10.26 min Scan# 820
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16




Instrument :
 BNA_F
 ClientSampleId :
 FK-GF-01-013-AMSD

Tgt Ion: 65 Resp: 44814
 Ion Ratio Lower Upper
 65 100
 92 69.4 59.1 88.7
 138 101.4 93.9 140.9



#48
 Acenaphthylene
 Concen: 40.21 ng
 RT: 10.64 min Scan# 853
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Tgt Ion: 152 Resp: 271064
 Ion Ratio Lower Upper
 152 100
 151 19.1 15.9 23.9
 153 12.6 10.6 16.0









#49
Dimethylphthalate
Concen: 47.25 ng
RT: 10.50 min Scan# 841
Delta R.T. -0.00 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16




Instrument :
BNA_F
ClientSampleId :
FK-GF-01-013-AMSD


Tgt Ion:163 Resp: 218678
Ion Ratio Lower Upper
163 100
194 3.3 2.7 4.1
164 10.0 8.2 12.2



#50
2,6-Dinitrotoluene
Concen: 43.65 ng
RT: 10.57 min Scan# 847
Delta R.T. -0.01 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16

Tgt Ion:165 Resp: 41877
Ion Ratio Lower Upper
165 100
63 61.9 44.1 66.1
89 62.2 44.3 66.5





#51

Acenaphthene

Concen: 39.02 ng

RT: 10.84 min Scan# 871

Delta R.T. -0.01 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16

Instrument :

BNA_F

ClientSampleId :

FK-GF-01-013-AMSD


Tgt Ion:154 Resp: 150824

Ion Ratio Lower Upper

154 100

153 114.7 90.0 135.0




152 54.6 42.6 63.8



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 35.41 ng

RT: 10.77 min Scan# 865

Delta R.T. -0.00 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16


Tgt Ion:138 Resp: 39454

Ion Ratio Lower Upper

138 100

108 10.3 7.3 10.9



92 111.7 76.6 114.8




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0









#53
2,4-Dinitrophenol
Concen: 69.40 ng
RT: 10.91 min Scan# 877
Delta R.T. -0.00 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16




Instrument :
BNA_F
ClientSampleId :
FK-GF-01-013-AMSD


Tgt Ion:184 Resp: 24041
Ion Ratio Lower Upper
184 100
63 60.4 69.2 103.8#
154 64.8 53.3 79.9



#54
Dibenzofuran
Concen: 41.91 ng
RT: 11.06 min Scan# 890
Delta R.T. -0.01 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16

Tgt Ion:168 Resp: 240283
Ion Ratio Lower Upper
168 100
139 38.4 27.3 40.9
169 13.7 10.6 16.0





#55

4-Nitrophenol

Concen: 82.46 ng

RT: 11.01 min Scan# 886

Delta R.T. -0.00 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16

Instrument :

BNA_F

ClientSampleId :

FK-GF-01-013-AMSD





Tgt Ion:139 Resp: 68059

Ion Ratio Lower Upper

139 100

109 62.3 37.7 77.7

65 83.3 49.9 89.9



#56

2,4-Dinitrotoluene

Concen: 47.14 ng

RT: 11.07 min Scan# 891

Delta R.T. -0.00 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16

Tgt Ion:165 Resp: 58970




Ion Ratio Lower Upper

165 100

63 35.1 52.4 78.6#

89 59.0 72.4 108.6#

182 2.5 2.0 3.0





#57

Fluorene

Concen: 40.96 ng

RT: 11.48 min Scan# 927

Delta R.T. -0.01 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16

Instrument :

BNA_F

ClientSampleId :

FK-GF-01-013-AMSD


Tgt Ion:166 Resp: 194966

Ion Ratio Lower Upper

166 100

165 98.6 79.0 118.4




167 13.9 10.3 15.5



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#58

2,3,4,6-Tetrachlorophenol

Concen: 40.45 ng

RT: 11.22 min Scan# 904

Delta R.T. -0.01 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16

Tgt Ion:232 Resp: 43385


Ion Ratio Lower Upper

232 100

131 48.8 0.0 0.0#

130 1.6 0.0 0.0#

166 32.7 0.0 0.0#






Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E





#59

Diethylphthalate

Concen: 44.43 ng

RT: 11.38 min Scan# 918

Delta R.T. -0.00 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16

Instrument :

BNA_F

ClientSampleId :

FK-GF-01-013-AMSD


Tgt Ion:149 Resp: 200348

Ion Ratio Lower Upper

149 100

177 22.2 18.3 27.5




150 12.2 9.8 14.8



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E



#60

4-Chlorophenyl-phenylether

Concen: 40.87 ng

RT: 11.50 min Scan# 929

Delta R.T. -0.00 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16


Tgt Ion:204 Resp: 88797

Ion Ratio Lower Upper

204 100

206 32.9 26.6 39.8



141 50.1 45.2 67.8




Abundance Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E









#61
4-Nitroaniline
Concen: 42.29 ng
RT: 11.53 min Scan# 931
Delta R.T. -0.00 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16




Instrument :
BNA_F
ClientSampleId :
FK-GF-01-013-AMSD

Tgt Ion:138 Resp: 46962
Ion Ratio Lower Upper
138 100
92 51.1 16.7 56.7
108 66.2 37.6 77.6



#62
Azobenzene
Concen: 41.75 ng
RT: 11.70 min Scan# 946
Delta R.T. -0.00 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16

Tgt Ion: 77 Resp: 179958
Ion Ratio Lower Upper
77 100
182 34.1 6.5 46.5
105 25.3 3.8 43.8
51 24.1 5.5 45.5









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 12.65 min Scan# 1029
Delta R.T. -0.00 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16

Instrument :
BNA_F
ClientSampleId :
FK-GF-01-013-AMSD

Tgt Ion:188 Resp: 119212
Ion Ratio Lower Upper
188 100
94 7.4 8.7 13.1#
80 7.8 9.3 13.9#




Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BF0
Ion 80.00 (79.70 to 80.70): BF0





#64
4,6-Dinitro-2-methylphenol
Concen: 36.72 ng
RT: 11.57 min Scan# 935
Delta R.T. -0.00 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16

Tgt Ion:198 Resp: 20041
Ion Ratio Lower Upper
198 100
51 27.4 28.8 68.8#
105 31.4 30.5 70.5



Abundance Ion 198.00 (197.70 to 198.70): E
Ion 51.00 (50.70 to 51.70): BF0
Ion 105.00 (104.70 to 105.70): E





#65

n-Nitrosodiphenylamine

Concen: 43.46 ng

RT: 11.65 min Scan# 942

Delta R.T. -0.00 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16

Instrument :

BNA_F

ClientSampleId :

FK-GF-01-013-AMSD


Tgt Ion:169 Resp: 172516

Ion Ratio Lower Upper

169 100

168 67.0 53.9 80.9



167 35.9 28.9 43.3



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#66

4-Bromophenyl-phenylether

Concen: 41.08 ng

RT: 12.10 min Scan# 981

Delta R.T. -0.01 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16


Tgt Ion:248 Resp: 52266

Ion Ratio Lower Upper

248 100

250 97.8 77.4 116.0



141 101.2 45.2 67.8#



Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E









#67
Hexachlorobenzene
Concen: 40.49 ng
RT: 12.16 min Scan# 986
Delta R.T. -0.00 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16

Instrument :
BNA_F
ClientSampleId :
FK-GF-01-013-AMSD

Tgt Ion:284 Resp: 56596
Ion Ratio Lower Upper
284 100
142 40.3 20.4 30.6#
249 33.5 23.8 35.6




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E





#68
Atrazine
Concen: 47.66 ng
RT: 12.33 min Scan# 1001
Delta R.T. -0.00 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16

Tgt Ion:200 Resp: 53018
Ion Ratio Lower Upper
200 100
173 27.5 4.8 44.8
215 44.4 26.9 66.9



Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 65.65 ng
 RT: 12.42 min Scan# 1009
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Instrument :
 BNA_F
 ClientSampleId :
 FK-GF-01-013-AMSD

Tgt Ion: 266 Resp: 47644
 Ion Ratio Lower Upper
 266 100
 268 65.4 52.0 78.0
 264 62.7 49.0 73.4




Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E






#70
 Phenanthrene
 Concen: 41.88 ng
 RT: 12.67 min Scan# 1031
 Delta R.T. -0.01 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Tgt Ion: 178 Resp: 286581
 Ion Ratio Lower Upper
 178 100
 176 19.1 15.2 22.8
 179 15.5 12.1 18.1



Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E









#71
 Anthracene
 Concen: 43.46 ng
 RT: 12.74 min Scan# 1037
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16




Instrument :
 BNA_F
 ClientSampleId :
 FK-GF-01-013-AMSD

Tgt Ion:178 Resp: 293860
 Ion Ratio Lower Upper
 178 100
 176 18.6 14.6 22.0
 179 15.3 12.6 19.0



#72
 Carbazole
 Concen: 44.30 ng
 RT: 12.96 min Scan# 1056
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Tgt Ion:167 Resp: 284926
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.2 25.8
 139 11.2 10.8 16.2









#73
Di-n-butylphthalate
Concen: 43.18 ng
RT: 13.41 min Scan# 1096
Delta R.T. -0.00 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16




Instrument :
BNA_F
ClientSampleId :
FK-GF-01-013-AMSD


Tgt Ion:149 Resp: 311429
Ion Ratio Lower Upper
149 100
150 8.8 7.5 11.3
104 3.6 3.7 5.5#



#74
Fluoranthene
Concen: 41.37 ng
RT: 14.15 min Scan# 1160
Delta R.T. -0.01 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16

Tgt Ion:202 Resp: 312302
Ion Ratio Lower Upper
202 100
101 12.9 0.0 28.7
203 17.4 0.0 37.2









#75
Chrysene-d12
Concen: 20.00 ng
RT: 15.95 min Scan# 1318
Delta R.T. 0.01 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16




Instrument :
BNA_F
ClientSampleId :
FK-GF-01-013-AMSD


Tgt Ion:240 Resp: 129924
Ion Ratio Lower Upper
240 100
120 13.6 8.2 12.2#
236 25.2 20.0 30.0



#76
Benzidine
Concen: 42.84 ng
RT: 14.34 min Scan# 1177
Delta R.T. -0.00 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16

Tgt Ion:184 Resp: 137554
Ion Ratio Lower Upper
184 100
185 14.8 11.3 16.9
183 11.4 8.9 13.3









#77
 Pyrene
 Concen: 40.99 ng
 RT: 14.43 min Scan# 1185
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16




Instrument :
 BNA_F
 ClientSampleId :
 FK-GF-01-013-AMSD

Tgt Ion: 202 Resp: 334925
 Ion Ratio Lower Upper
 202 100
 200 20.5 16.9 25.3
 203 16.9 14.1 21.1



#78
 Terphenyl-d14
 Concen: 69.81 ng
 RT: 14.65 min Scan# 1204
 Delta R.T. -0.00 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Tgt Ion: 244 Resp: 371336
 Ion Ratio Lower Upper
 244 100
 212 6.1 4.9 7.3
 122 7.1 6.2 9.4









#79
Butylbenzylphthalate
Concen: 42.35 ng
RT: 15.28 min Scan# 1259
Delta R.T. 0.01 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16




Instrument :
BNA_F
ClientSampleId :
FK-GF-01-013-AMSD

Tgt Ion:149 Resp: 136411
Ion Ratio Lower Upper
149 100
91 59.9 50.1 75.1
206 20.1 16.2 24.2



#80
Benzo(a)anthracene
Concen: 40.45 ng
RT: 15.94 min Scan# 1317
Delta R.T. 0.01 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16

Tgt Ion:228 Resp: 311541
Ion Ratio Lower Upper
228 100
226 27.4 21.7 32.5
229 20.0 16.0 24.0









#81
 3,3'-Dichlorobenzidine
 Concen: 33.33 ng
 RT: 15.93 min Scan# 1316
 Delta R.T. 0.02 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Instrument :
 BNA_F
 ClientSampleId :
 FK-GF-01-013-AMSD

Tgt Ion:252 Resp: 84673
 Ion Ratio Lower Upper
 252 100
 254 63.6 51.9 77.9
 126 12.5 9.8 14.8




Abundance Ion 252.00 (251.70 to 252.70): E
 Ion 254.00 (253.70 to 254.70): E
 Ion 126.00 (125.70 to 126.70): E





#82
 Chrysene
 Concen: 42.76 ng
 RT: 15.99 min Scan# 1321
 Delta R.T. 0.02 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Tgt Ion:228 Resp: 296141
 Ion Ratio Lower Upper
 228 100
 226 30.2 24.1 36.1
 229 19.4 15.5 23.3



Abundance Ion 228.00 (227.70 to 228.70): E
 Ion 226.00 (225.70 to 226.70): E
 Ion 229.00 (228.70 to 229.70): E





#83

Bis(2-ethylhexyl)phthalate

Concen: 41.34 ng

RT: 16.02 min Scan# 1324

Delta R.T. 0.02 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16

Instrument :

BNA_F

ClientSampleId :

FK-GF-01-013-AMSD


Tgt Ion:149 Resp: 201717

Ion Ratio Lower Upper

149 100

167 25.6 22.4 33.6




279 2.4 3.4 5.0#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#84

Di-n-octyl phthalate

Concen: 39.63 ng

RT: 16.87 min Scan# 1398

Delta R.T. 0.06 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16


Tgt Ion:149 Resp: 330654

Ion Ratio Lower Upper

149 100

167 1.5 1.2 1.8



43 9.0 7.3 10.9




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BF0









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 39.48 ng
 RT: 19.19 min Scan# 1601
 Delta R.T. 0.14 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Instrument :
 BNA_F
 ClientSampleId :
 FK-GF-01-013-AMSD

Tgt Ion: 276 Resp: 341243
 Ion Ratio Lower Upper
 276 100
 138 29.5 0.0 0.0#
 277 24.7 0.0 0.0#




Abundance Ion 276.00 (275.70 to 276.70): E
 Ion 138.00 (137.70 to 138.70): E
 Ion 277.00 (276.70 to 277.70): E





#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 17.79 min Scan# 1479
 Delta R.T. 0.09 min
 Lab File: BF078092.D
 Acq: 30 Mar 2015 20:16

Tgt Ion: 264 Resp: 125192
 Ion Ratio Lower Upper
 264 100
 260 23.4 18.6 27.8
 265 21.7 17.3 25.9



Abundance Ion 264.00 (263.70 to 264.70): E
 Ion 260.00 (259.70 to 260.70): E
 Ion 265.00 (264.70 to 265.70): E








#87
Benzo(b)fluoranthene
Concen: 38.61 ng
RT: 17.32 min Scan# 1438
Delta R.T. 0.08 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16




Instrument :
BNA_F
ClientSampleId :
FK-GF-01-013-AMSD

Tgt Ion:252 Resp: 315518
Ion Ratio Lower Upper
252 100
253 21.3 18.0 27.0
125 8.3 9.9 14.9#



#88
Benzo(k)fluoranthene
Concen: 46.48 ng
RT: 17.36 min Scan# 1441
Delta R.T. 0.08 min
Lab File: BF078092.D
Acq: 30 Mar 2015 20:16

Tgt Ion:252 Resp: 296695
Ion Ratio Lower Upper
252 100
253 21.4 17.2 25.8
125 8.2 8.1 12.1





#89

Benzo(a)pyrene

Concen: 43.48 ng

RT: 17.72 min Scan# 1473

Delta R.T. 0.09 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16

Instrument :

BNA_F

ClientSampleId :

FK-GF-01-013-AMSD


Tgt Ion:252 Resp: 296491

Ion Ratio Lower Upper

252 100

253 21.4 17.6 26.4



125 13.0 7.4 11.2#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 39.58 ng

RT: 19.21 min Scan# 1603

Delta R.T. 0.13 min

Lab File: BF078092.D

Acq: 30 Mar 2015 20:16


Tgt Ion:278 Resp: 267721

Ion Ratio Lower Upper

278 100

139 19.3 13.4 20.0



279 23.3 19.0 28.4




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(g,h,i)perylene

Concen: 41.26 ng

RT: 19.57 min Scan# 1635

Delta R.T. 0.13 min

Lab File: BF078092.D


Acq: 30 Mar 2015 20:16

Instrument :

BNA_F

ClientSampleId :

FK-GF-01-013-AMSD



Tgt Ion:	276	Resp:	284113
Ion	Ratio	Lower	Upper
276	100		
277	23.9	18.5	27.7
138	26.8	19.0	28.6

