

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF033017\
 Data File : BF094068.D
 Acq On : 30 Mar 2017 19:59
 Operator : SJ/MA
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Quant Time: Mar 31 04:17:47 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 27 16:10:49 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	0.00	152	0	0.00	ng	-5.95
21) Naphthalene-d8	0.00	136	0	0.00	ng	-7.46
38) Acenaphthene-d10	0.00	164	0	0.00	ng	-9.60
63) Phenanthrene-d10	0.00	188	0	0.00	ng	-11.42
75) Chrysene-d12	0.00	240	0	0.00	ng	-14.68
86) Perylene-d12	0.00	264	0	0.00	ng	-16.31
System Monitoring Compounds						
5) 2-Fluorophenol	0.00	112	0	0.00	ng	
7) Phenol-d6	0.00	99	0	0.00	ng	
23) Nitrobenzene-d5	0.00	82	0	0.00	ng	
41) 2,4,6-Tribromophenol	0.00	330	0	0.00	ng	
44) 2-Fluorobiphenyl	0.00	172	0	0.00	ng	
78) Terphenyl-d14	13.36	244	394	0.00	ng	-0.05

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF033017\
Data File : BF094068.D
Acq On : 30 Mar 2017 19:59
Operator : SJ/MA
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Quant Time: Mar 31 04:17:47 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 27 16:10:49 2017
Response via : Initial Calibration

