

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF033017\
 Data File : BF094062.D
 Acq On : 30 Mar 2017 17:01
 Operator : SJ/MA
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Mar 31 03:18:10 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 27 16:10:49 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.92	152	201761	20.00	ng	-0.04
21) Naphthalene-d8	7.42	136	830808	20.00	ng	-0.04
38) Acenaphthene-d10	9.57	164	382339	20.00	ng	-0.03
63) Phenanthrene-d10	11.39	188	691459	20.00	ng	-0.03
75) Chrysene-d12	14.65	240	551902	20.00	ng	-0.03
86) Perylene-d12	16.27	264	461775	20.00	ng	-0.04

System Monitoring Compounds						
5) 2-Fluorophenol	4.45	112	926805	77.59	ng	-0.02
7) Phenol-d6	5.53	99	1122978	75.99	ng	-0.02
23) Nitrobenzene-d5	6.57	82	1154970	79.34	ng	-0.03
41) 2,4,6-Tribromophenol	10.55	330	264285	87.47	ng	-0.03
44) 2-Fluorobiphenyl	8.75	172	2005016	79.99	ng	-0.03
78) Terphenyl-d14	13.37	244	1964210	79.77	ng	-0.04

Target Compounds				Qvalue		
2) 1,4-Dioxane	2.28	88	223513	38.95	ng	95
3) Pyridine	2.77	79	603402	38.58	ng	99
4) n-Nitrosodimethylamine	2.73	42	241982	37.60	ng	89
6) Aniline	5.54	93	720262	35.04	ng	# 65
8) 2-Chlorophenol	5.69	128	533408	40.62	ng	94
9) Benzaldehyde	5.41	77	364908	36.57	ng	97
10) Phenol	5.55	94	639875	38.05	ng	77
11) bis(2-Chloroethyl)ether	5.62	93	522253	39.74	ng	98
12) 1,3-Dichlorobenzene	5.85	146	587225	39.87	ng	99
13) 1,4-Dichlorobenzene	5.94	146	591138	39.63	ng	98
14) 1,2-Dichlorobenzene	6.12	146	548482	39.93	ng	95
15) Benzyl Alcohol	6.09	79	417754	38.62	ng	98
16) 2,2'-oxybis(1-Chloropropan	6.25	45	720259	35.57	ng	83
17) 2-Methylphenol	6.23	107	440742	39.19	ng	# 93
18) Hexachloroethane	6.50	117	214701	40.95	ng	# 83
19) n-Nitroso-di-n-propylamine	6.40	70	379919	40.00	ng	98
20) 3+4-Methylphenols	6.42	107	586569	43.07	ng	# 63
22) Acetophenone	6.39	105	769808	41.57	ng	# 87
24) Nitrobenzene	6.59	77	567805	39.55	ng	96
25) Isophorone	6.88	82	1015596	39.70	ng	96
26) 2-Nitrophenol	6.96	139	301169	43.98	ng	95
27) 2,4-Dimethylphenol	7.04	122	474487	40.22	ng	97
28) bis(2-Chloroethoxy)methane	7.14	93	660616	39.16	ng	99
29) 2,4-Dichlorophenol	7.27	162	455615	41.30	ng	99
30) 1,2,4-Trichlorobenzene	7.36	180	458903	40.39	ng	97
31) Naphthalene	7.45	128	1586831	39.72	ng	99
32) Benzoic acid	7.21	122	320158	40.76	ng	97
33) 4-Chloroaniline	7.52	127	654370	40.42	ng	95
34) Hexachlorobutadiene	7.60	225	236463	40.58	ng	98
35) Caprolactam	7.98	113	146199	41.56	ng	92
36) 4-Chloro-3-methylphenol	8.14	107	481638	41.78	ng	90
37) 2-Methylnaphthalene	8.29	142	1074340	40.34	ng	99
39) 1,2,4,5-Tetrachlorobenzene	8.50	216	428867	41.00	ng	100
40) Hexachlorocyclopentadiene	8.49	237	195150	39.87	ng	99


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF033017\
 Data File : BF094062.D
 Acq On : 30 Mar 2017 17:01
 Operator : SJ/MA
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Mar 31 03:18:10 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 27 16:10:49 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	8.65	196	307157	41.51	ng	98
43) 2,4,5-Trichlorophenol	8.70	196	331097	41.35	ng	94
45) 1,1'-Biphenyl	8.87	154	1219566	39.55	ng	98
46) 2-Chloronaphthalene	8.89	162	958254	39.69	ng	95
47) 2-Nitroaniline	9.02	65	299854	41.87	ng	# 85
48) Acenaphthylene	9.39	152	1575493	39.27	ng	100
49) Dimethylphthalate	9.26	163	1125393	41.41	ng	100
50) 2,6-Dinitrotoluene	9.32	165	265514	42.62	ng	93
51) Acenaphthene	9.61	154	911252	38.92	ng	99
52) 3-Nitroaniline	9.53	138	304104	41.57	ng	93
53) 2,4-Dinitrophenol	9.66	184	138241	43.71	ng	# 67
54) Dibenzofuran	9.82	168	1224652	38.43	ng	100
55) 4-Nitrophenol	9.77	139	221385	38.64	ng	# 72
56) 2,4-Dinitrotoluene	9.82	165	318968	40.76	ng	# 73
57) Fluorene	10.24	166	1006079	38.07	ng	99
58) 2,3,4,6-Tetrachlorophenol	9.98	232	224853	40.93	ng	98
59) Diethylphthalate	10.12	149	1093382	40.70	ng	99
60) 4-Chlorophenyl-phenylether	10.25	204	436716	38.79	ng	95
61) 4-Nitroaniline	10.29	138	309477	44.08	ng	96
62) Azobenzene	10.45	77	993678	39.11	ng	94
64) 4,6-Dinitro-2-methylphenol	10.32	198	188416	44.49	ng	# 74
65) n-Nitrosodiphenylamine	10.40	169	933087	39.02	ng	99
66) 4-Bromophenyl-phenylether	10.85	248	275608	40.53	ng	98
67) Hexachlorobenzene	10.92	284	282529	41.04	ng	# 92
68) Atrazine	11.07	200	294052	41.06	ng	97
69) Pentachlorophenol	11.17	266	141578	33.04	ng	99
70) Phenanthrene	11.42	178	1505850	39.15	ng	99
71) Anthracene	11.49	178	1565120	39.52	ng	99
72) Carbazole	11.69	167	1584165	39.01	ng	99
73) Di-n-butylphthalate	12.13	149	1712412	40.55	ng	100
74) Fluoranthene	12.89	202	1564606	39.72	ng	99
76) Benzidine	13.06	184	786547	36.01	ng	# 94
77) Pyrene	13.16	202	1607853	40.14	ng	100
79) Butylbenzylphthalate	13.97	149	740452	43.39	ng	# 83
80) Benzo(a)anthracene	14.63	228	1312079	40.77	ng	99
81) 3,3'-Dichlorobenzidine	14.61	252	488203	42.44	ng	# 96
82) Chrysene	14.68	228	1145624	38.36	ng	98
83) Bis(2-ethylhexyl)phthalate	14.69	149	934224	40.12	ng	# 99
84) Di-n-octyl phthalate	15.44	149	1675944	43.09	ng	98
85) Indeno(1,2,3-cd)pyrene	17.64	276	1015096	39.24	ng	100
87) Benzo(b)fluoranthene	15.87	252	1223988	43.24	ng	99
88) Benzo(k)fluoranthene	15.90	252	1028921m	37.15	ng	
89) Benzo(a)pyrene	16.22	252	1053958	40.43	ng	99
90) Dibenzo(a,h)anthracene	17.66	278	863474	39.12	ng	98
91) Benzo(g,h,i)perylene	18.04	276	853156	38.35	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 5.92 min Scan# 685

Delta R.T. -0.04 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:152 Resp: 201761

Ion Ratio Lower Upper

152 100

150 154.0 126.2 189.2




115 60.0 52.2 78.2



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 38.95 ng

RT: 2.28 min Scan# 66

Delta R.T. -0.05 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01


Tgt Ion: 88 Resp: 223513

Ion Ratio Lower Upper

88 100

58 72.0 61.4 92.0



43 27.5 23.4 35.0




Abundance Ion 88.00 (87.70 to 88.70): BF0

Ion 58.00 (57.70 to 58.70): BF0

Ion 43.00 (42.70 to 43.70): BF0





#3

Pyridine

Concen: 38.58 ng

RT: 2.77 min Scan# 150

Delta R.T. -0.04 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion: 79 Resp: 603402

Ion Ratio Lower Upper

79 100

52 69.7 54.8 82.2

51 33.6 27.0 40.4



#4

n-Nitrosodimethylamine

Concen: 37.60 ng

RT: 2.73 min Scan# 144

Delta R.T. -0.04 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01




Tgt Ion: 42 Resp: 241982


Ion Ratio Lower Upper

42 100

74 142.9 103.9 155.9

44 7.1 5.7 8.5





#5

2-Fluorophenol

Concen: 77.59 ng

RT: 4.45 min Scan# 436

Delta R.T. -0.02 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 112 Resp: 926805

Ion Ratio Lower Upper

112 100

64 59.2 46.0 69.0




63 29.2 23.4 35.0



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0



#6

Aniline

Concen: 35.04 ng

RT: 5.54 min Scan# 621

Delta R.T. -0.04 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01


Tgt Ion: 93 Resp: 720262

Ion Ratio Lower Upper

93 100

66 63.4 32.9 49.3#



65 35.1 16.0 24.0#




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 66.00 (65.70 to 66.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#7

Phenol-d6

Concen: 75.99 ng

RT: 5.53 min Scan# 620

Delta R.T. -0.02 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 99 Resp: 1122978

Ion Ratio Lower Upper

99 100

42 18.2 13.5 20.3




71 33.2 24.5 36.7



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#8

2-Chlorophenol

Concen: 40.62 ng

RT: 5.69 min Scan# 646

Delta R.T. -0.03 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01


Tgt Ion: 128 Resp: 533408

Ion Ratio Lower Upper

128 100

130 32.6 12.9 52.9



64 42.4 28.4 68.4




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF0





#9

Benzaldehyde

Concen: 36.57 ng

RT: 5.41 min Scan# 599

Delta R.T. -0.04 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 77 Resp: 364908

Ion Ratio Lower Upper

77 100

105 101.9 83.8 123.8




106 94.0 79.1 119.1



Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 105.00 (104.70 to 105.70): BF0

Ion 106.00 (105.70 to 106.70): BF0



#10

Phenol

Concen: 38.05 ng

RT: 5.55 min Scan# 622

Delta R.T. -0.02 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01


Tgt Ion: 94 Resp: 639875

Ion Ratio Lower Upper

94 100

65 38.0 9.9 49.9



66 61.3 23.1 63.1




Abundance Ion 94.00 (93.70 to 94.70): BF0

Ion 65.00 (64.70 to 65.70): BF0

Ion 66.00 (65.70 to 66.70): BF0









#11
bis(2-Chloroethyl)ether
Concen: 39.74 ng
RT: 5.62 min Scan# 635
Delta R.T. -0.04 min
Lab File: BF094062.D
Acq: 30 Mar 2017 17:01




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 93 Resp: 522253
Ion Ratio Lower Upper
93 100
63 76.8 59.2 99.2
95 32.4 12.8 52.8



#12
1,3-Dichlorobenzene
Concen: 39.87 ng
RT: 5.85 min Scan# 674
Delta R.T. -0.04 min
Lab File: BF094062.D
Acq: 30 Mar 2017 17:01

Tgt Ion: 146 Resp: 587225
Ion Ratio Lower Upper
146 100
148 63.8 51.8 77.6
75 28.7 23.1 34.7









#13
 1,4-Dichlorobenzene
 Concen: 39.63 ng
 RT: 5.94 min Scan# 689
 Delta R.T. -0.04 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:146 Resp: 591138
 Ion Ratio Lower Upper
 146 100
 148 64.8 52.2 78.2
 111 40.8 30.3 45.5




Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E








#14
 1,2-Dichlorobenzene
 Concen: 39.93 ng
 RT: 6.12 min Scan# 719
 Delta R.T. -0.04 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Tgt Ion:146 Resp: 548482
 Ion Ratio Lower Upper
 146 100
 148 65.6 50.4 75.6
 111 42.8 38.2 57.4



Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E





#15

Benzyl Alcohol

Concen: 38.62 ng

RT: 6.09 min Scan# 714

Delta R.T. -0.03 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion: 79 Resp: 417754

Ion Ratio Lower Upper

79 100

108 77.2 63.4 95.0

77 59.8 49.2 73.8



#16

2,2'-oxybis(1-Chloropropane)

Concen: 35.57 ng

RT: 6.25 min Scan# 741

Delta R.T. -0.03 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01


Tgt Ion: 45 Resp: 720259


Ion Ratio Lower Upper

45 100

77 19.7 0.0 33.2

79 16.5 0.0 30.0









#17
2-Methylphenol
Concen: 39.19 ng
RT: 6.23 min Scan# 739
Delta R.T. -0.02 min
Lab File: BF094062.D
Acq: 30 Mar 2017 17:01




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:107 Resp: 440742
Ion Ratio Lower Upper
107 100
108 112.8 93.4 140.0
77 54.0 35.8 53.6#
79 50.4 34.8 52.2



#18
Hexachloroethane
Concen: 40.95 ng
RT: 6.50 min Scan# 785
Delta R.T. -0.04 min
Lab File: BF094062.D
Acq: 30 Mar 2017 17:01

Tgt Ion:117 Resp: 214701
Ion Ratio Lower Upper
117 100
119 94.9 77.4 116.0
201 86.8 94.6 141.8#









#19
n-Nitroso-di-n-propylamine
Concen: 40.00 ng
RT: 6.40 min Scan# 768
Delta R.T. -0.04 min
Lab File: BF094062.D
Acq: 30 Mar 2017 17:01




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 70 Resp: 379919
Ion Ratio Lower Upper
70 100
42 57.9 47.2 70.8
101 10.3 7.2 10.8
130 21.0 15.9 23.9



#20
3+4-Methylphenols
Concen: 43.07 ng
RT: 6.42 min Scan# 770
Delta R.T. -0.02 min
Lab File: BF094062.D
Acq: 30 Mar 2017 17:01

Tgt Ion: 107 Resp: 586569
Ion Ratio Lower Upper
107 100
108 92.6 71.0 111.0
77 31.0 90.5 130.5#
79 28.8 8.4 48.4









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 7.42 min Scan# 941
Delta R.T. -0.04 min
Lab File: BF094062.D
Acq: 30 Mar 2017 17:01




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:	136	Resp:	830808
Ion	Ratio	Lower	Upper
136	100		
137	11.3	8.5	12.7
54	7.4	4.9	7.3#
68	5.7	4.1	6.1



#22
Acetophenone
Concen: 41.57 ng
RT: 6.39 min Scan# 766
Delta R.T. -0.04 min
Lab File: BF094062.D
Acq: 30 Mar 2017 17:01

Tgt Ion:	105	Resp:	769808
Ion	Ratio	Lower	Upper
105	100		
71	2.9	2.8	4.2
51	25.6	30.7	46.1#
120	22.0	17.4	26.0









#23
 Nitrobenzene-d5
 Concen: 79.34 ng
 RT: 6.57 min Scan# 797
 Delta R.T. -0.03 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion: 82 Resp: 1154970
 Ion Ratio Lower Upper
 82 100
 128 47.7 34.0 51.0
 54 49.3 42.2 63.2



#24
 Nitrobenzene
 Concen: 39.55 ng
 RT: 6.59 min Scan# 800
 Delta R.T. -0.04 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Tgt Ion: 77 Resp: 567805
 Ion Ratio Lower Upper
 77 100
 123 48.5 35.8 53.8
 65 13.1 10.6 15.8





#25

Isophorone

Concen: 39.70 ng

RT: 6.88 min Scan# 849

Delta R.T. -0.04 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 82 Resp: 1015596

Ion Ratio Lower Upper




82 100

95 6.7 5.3 7.9

138 18.8 13.1 19.7



Abundance Ion 82.00 (81.70 to 82.70): BF0
Ion 95.00 (94.70 to 95.70): BF0
Ion 138.00 (137.70 to 138.70): BF0



#26

2-Nitrophenol

Concen: 43.98 ng

RT: 6.96 min Scan# 863

Delta R.T. -0.04 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01


Tgt Ion: 139 Resp: 301169

Ion Ratio Lower Upper



139 100


109 26.7 21.0 31.6

65 46.4 33.0 49.6



Abundance Ion 139.00 (138.70 to 139.70): E
Ion 109.00 (108.70 to 109.70): E
Ion 65.00 (64.70 to 65.70): BF0





#27

2,4-Dimethylphenol

Concen: 40.22 ng

RT: 7.04 min Scan# 876

Delta R.T. -0.02 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:122 Resp: 474487

Ion Ratio Lower Upper

122 100

107 108.5 89.2 133.8




121 58.2 45.2 67.8



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 107.00 (106.70 to 107.70): E

Ion 121.00 (120.70 to 121.70): E



#28

bis(2-Chloroethoxy)methane

Concen: 39.16 ng

RT: 7.14 min Scan# 893

Delta R.T. -0.04 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01


Tgt Ion: 93 Resp: 660616

Ion Ratio Lower Upper

93 100

95 32.2 26.5 39.7



123 11.2 9.0 13.4




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 95.00 (94.70 to 95.70): BF0

Ion 123.00 (122.70 to 123.70): E









#29
 2,4-Dichlorophenol
 Concen: 41.30 ng
 RT: 7.27 min Scan# 915
 Delta R.T. -0.02 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion	Ratio	Lower	Upper
162	100		
164	64.3	44.6	84.6
98	35.4	14.8	54.8




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 98.00 (97.70 to 98.70): BF0






#30
 1,2,4-Trichlorobenzene
 Concen: 40.39 ng
 RT: 7.36 min Scan# 930
 Delta R.T. -0.04 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Tgt Ion	Ratio	Lower	Upper
180	100		
182	98.0	75.8	113.6
145	33.4	25.3	37.9



Abundance Ion 180.00 (179.70 to 180.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 145.00 (144.70 to 145.70): E





#31

Naphthalene

Concen: 39.72 ng

RT: 7.45 min Scan# 946

Delta R.T. -0.04 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:128 Resp: 1586831

Ion Ratio Lower Upper

128 100

129 11.4 9.0 13.6




127 13.9 10.8 16.2



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 40.76 ng

RT: 7.21 min Scan# 905

Delta R.T. -0.01 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01


Tgt Ion:122 Resp: 320158

Ion Ratio Lower Upper

122 100

105 121.0 103.3 143.3



77 90.5 73.7 113.7




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF0





#33

4-Chloroaniline

Concen: 40.42 ng

RT: 7.52 min Scan# 957

Delta R.T. -0.04 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01


Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:	127	Resp:	654370
Ion	Ratio	Lower	Upper
127	100		
129	32.4	26.2	39.2
65	29.5	27.8	41.8
92	19.1	16.8	25.2






Abundance Ion 127.00 (126.70 to 127.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 65.00 (64.70 to 65.70): BF0

Ion 92.00 (91.70 to 92.70): BF0



#34

Hexachlorobutadiene

Concen: 40.58 ng


RT: 7.60 min Scan# 972

Delta R.T. -0.04 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01



Tgt Ion:	225	Resp:	236463
Ion	Ratio	Lower	Upper
225	100		
223	61.9	48.8	73.2
227	62.0	51.1	76.7




Abundance Ion 225.00 (224.70 to 225.70): E

Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E









#35
 Caprolactam
 Concen: 41.56 ng
 RT: 7.98 min Scan# 1036
 Delta R.T. -0.02 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:113 Resp: 146199
 Ion Ratio Lower Upper
 113 100
 55 160.5 150.2 190.2
 56 117.9 107.8 147.8




Abundance Ion 113.00 (112.70 to 113.70): E
 Ion 55.00 (54.70 to 55.70): BF0
 Ion 56.00 (55.70 to 56.70): BF0






#36
 4-Chloro-3-methylphenol
 Concen: 41.78 ng
 RT: 8.14 min Scan# 1063
 Delta R.T. -0.02 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Tgt Ion:107 Resp: 481638
 Ion Ratio Lower Upper
 107 100
 144 26.7 24.2 36.4
 142 80.8 73.4 110.0



Abundance Ion 107.00 (106.70 to 107.70): E
 Ion 144.00 (143.70 to 144.70): E
 Ion 142.00 (141.70 to 142.70): E









#37
 2-Methylnaphthalene
 Concen: 40.34 ng
 RT: 8.29 min Scan# 1089
 Delta R.T. -0.04 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:142 Resp: 1074340
 Ion Ratio Lower Upper
 142 100
 141 89.4 71.3 106.9
 115 31.9 27.1 40.7



#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 9.57 min Scan# 1306
 Delta R.T. -0.03 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Tgt Ion:164 Resp: 382339
 Ion Ratio Lower Upper
 164 100
 162 104.6 82.6 123.8
 160 46.5 36.2 54.2









#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 41.00 ng
 RT: 8.50 min Scan# 1124
 Delta R.T. -0.04 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:	216	Resp:	428867
Ion	Ratio	Lower	Upper
216	100		
214	79.2	63.4	95.2
179	22.4	17.9	26.9
108	20.4	16.5	24.7




Abundance Ion 216.00 (215.70 to 216.70): E
 Ion 214.00 (213.70 to 214.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 108.00 (107.70 to 108.70): E






#40
 Hexachlorocyclopentadiene
 Concen: 39.87 ng
 RT: 8.49 min Scan# 1122
 Delta R.T. -0.04 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Tgt Ion:	237	Resp:	195150
Ion	Ratio	Lower	Upper
237	100		
235	61.7	42.6	82.6
272	12.2	0.0	31.9



Abundance Ion 236.80 (236.50 to 237.50): E
 Ion 234.90 (234.60 to 235.60): E
 Ion 271.75 (271.45 to 272.45): E









#41
 2,4,6-Tribromophenol
 Concen: 87.47 ng
 RT: 10.55 min Scan# 1472
 Delta R.T. -0.03 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:330 Resp: 264285
 Ion Ratio Lower Upper
 330 100
 332 96.1 76.6 115.0
 141 39.3 31.4 47.2




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#42
 2,4,6-Trichlorophenol
 Concen: 41.51 ng
 RT: 8.65 min Scan# 1149
 Delta R.T. -0.03 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Tgt Ion:196 Resp: 307157
 Ion Ratio Lower Upper
 196 100
 198 95.3 74.2 111.4
 200 29.9 24.0 36.0



Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E





#43

2,4,5-Trichlorophenol

Concen: 41.35 ng

RT: 8.70 min Scan# 1158

Delta R.T. -0.02 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:196 Resp: 331097


Ion Ratio Lower Upper

196 100

198 91.3 75.7 113.5

97 52.0 47.7 71.5

132 32.0 28.0 42.0






Abundance Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 97.00 (96.70 to 97.70): BF0

Ion 132.00 (131.70 to 132.70): E



#44

2-Fluorobiphenyl

Concen: 79.99 ng

RT: 8.75 min Scan# 1167

Delta R.T. -0.03 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01


Tgt Ion:172 Resp: 2005016

Ion Ratio Lower Upper

172 100

171 36.7 29.6 44.4



170 24.4 19.8 29.8




Abundance Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E









#45
 1,1'-Biphenyl
 Concen: 39.55 ng
 RT: 8.87 min Scan# 1187
 Delta R.T. -0.03 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:154 Resp: 1219566
 Ion Ratio Lower Upper
 154 100
 153 41.4 21.2 61.2
 76 12.5 0.0 29.9




Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): BF0






#46
 2-Chloronaphthalene
 Concen: 39.69 ng
 RT: 8.89 min Scan# 1190
 Delta R.T. -0.04 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Tgt Ion:162 Resp: 958254
 Ion Ratio Lower Upper
 162 100
 127 40.5 28.3 42.5
 164 33.4 27.0 40.4



Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E





#47

2-Nitroaniline

Concen: 41.87 ng

RT: 9.02 min Scan# 1212

Delta R.T. -0.03 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 65 Resp: 299854

Ion Ratio Lower Upper




65 100

92 72.0 53.9 80.9

138 119.2 78.8 118.2#



Abundance Ion 65.00 (64.70 to 65.70): BF0
Ion 92.00 (91.70 to 92.70): BF0
Ion 138.00 (137.70 to 138.70): E



#48

Acenaphthylene

Concen: 39.27 ng

RT: 9.39 min Scan# 1276

Delta R.T. -0.04 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01


Tgt Ion: 152 Resp: 1575493

Ion Ratio Lower Upper



152 100


151 20.9 16.8 25.2

153 13.4 10.8 16.2



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E









#49
Dimethylphthalate
Concen: 41.41 ng
RT: 9.26 min Scan# 1253
Delta R.T. -0.03 min
Lab File: BF094062.D
Acq: 30 Mar 2017 17:01




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:163 Resp: 1125393
Ion Ratio Lower Upper
163 100
194 3.4 2.6 4.0
164 10.4 8.4 12.6



#50
2,6-Dinitrotoluene
Concen: 42.62 ng
RT: 9.32 min Scan# 1264
Delta R.T. -0.03 min
Lab File: BF094062.D
Acq: 30 Mar 2017 17:01

Tgt Ion:165 Resp: 265514
Ion Ratio Lower Upper
165 100
63 49.9 34.3 51.5
89 49.2 37.1 55.7





#51

Acenaphthene

Concen: 38.92 ng

RT: 9.61 min Scan# 1313

Delta R.T. -0.04 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:154 Resp: 911252

Ion Ratio Lower Upper

154 100

153 114.0 90.5 135.7




152 54.1 43.6 65.4



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 41.57 ng

RT: 9.53 min Scan# 1299

Delta R.T. -0.02 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01


Tgt Ion:138 Resp: 304104

Ion Ratio Lower Upper

138 100

108 10.4 9.1 13.7



92 108.5 93.8 140.6




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0









#53
2,4-Dinitrophenol
Concen: 43.71 ng
RT: 9.66 min Scan# 1321
Delta R.T. -0.02 min
Lab File: BF094062.D
Acq: 30 Mar 2017 17:01

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:184 Resp: 138241
Ion Ratio Lower Upper
184 100
63 56.3 62.7 94.1#
154 61.3 81.1 121.7#




Abundance Ion 184.00 (183.70 to 184.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 154.00 (153.70 to 154.70): E






#54
Dibenzofuran
Concen: 38.43 ng
RT: 9.82 min Scan# 1349
Delta R.T. -0.04 min
Lab File: BF094062.D
Acq: 30 Mar 2017 17:01

Tgt Ion:168 Resp: 1224652
Ion Ratio Lower Upper
168 100
139 38.0 30.6 45.8
169 13.3 10.8 16.2



Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E









#55
4-Nitrophenol
Concen: 38.64 ng
RT: 9.77 min Scan# 1340
Delta R.T. -0.01 min
Lab File: BF094062.D
Acq: 30 Mar 2017 17:01




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:139 Resp: 221385
Ion Ratio Lower Upper
139 100
109 61.2 34.1 74.1
65 84.0 31.4 71.4#



#56
2,4-Dinitrotoluene
Concen: 40.76 ng
RT: 9.82 min Scan# 1348
Delta R.T. -0.03 min
Lab File: BF094062.D
Acq: 30 Mar 2017 17:01

Tgt Ion:165 Resp: 318968
Ion Ratio Lower Upper
165 100
63 54.6 25.4 38.0#
89 72.9 46.4 69.6#
182 2.3 1.8 2.6









#57
 Fluorene
 Concen: 38.07 ng
 RT: 10.24 min Scan# 1420
 Delta R.T. -0.04 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01




Instrument :
 BNA_F
 ClientSampleId :
 SSTCCCC040


Tgt Ion:166 Resp: 1006079
 Ion Ratio Lower Upper
 166 100
 165 98.0 78.8 118.2
 167 14.0 10.6 16.0



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 40.93 ng
 RT: 9.98 min Scan# 1376
 Delta R.T. -0.03 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Tgt Ion:232 Resp: 224853
 Ion Ratio Lower Upper
 232 100
 131 53.7 44.8 67.2
 130 2.6 2.3 3.5
 166 35.9 29.0 43.4





#59

Diethylphthalate

Concen: 40.70 ng

RT: 10.12 min Scan# 1400

Delta R.T. -0.04 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:149 Resp: 1093382

Ion Ratio Lower Upper

149 100

177 21.4 16.7 25.1




150 12.5 10.2 15.2



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E



#60

4-Chlorophenyl-phenylether

Concen: 38.79 ng

RT: 10.25 min Scan# 1421

Delta R.T. -0.04 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01


Tgt Ion:204 Resp: 436716

Ion Ratio Lower Upper

204 100

206 33.3 26.2 39.2



141 69.8 60.8 91.2




Abundance Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E





#61

4-Nitroaniline

Concen: 44.08 ng

RT: 10.29 min Scan# 1428

Delta R.T. -0.02 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:138 Resp: 309477

Ion Ratio Lower Upper

138 100

92 46.8 21.8 61.8




108 62.9 42.3 82.3



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BF0

Ion 108.00 (107.70 to 108.70): E



#62

Azobenzene

Concen: 39.11 ng

RT: 10.45 min Scan# 1455

Delta R.T. -0.03 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01

Tgt Ion: 77 Resp: 993678


Ion Ratio Lower Upper

77 100

182 26.7 0.1 40.1

105 24.4 3.6 43.6

51 27.4 9.4 49.4






Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 182.00 (181.70 to 182.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 51.00 (50.70 to 51.70): BF0





#63

Phenanthrene-d10

Concen: 20.00 ng

RT: 11.39 min Scan# 1616

Delta R.T. -0.03 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:188 Resp: 691459

Ion Ratio Lower Upper

188 100

94 10.1 9.4 14.2

80 10.4 10.2 15.2



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF0

Ion 80.00 (79.70 to 80.70): BF0

800000

600000

400000

200000

0



11.39

11.35

11.40

11.45

Time-->



#64

4,6-Dinitro-2-methylphenol

Concen: 44.49 ng

RT: 10.32 min Scan# 1434

Delta R.T. -0.03 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01


Tgt Ion:198 Resp: 188416

Ion Ratio Lower Upper

198 100

51 48.0 53.2 93.2#

105 47.9 45.8 85.8



Abundance Ion 198.00 (197.70 to 198.70): E

Ion 51.00 (50.70 to 51.70): BF0

Ion 105.00 (104.70 to 105.70): E

300000

200000

100000


0


10.32

10.30

10.40

Time-->









#65
 n-Nitrosodiphenylamine
 Concen: 39.02 ng
 RT: 10.40 min Scan# 1447
 Delta R.T. -0.03 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:169 Resp: 933087
 Ion Ratio Lower Upper
 169 100
 168 67.2 53.2 79.8
 167 37.5 29.5 44.3




Abundance Ion 169.00 (168.70 to 169.70): E
 Ion 168.00 (167.70 to 168.70): E
 Ion 167.00 (166.70 to 167.70): E






#66
 4-Bromophenyl-phenylether
 Concen: 40.53 ng
 RT: 10.85 min Scan# 1523
 Delta R.T. -0.04 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Tgt Ion:248 Resp: 275608
 Ion Ratio Lower Upper
 248 100
 250 97.1 76.8 115.2
 141 89.3 68.6 103.0



Abundance Ion 248.00 (247.70 to 248.70): E
 Ion 250.00 (249.70 to 250.70): E
 Ion 141.00 (140.70 to 141.70): E









#67
Hexachlorobenzene
Concen: 41.04 ng
RT: 10.92 min Scan# 1536
Delta R.T. -0.03 min
Lab File: BF094062.D
Acq: 30 Mar 2017 17:01




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 284 Resp: 282529
Ion Ratio Lower Upper
284 100
142 36.7 22.8 34.2#
249 30.4 24.0 36.0



#68
Atrazine
Concen: 41.06 ng
RT: 11.07 min Scan# 1562
Delta R.T. -0.02 min
Lab File: BF094062.D
Acq: 30 Mar 2017 17:01

Tgt Ion: 200 Resp: 294052
Ion Ratio Lower Upper
200 100
173 28.0 6.3 46.3
215 45.4 27.2 67.2









#69
 Pentachlorophenol
 Concen: 33.04 ng
 RT: 11.17 min Scan# 1578
 Delta R.T. -0.03 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion: 266 Resp: 141578
 Ion Ratio Lower Upper
 266 100
 268 63.9 51.1 76.7
 264 63.3 51.7 77.5




Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E






#70
 Phenanthrene
 Concen: 39.15 ng
 RT: 11.42 min Scan# 1621
 Delta R.T. -0.04 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Tgt Ion: 178 Resp: 1505850
 Ion Ratio Lower Upper
 178 100
 176 19.4 16.2 24.4
 179 15.5 12.6 19.0



Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E





#71

Anthracene

Concen: 39.52 ng

RT: 11.49 min Scan# 1632

Delta R.T. -0.04 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:178 Resp: 1565120

Ion Ratio Lower Upper

178 100

176 19.1 15.8 23.8




179 15.6 12.7 19.1



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#72

Carbazole

Concen: 39.01 ng

RT: 11.69 min Scan# 1667

Delta R.T. -0.03 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01


Tgt Ion:167 Resp: 1584165

Ion Ratio Lower Upper

167 100

166 22.5 18.1 27.1



139 13.5 11.7 17.5




Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E





#73

Di-n-butylphthalate

Concen: 40.55 ng

RT: 12.13 min Scan# 1742

Delta R.T. -0.04 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:149 Resp: 1712412

Ion Ratio Lower Upper

149 100

150 9.6 7.7 11.5




104 4.8 4.0 6.0



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 39.72 ng

RT: 12.89 min Scan# 1870

Delta R.T. -0.04 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01


Tgt Ion:202 Resp: 1564606

Ion Ratio Lower Upper

202 100

101 13.4 0.0 33.2



203 17.9 0.0 38.1




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.00 ng

RT: 14.65 min Scan# 2170

Delta R.T. -0.03 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:240 Resp: 551902

Ion Ratio Lower Upper

240 100

120 12.6 5.8 8.8#




236 25.8 20.5 30.7



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 36.01 ng

RT: 13.06 min Scan# 1899

Delta R.T. -0.03 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01


Tgt Ion:184 Resp: 786547

Ion Ratio Lower Upper

184 100

185 15.4 15.5 23.3#



183 12.3 9.8 14.6




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E






#77
 Pyrene
 Concen: 40.14 ng
 RT: 13.16 min Scan# 1917
 Delta R.T. -0.04 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion: 202 Resp: 1607853

Ion	Ratio	Lower	Upper
202	100		
200	21.9	17.6	26.4
203	17.9	14.4	21.6




Abundance Ion 202.00 (201.70 to 202.70): E
 Ion 200.00 (199.70 to 200.70): E
 Ion 203.00 (202.70 to 203.70): E





#78
 Terphenyl-d14
 Concen: 79.77 ng
 RT: 13.37 min Scan# 1952
 Delta R.T. -0.04 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01


Tgt Ion: 244 Resp: 1964210

Ion	Ratio	Lower	Upper
244	100		
212	7.2	6.0	9.0
122	14.1	10.3	15.5



Abundance Ion 244.00 (243.70 to 244.70): E
 Ion 212.00 (211.70 to 212.70): E
 Ion 122.00 (121.70 to 122.70): E









#79
Butylbenzylphthalate
Concen: 43.39 ng
RT: 13.97 min Scan# 2055
Delta R.T. -0.04 min
Lab File: BF094062.D
Acq: 30 Mar 2017 17:01




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:149 Resp: 740452
Ion Ratio Lower Upper
149 100
91 72.1 45.0 67.4#
206 18.5 17.0 25.6



#80
Benzo(a)anthracene
Concen: 40.77 ng
RT: 14.63 min Scan# 2167
Delta R.T. -0.04 min
Lab File: BF094062.D
Acq: 30 Mar 2017 17:01

Tgt Ion:228 Resp: 1312079
Ion Ratio Lower Upper
228 100
226 27.6 22.6 33.8
229 20.2 16.3 24.5









#81
 3,3'-Dichlorobenzidine
 Concen: 42.44 ng
 RT: 14.61 min Scan# 2163
 Delta R.T. -0.03 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion	Ratio	Lower	Upper
252	100		
254	66.1	51.5	77.3
126	14.3	15.8	23.8#




Abundance Ion 252.00 (251.70 to 252.70): E
 Ion 254.00 (253.70 to 254.70): E
 Ion 126.00 (125.70 to 126.70): E






#82
 Chrysene
 Concen: 38.36 ng
 RT: 14.68 min Scan# 2175
 Delta R.T. -0.04 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.3	24.9	37.3
229	20.6	15.8	23.6



Abundance Ion 228.00 (227.70 to 228.70): E
 Ion 226.00 (225.70 to 226.70): E
 Ion 229.00 (228.70 to 229.70): E









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 40.12 ng
 RT: 14.69 min Scan# 2177
 Delta R.T. -0.03 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:149 Resp: 934224
 Ion Ratio Lower Upper
 149 100
 167 25.7 21.0 31.4
 279 3.4 1.9 2.9#



#84
 Di-n-octyl phthalate
 Concen: 43.09 ng
 RT: 15.44 min Scan# 2305
 Delta R.T. -0.04 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Tgt Ion:149 Resp: 1675944
 Ion Ratio Lower Upper
 149 100
 167 1.6 1.2 1.8
 43 9.6 8.5 12.7









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 39.24 ng
 RT: 17.64 min Scan# 2678
 Delta R.T. -0.05 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:276 Resp: 1015096
 Ion Ratio Lower Upper
 276 100
 138 34.8 27.8 41.6
 277 25.7 20.2 30.4



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 16.27 min Scan# 2446
 Delta R.T. -0.04 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Tgt Ion:264 Resp: 461775
 Ion Ratio Lower Upper
 264 100
 260 23.3 19.5 29.3
 265 21.4 17.2 25.8









#87
Benzo(b)fluoranthene
Concen: 43.24 ng
RT: 15.87 min Scan# 2377
Delta R.T. -0.03 min
Lab File: BF094062.D
Acq: 30 Mar 2017 17:01




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:252 Resp: 1223988
Ion Ratio Lower Upper
252 100
253 22.4 18.1 27.1
125 13.1 9.8 14.8



#88
Benzo(k)fluoranthene
Concen: 37.15 ng m
RT: 15.90 min Scan# 2383
Delta R.T. -0.03 min
Lab File: BF094062.D
Acq: 30 Mar 2017 17:01

Tgt Ion:252 Resp: 1028921
Ion Ratio Lower Upper
252 100
253 21.7 18.4 27.6
125 12.3 13.1 19.7#





#89

Benzo(a)pyrene

Concen: 40.43 ng

RT: 16.22 min Scan# 2437

Delta R.T. -0.03 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:252 Resp: 1053958

Ion Ratio Lower Upper

252 100

253 21.5 17.5 26.3




125 13.6 11.0 16.4



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 39.12 ng

RT: 17.66 min Scan# 2682

Delta R.T. -0.05 min

Lab File: BF094062.D

Acq: 30 Mar 2017 17:01


Tgt Ion:278 Resp: 863474

Ion Ratio Lower Upper

278 100

139 22.1 16.6 24.8



279 24.2 19.3 28.9




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91
 Benzo(g,h,i)perylene
 Concen: 38.35 ng
 RT: 18.04 min Scan# 2747
 Delta R.T. -0.05 min
 Lab File: BF094062.D
 Acq: 30 Mar 2017 17:01

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:	276	Resp:	853156
Ion	Ratio	Lower	Upper
276	100		
277	24.6	18.6	28.0
138	31.2	25.4	38.0

