

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF033022\
 Data File : BF127579.D
 Acq On : 30 Mar 2022 10:56
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 30 16:58:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 29 07:13:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2	1,4-Dioxane	0.550	0.580	-5.5	128	-0.02
3	Pyridine	1.553	1.453	6.4	106	-0.01
4	n-Nitrosodimethylamine	0.853	0.808	5.3	108	-0.02
5 S	2-Fluorophenol	1.253	1.151	8.1	105	0.00
6	Aniline	2.044	1.973	3.5	109	0.00
7 S	Phenol-d6	1.713	1.587	7.4	106	0.00
8	2-Chlorophenol	1.280	1.245	2.7	110	0.00
9	Benzaldehyde	0.965	0.856	11.3	107	0.00
10 C	Phenol	1.865	1.765	5.4	107	0.00
11	bis(2-Chloroethyl)ether	1.365	1.354	0.8	113	0.00
12	1,3-Dichlorobenzene	1.440	1.325	8.0	106	0.00
13 C	1,4-Dichlorobenzene	1.449	1.340	7.5	106	0.00
14	1,2-Dichlorobenzene	1.411	1.261	10.6	108	0.00
15	Benzyl Alcohol	1.245	1.186	4.7	107	0.00
16	2,2'-oxybis(1-Chloropropane	2.456	2.430	1.1	112	0.00
17	2-Methylphenol	1.081	1.039	3.9	110	0.00
18	Hexachloroethane	0.530	0.505	4.7	110	0.00
19 P	n-Nitroso-di-n-propylamine	1.027	0.999	2.7	113	0.00
20	3+4-Methylphenols	1.333	1.297	2.7	112	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	118	0.00
22	Acetophenone	0.511	0.458	10.4	111	0.00
23 S	Nitrobenzene-d5	0.468	0.425	9.2	110	0.00
24	Nitrobenzene	0.446	0.424	4.9	114	0.00
25	Isophorone	0.745	0.748	-0.4	119	0.00
26 C	2-Nitrophenol	0.187	0.184	1.6	117	0.00
27	2,4-Dimethylphenol	0.282	0.272	3.5	114	0.00
28	bis(2-Chloroethoxy)methane	0.484	0.479	1.0	119	0.00
29 C	2,4-Dichlorophenol	0.320	0.305	4.7	114	0.00
30	1,2,4-Trichlorobenzene	0.372	0.337	9.4	110	0.00
31	Naphthalene	1.062	0.983	7.4	112	0.00
32	Benzoic acid	0.140	0.145	-3.6	111	0.00
33	4-Chloroaniline	0.411	0.403	1.9	118	0.00
34 C	Hexachlorobutadiene	0.236	0.216	8.5	110	0.00
35	Caprolactam	0.099	0.102	-3.0	119	0.00
36 C	4-Chloro-3-methylphenol	0.347	0.336	3.2	116	0.00
37	2-Methylnaphthalene	0.683	0.655	4.1	116	0.00
38	1-Methylnaphthalene	0.675	0.630	6.7	114	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	121	0.00
40	1,2,4,5-Tetrachlorobenzene	0.636	0.569	10.5	113	0.00
41 P	Hexachlorocyclopentadiene	0.147	0.131	10.9	107	0.00
42 S	2,4,6-Tribromophenol	0.213	0.195	8.5	116	0.00
43 C	2,4,6-Trichlorophenol	0.384	0.373	2.9	118	0.00
44	2,4,5-Trichlorophenol	0.396	0.393	0.8	116	0.00
45 S	2-Fluorobiphenyl	1.370	1.207	11.9	114	0.00
46	1,1'-Biphenyl	1.511	1.364	9.7	115	0.00
47	2-Chloronaphthalene	1.189	1.091	8.2	116	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF033022\
 Data File : BF127579.D
 Acq On : 30 Mar 2022 10:56
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 30 16:58:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 29 07:13:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.456	0.443	2.9	121	0.00
49	Acenaphthylene	1.798	1.618	10.0	112	0.00
50	Dimethylphthalate	1.391	1.325	4.7	120	0.00
51	2,6-Dinitrotoluene	0.292	0.279	4.5	119	0.00
52 C	Acenaphthene	1.049	0.988	5.8	117	0.00
53	3-Nitroaniline	0.320	0.310	3.1	120	0.00
54 P	2,4-Dinitrophenol	0.114	0.120	-5.3	116	0.00
55	Dibenzofuran	1.710	1.464	14.4	113	0.00
56 P	4-Nitrophenol	0.163	0.139	14.7	97	0.00
57	2,4-Dinitrotoluene	0.376	0.338	10.1	112	0.00
58	Fluorene	1.353	1.183	12.6	118	0.00
59	2,3,4,6-Tetrachlorophenol	0.350	0.333	4.9	121	0.00
60	Diethylphthalate	1.271	1.221	3.9	123	0.00
61	4-Chlorophenyl-phenylether	0.719	0.657	8.6	122	0.00
62	4-Nitroaniline	0.300	0.287	4.3	121	0.00
63	Azobenzene	1.500	1.453	3.1	123	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	120	0.00
65	4,6-Dinitro-2-methylphenol	0.123	0.127	-3.3	119	0.00
66 c	n-Nitrosodiphenylamine	0.609	0.578	5.1	119	0.00
67	4-Bromophenyl-phenylether	0.234	0.228	2.6	121	0.00
68	Hexachlorobenzene	0.255	0.242	5.1	118	0.00
69	Atrazine	0.203	0.194	4.4	114	0.00
70 C	Pentachlorophenol	0.082	0.078	4.9	105	0.00
71	Phenanthrene	1.043	0.974	6.6	118	0.00
72	Anthracene	1.035	0.967	6.6	118	0.00
73	Carbazole	0.990	0.932	5.9	118	0.00
74	Di-n-butylphthalate	1.024	1.071	-4.6	129	0.00
75 C	Fluoranthene	1.147	1.088	5.1	118	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	132	0.00
77	Benzidine	0.477	0.436	8.6	134	0.00
78	Pyrene	1.716	1.473	14.2	119	0.00
79 S	Terphenyl-d14	1.222	1.062	13.1	120	0.00
80	Butylbenzylphthalate	0.565	0.576	-1.9	134	0.00
81	Benzo(a)anthracene	1.337	1.253	6.3	127	0.00
82	3,3'-Dichlorobenzidine	0.393	0.362	7.9	126	0.00
83	Chrysene	1.267	1.175	7.3	126	0.00
84	Bis(2-ethylhexyl)phthalate	0.653	0.721	-10.4	144	0.00
85 c	Di-n-octyl phthalate	0.869	1.107	-27.4#	168#	0.00
86 I	Perylene-d12	1.000	1.000	0.0	131	0.00
87	Indeno(1,2,3-cd)pyrene	1.353	1.106	18.3	101	0.00
88	Benzo(b)fluoranthene	1.261	1.206	4.4	124	0.00
89	Benzo(k)fluoranthene	1.307	1.297	0.8	141	0.00
90 C	Benzo(a)pyrene	1.035	0.998	3.6	128	0.00
91	Dibenzo(a,h)anthracene	1.117	0.916	18.0	101	0.00
92	Benzo(g,h,i)perylene	1.155	0.925	19.9	99	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF033022\
Data File : BF127579.D
Acq On : 30 Mar 2022 10:56
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 30 16:58:21 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 29 07:13:31 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1