

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF033022\
 Data File : BF127579.D
 Acq On : 30 Mar 2022 10:56
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 30 16:58:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 29 07:13:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	112	0.00
2	1,4-Dioxane	40.000	42.185	-5.5	128	-0.02
3	Pyridine	40.000	37.430	6.4	106	-0.01
4	n-Nitrosodimethylamine	40.000	37.888	5.3	108	-0.02
5 S	2-Fluorophenol	80.000	73.485	8.1	105	0.00
6	Aniline	40.000	38.619	3.5	109	0.00
7 S	Phenol-d6	80.000	74.128	7.3	106	0.00
8	2-Chlorophenol	40.000	38.908	2.7	110	0.00
9	Benzaldehyde	40.000	38.822	2.9	107	0.00
10 C	Phenol	40.000	37.845	5.4	107	0.00
11	bis(2-Chloroethyl)ether	40.000	39.698	0.8	113	0.00
12	1,3-Dichlorobenzene	40.000	36.815	8.0	106	0.00
13 C	1,4-Dichlorobenzene	40.000	37.000	7.5	106	0.00
14	1,2-Dichlorobenzene	40.000	35.737	10.7	108	0.00
15	Benzyl Alcohol	40.000	38.115	4.7	107	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.581	1.0	112	0.00
17	2-Methylphenol	40.000	38.474	3.8	110	0.00
18	Hexachloroethane	40.000	38.106	4.7	110	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.908	2.7	113	0.00
20	3+4-Methylphenols	40.000	38.914	2.7	112	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	118	0.00
22	Acetophenone	40.000	35.839	10.4	111	0.00
23 S	Nitrobenzene-d5	80.000	72.589	9.3	110	0.00
24	Nitrobenzene	40.000	37.974	5.1	114	0.00
25	Isophorone	40.000	40.175	-0.4	119	0.00
26 C	2-Nitrophenol	40.000	39.308	1.7	117	0.00
27	2,4-Dimethylphenol	40.000	38.646	3.4	114	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.547	1.1	119	0.00
29 C	2,4-Dichlorophenol	40.000	38.145	4.6	114	0.00
30	1,2,4-Trichlorobenzene	40.000	36.233	9.4	110	0.00
31	Naphthalene	40.000	37.021	7.4	112	0.00
32	Benzoic acid	40.000	37.366	6.6	111	0.00
33	4-Chloroaniline	40.000	39.142	2.1	118	0.00
34 C	Hexachlorobutadiene	40.000	36.481	8.8	110	0.00
35	Caprolactam	40.000	41.272	-3.2	119	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.745	3.1	116	0.00
37	2-Methylnaphthalene	40.000	38.370	4.1	116	0.00
38	1-Methylnaphthalene	40.000	37.302	6.7	114	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	121	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	35.760	10.6	113	0.00
41 P	Hexachlorocyclopentadiene	40.000	35.706	10.7	107	0.00
42 S	2,4,6-Tribromophenol	80.000	73.072	8.7	116	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.787	3.0	118	0.00
44	2,4,5-Trichlorophenol	40.000	39.697	0.8	116	0.00
45 S	2-Fluorobiphenyl	80.000	70.492	11.9	114	0.00
46	1,1'-Biphenyl	40.000	36.112	9.7	115	0.00
47	2-Chloronaphthalene	40.000	36.698	8.3	116	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF033022\
 Data File : BF127579.D
 Acq On : 30 Mar 2022 10:56
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 30 16:58:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 29 07:13:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.858	2.9	121	0.00
49	Acenaphthylene	40.000	35.994	10.0	112	0.00
50	Dimethylphthalate	40.000	38.098	4.8	120	0.00
51	2,6-Dinitrotoluene	40.000	38.197	4.5	119	0.00
52 C	Acenaphthene	40.000	37.669	5.8	117	0.00
53	3-Nitroaniline	40.000	38.727	3.2	120	0.00
54 P	2,4-Dinitrophenol	40.000	37.806	5.5	116	0.00
55	Dibenzofuran	40.000	38.182	4.5	113	0.00
56 P	4-Nitrophenol	40.000	33.975	15.1	97	0.00
57	2,4-Dinitrotoluene	40.000	35.991	10.0	112	0.00
58	Fluorene	40.000	38.592	3.5	118	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.037	4.9	121	0.00
60	Diethylphthalate	40.000	38.436	3.9	123	0.00
61	4-Chlorophenyl-phenylether	40.000	36.558	8.6	122	0.00
62	4-Nitroaniline	40.000	38.258	4.4	121	0.00
63	Azobenzene	40.000	38.757	3.1	123	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	120	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.200	-3.0	119	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.987	5.0	119	0.00
67	4-Bromophenyl-phenylether	40.000	38.989	2.5	121	0.00
68	Hexachlorobenzene	40.000	38.001	5.0	118	0.00
69	Atrazine	40.000	38.203	4.5	114	0.00
70 C	Pentachlorophenol	40.000	35.672	10.8	105	0.00
71	Phenanthrene	40.000	37.349	6.6	118	0.00
72	Anthracene	40.000	37.380	6.5	118	0.00
73	Carbazole	40.000	37.680	5.8	118	0.00
74	Di-n-butylphthalate	40.000	41.822	-4.6	129	0.00
75 C	Fluoranthene	40.000	37.958	5.1	118	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	132	0.00
77	Benzydine	40.000	36.535	8.7	134	0.00
78	Pyrene	40.000	34.317	14.2	119	0.00
79 S	Terphenyl-d14	80.000	69.562	13.0	120	0.00
80	Butylbenzylphthalate	40.000	40.780	-2.0	134	0.00
81	Benzo(a)anthracene	40.000	37.468	6.3	127	0.00
82	3,3'-Dichlorobenzidine	40.000	36.853	7.9	126	0.00
83	Chrysene	40.000	37.104	7.2	126	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.147	-10.4	144	0.00
85 c	Di-n-octyl phthalate	40.000	50.964	-27.4#	168	0.00
86 I	Perylene-d12	20.000	20.000	0.0	131	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	32.676	18.3	101	0.00
88	Benzo(b)fluoranthene	40.000	38.260	4.4	124	0.00
89	Benzo(k)fluoranthene	40.000	39.684	0.8	141	0.00
90 C	Benzo(a)pyrene	40.000	38.572	3.6	128	0.00
91	Dibenzo(a,h)anthracene	40.000	32.808	18.0	101	0.00
92	Benzo(g,h,i)perylene	40.000	32.029	19.9	99	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF033022\
Data File : BF127579.D
Acq On : 30 Mar 2022 10:56
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 30 16:58:21 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 29 07:13:31 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1