

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF033116\
 Data File : BF085932.D
 Acq On : 31 Mar 2016 17:25
 Operator : UM/SJ
 Sample : H1954-04
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-4

Quant Time: Mar 31 23:40:53 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 30 12:31:18 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.79	152	109589	20.00	ng	-0.01
21) Naphthalene-d8	8.07	136	439503	20.00	ng	-0.02
38) Acenaphthene-d10	9.83	164	191940	20.00	ng	-0.01
63) Phenanthrene-d10	11.30	188	327214	20.00	ng	-0.02
75) Chrysene-d12	13.95	240	213047	20.00	ng	-0.02
86) Perylene-d12	15.36	264	202708	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.38	112	458372	69.66	ng	-0.01
7) Phenol-d6	6.42	99	398780	47.66	ng	-0.02
23) Nitrobenzene-d5	7.36	82	680091	87.14	ng	-0.01
41) 2,4,6-Tribromophenol	10.62	330	242796	133.14	ng	-0.01
44) 2-Fluorobiphenyl	9.16	172	1134475	98.75	ng	-0.01
78) Terphenyl-d14	12.89	244	762058	73.56	ng	-0.01
Target Compounds						
2) 1,4-Dioxane	2.36	88	52238	16.63	ng	97
49) Dimethylphthalate	9.54	163	31311	2.15	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF033116\
Data File : BF085932.D
Acq On : 31 Mar 2016 17:25
Operator : UM/SJ
Sample : H1954-04
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MW-4

Quant Time: Mar 31 23:40:53 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032416.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 30 12:31:18 2016
Response via : Initial Calibration

