

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF033116\  
 Data File : BF085950.D  
 Acq On : 1 Apr 2016 2:06  
 Operator : UM/SJ  
 Sample : H2010-08  
 Misc :  
 ALS Vial : 25 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 AMBOY-SB23COMP0.5-15.0

Quant Time: Apr 01 03:07:10 2016  
 Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF032416.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Mar 30 12:31:18 2016  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min)       |
|-----------------------------|-------|------|----------|--------|-------|----------------|
| 1) 1,4-Dichlorobenzene-d4   | 6.79  | 152  | 109487   | 20.00  | ng    | -0.01          |
| 21) Naphthalene-d8          | 8.07  | 136  | 423604   | 20.00  | ng    | -0.02          |
| 38) Acenaphthene-d10        | 9.83  | 164  | 191870   | 20.00  | ng    | -0.01          |
| 63) Phenanthrene-d10        | 11.32 | 188  | 327442   | 20.00  | ng    | -0.01          |
| 75) Chrysene-d12            | 13.95 | 240  | 203041   | 20.00  | ng    | -0.02          |
| 86) Perylene-d12            | 15.36 | 264  | 195820   | 20.00  | ng    | -0.01          |
| System Monitoring Compounds |       |      |          |        |       |                |
| 5) 2-Fluorophenol           | 5.40  | 112  | 853054   | 129.76 | ng    | 0.00           |
| 7) Phenol-d6                | 6.44  | 99   | 1098761  | 131.45 | ng    | -0.01          |
| 23) Nitrobenzene-d5         | 7.36  | 82   | 704677   | 93.68  | ng    | -0.01          |
| 41) 2,4,6-Tribromophenol    | 10.62 | 330  | 225513   | 123.71 | ng    | -0.01          |
| 44) 2-Fluorobiphenyl        | 9.16  | 172  | 1133978  | 98.74  | ng    | -0.01          |
| 78) Terphenyl-d14           | 12.89 | 244  | 742254   | 75.18  | ng    | -0.01          |
| Target Compounds            |       |      |          |        |       |                |
| 49) Dimethylphthalate       | 9.54  | 163  | 164340   | 11.29  | ng    | Qvalue<br># 99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF033116\  
Data File : BF085950.D  
Acq On : 1 Apr 2016 2:06  
Operator : UM/SJ  
Sample : H2010-08  
Misc :  
ALS Vial : 25 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
AMBOY-SB23COMP0.5-15.0

Quant Time: Apr 01 03:07:10 2016  
Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF032416.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Mar 30 12:31:18 2016  
Response via : Initial Calibration

