

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF033125\
 Data File : BF142194.D
 Acq On : 31 Mar 2025 21:55
 Operator : RC/JU
 Sample : Q1671-01
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-1

Manual Integrations
 APPROVED

Reviewed By :Anahy Claudio 04/01/2025
 Supervised By :Jagrut Upadhyay 04/01/2025

Quant Time: Apr 01 00:49:20 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 15:46:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.863	152	99301	20.000	ng	-0.02
21) Naphthalene-d8	8.145	136	377149	20.000	ng	-0.02
39) Acenaphthene-d10	9.904	164	189878	20.000	ng	-0.01
64) Phenanthrene-d10	11.392	188	279944m	20.000	ng	-0.01
76) Chrysene-d12	14.033	240	246075	20.000	ng	# 0.00
86) Perylene-d12	15.504	264	295023	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.481	112	505066	84.887	ng	-0.01
7) Phenol-d6	6.487	99	658222	86.888	ng	-0.02
23) Nitrobenzene-d5	7.422	82	399134	59.556	ng	-0.02
42) 2,4,6-Tribromophenol	10.692	330	193201	80.205	ng	-0.01
45) 2-Fluorobiphenyl	9.222	172	853373	68.350	ng	-0.02
79) Terphenyl-d14	12.974	244	741560	44.554	ng	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF033125\
Data File : BF142194.D
Acq On : 31 Mar 2025 21:55
Operator : RC/JU
Sample : Q1671-01
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-1

Quant Time: Apr 01 00:49:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 04/01/2025
Supervised By :Jagrut Upadhyay 04/01/2025

