




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF033125\
Data File : BF142174.D
Acq On : 31 Mar 2025 11:53
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 31 13:27:35 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration



TIC: BF142174.D\data.ms

(70) Pentachlorophenol (C)




11.192min (-0.012) 85074.56 ng

response 205770

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.80	64.05
264.00	62.70	63.34
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF033125\
Data File : BF142174.D
Acq On : 31 Mar 2025 11:53
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 31 13:27:35 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration



TIC: BF142174.D\data.ms

(77) Benzidine

12.727min (-0.006) 153700.47 ng

response 802010

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	13.89
183.00	12.10	11.96
0.00	0.00	0.00