

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF040122\
 Data File : BF127607.D
 Acq On : 01 Apr 2022 12:38
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 01 16:12:43 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 29 07:13:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	120	0.00
2	1,4-Dioxane	40.000	46.719	-16.8	153	0.00
3	Pyridine	40.000	41.933	-4.8	127	0.00
4	n-Nitrosodimethylamine	40.000	40.208	-0.5	123	0.00
5 S	2-Fluorophenol	80.000	80.104	-0.1	123	0.00
6	Aniline	40.000	39.985	0.0	121	0.00
7 S	Phenol-d6	80.000	77.745	2.8	120	0.00
8	2-Chlorophenol	40.000	40.000	0.0	121	0.00
9	Benzaldehyde	40.000	39.924	0.2	118	0.00
10 C	Phenol	40.000	39.494	1.3	120	0.00
11	bis(2-Chloroethyl)ether	40.000	41.863	-4.7	128	0.00
12	1,3-Dichlorobenzene	40.000	38.454	3.9	119	0.00
13 C	1,4-Dichlorobenzene	40.000	38.696	3.3	119	0.00
14	1,2-Dichlorobenzene	40.000	36.527	8.7	118	0.00
15	Benzyl Alcohol	40.000	37.461	6.3	113	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.327	-0.8	123	0.00
17	2-Methylphenol	40.000	39.371	1.6	121	0.00
18	Hexachloroethane	40.000	39.856	0.4	123	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.936	2.7	122	0.00
20	3+4-Methylphenols	40.000	39.335	1.7	122	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	122	0.00
22	Acetophenone	40.000	37.433	6.4	120	0.00
23 S	Nitrobenzene-d5	80.000	76.892	3.9	121	0.00
24	Nitrobenzene	40.000	39.255	1.9	122	0.00
25	Isophorone	40.000	41.288	-3.2	126	0.00
26 C	2-Nitrophenol	40.000	40.317	-0.8	124	0.00
27	2,4-Dimethylphenol	40.000	40.264	-0.7	123	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.411	-3.5	129	0.00
29 C	2,4-Dichlorophenol	40.000	39.756	0.6	123	0.00
30	1,2,4-Trichlorobenzene	40.000	37.761	5.6	119	0.00
31	Naphthalene	40.000	38.696	3.3	121	0.00
32	Benzoic acid	40.000	36.711	8.2	113	0.00
33	4-Chloroaniline	40.000	40.156	-0.4	125	0.00
34 C	Hexachlorobutadiene	40.000	37.685	5.8	118	0.00
35	Caprolactam	40.000	42.185	-5.5	126	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.948	0.1	124	0.00
37	2-Methylnaphthalene	40.000	39.529	1.2	123	0.00
38	1-Methylnaphthalene	40.000	37.897	5.3	120	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	120	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.790	5.5	119	0.00
41 P	Hexachlorocyclopentadiene	40.000	36.519	8.7	111	0.00
42 S	2,4,6-Tribromophenol	80.000	74.291	7.1	117	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.130	2.2	119	0.00
44	2,4,5-Trichlorophenol	40.000	40.223	-0.6	117	0.00
45 S	2-Fluorobiphenyl	80.000	73.336	8.3	118	0.00
46	1,1'-Biphenyl	40.000	38.093	4.8	121	0.00
47	2-Chloronaphthalene	40.000	38.345	4.1	120	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF040122\
 Data File : BF127607.D
 Acq On : 01 Apr 2022 12:38
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 01 16:12:43 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 29 07:13:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.561	-1.4	125	0.00
49	Acenaphthylene	40.000	37.238	6.9	115	0.00
50	Dimethylphthalate	40.000	38.978	2.6	122	0.00
51	2,6-Dinitrotoluene	40.000	38.554	3.6	119	0.00
52 C	Acenaphthene	40.000	39.105	2.2	121	0.00
53	3-Nitroaniline	40.000	40.058	-0.1	124	0.00
54 P	2,4-Dinitrophenol	40.000	36.807	8.0	111	0.00
55	Dibenzofuran	40.000	38.971	2.6	115	0.00
56 P	4-Nitrophenol	40.000	44.213	-10.5	126	0.01
57	2,4-Dinitrotoluene	40.000	35.958	10.1	112	0.00
58	Fluorene	40.000	39.781	0.5	121	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.321	4.2	121	0.00
60	Diethylphthalate	40.000	38.859	2.9	124	0.00
61	4-Chlorophenyl-phenylether	40.000	36.659	8.4	122	0.00
62	4-Nitroaniline	40.000	39.866	0.3	125	0.00
63	Azobenzene	40.000	39.601	1.0	125	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	119	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.424	-3.6	119	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.213	2.0	122	0.00
67	4-Bromophenyl-phenylether	40.000	39.710	0.7	122	0.00
68	Hexachlorobenzene	40.000	38.002	5.0	117	0.00
69	Atrazine	40.000	38.403	4.0	114	0.00
70 C	Pentachlorophenol	40.000	36.381	9.0	107	0.00
71	Phenanthrene	40.000	38.082	4.8	119	0.00
72	Anthracene	40.000	38.349	4.1	119	0.00
73	Carbazole	40.000	38.821	2.9	120	0.00
74	Di-n-butylphthalate	40.000	43.243	-8.1	132	0.00
75 C	Fluoranthene	40.000	38.503	3.7	119	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	121	0.00
77	Benzidine	40.000	43.089	-7.7	145	0.00
78	Pyrene	40.000	37.422	6.4	119	0.00
79 S	Terphenyl-d14	80.000	74.216	7.2	118	0.00
80	Butylbenzylphthalate	40.000	44.043	-10.1	132	0.00
81	Benzo(a)anthracene	40.000	39.086	2.3	122	0.00
82	3,3'-Dichlorobenzidine	40.000	39.814	0.5	125	0.00
83	Chrysene	40.000	39.201	2.0	123	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	47.273	-18.2	142	0.00
85 c	Di-n-octyl phthalate	40.000	55.248	-38.1#	167	0.00
86 I	Perylene-d12	20.000	20.000	0.0	136	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	35.893	10.3	115	0.01
88	Benzo(b)fluoranthene	40.000	39.443	1.4	133	0.00
89	Benzo(k)fluoranthene	40.000	37.411	6.5	138	0.00
90 C	Benzo(a)pyrene	40.000	39.377	1.6	136	0.01
91	Dibenzo(a,h)anthracene	40.000	36.890	7.8	118	0.01
92	Benzo(g,h,i)perylene	40.000	35.859	10.4	115	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF040122\
Data File : BF127607.D
Acq On : 01 Apr 2022 12:38
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 01 16:12:43 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 29 07:13:31 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1