

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040215\
 Data File : BF078208.D
 Acq On : 2 Apr 2015 19:15
 Operator : TP/IZ
 Sample : G1710-10
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 W-22

Quant Time: Apr 03 01:17:22 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 03 00:57:15 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.01	152	41721	20.00	ng	0.00
21) Naphthalene-d8	8.59	136	169919	20.00	ng	0.00
38) Acenaphthene-d10	10.76	164	85113	20.00	ng	0.00
63) Phenanthrene-d10	12.59	188	166284	20.00	ng	0.00
75) Chrysene-d12	15.87	240	169030	20.00	ng	-0.01
86) Perylene-d12	17.62	264	166321	20.00	ng	-0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.30	112	133943	52.93	ng	0.01
7) Phenol-d6	6.60	99	102694	32.38	ng	0.00
23) Nitrobenzene-d5	7.71	82	234494	83.65	ng	-0.01
41) 2,4,6-Tribromophenol	11.74	330	100139	141.86	ng	0.00
44) 2-Fluorobiphenyl	9.94	172	505967	84.97	ng	0.00
78) Terphenyl-d14	14.59	244	575697	76.96	ng	0.00
Target Compounds						
49) Dimethylphthalate	10.45	163	13909	2.18	ng	Qvalue # 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040215\
Data File : BF078208.D
Acq On : 2 Apr 2015 19:15
Operator : TP/IZ
Sample : G1710-10
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
W-22

Quant Time: Apr 03 01:17:22 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 03 00:57:15 2015
Response via : Initial Calibration

