


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF040216\
 Data File : BF085962.D
 Acq On : 1 Apr 2016 15:27
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF040216.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Apr 01 23:17:06 2016



AutoFind: Scans 827, 828, 829; Background Corrected with Scan 824

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.8	12865	PASS
68	69	0.00	2	1.4	173	PASS
69	198	0.00	100	37.5	12416	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	54.8	18173	PASS
197	198	0.00	2	0.6	186	PASS
198	198	100	100	100.0	33138	PASS
199	198	5	9	6.5	2141	PASS
275	198	10	60	22.7	7509	PASS
365	198	1	100	2.7	886	PASS
441	442	0.01	24	15.6	4223	PASS
442	198	50	100	81.6	27056	PASS
443	442	15	24	18.9	5105	PASS