


Data Path : Z:\HPCHEM1\BNA F\DATA\BF040216\
 Data File : BF086023.D
 Acq On : 2 Apr 2016 20:40
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF040216.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Apr 01 23:17:06 2016



AutoFind: Scans 827, 828, 829; Background Corrected with Scan 824

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.1	13021	PASS
68	69	0.00	2	1.7	223	PASS
69	198	0.00	100	36.7	12880	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	54.6	19161	PASS
197	198	0.00	2	0.6	200	PASS
198	198	100	100	100.0	35099	PASS
199	198	5	9	7.2	2524	PASS
275	198	10	60	22.8	7995	PASS
365	198	1	100	2.7	958	PASS
441	442	0.01	24	15.5	4071	PASS
442	198	50	100	74.9	26296	PASS
443	442	15	24	19.0	4995	PASS