

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF040218\
 Data File : BF104173.D
 Acq On : 3 Apr 2018 10:32
 Operator : JU/SJ
 Sample : J2173-05DL 5X
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 QNWP8-MW24D-GWDL

Manual Integrations
 APPROVED

Sohil
 4/3/2018 2:28:48 PM

Quant Time: Apr 03 12:50:09 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 02 13:40:00 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.96	152	76990	20.00	ng	0.00
21) Naphthalene-d8	8.25	136	322668	20.00	ng	0.00
38) Acenaphthene-d10	10.00	164	129736	20.00	ng	0.00
63) Phenanthrene-d10	11.50	188	206912	20.00	ng	0.00
75) Chrysene-d12	14.15	240	192873	20.00	ng	0.00
86) Perylene-d12	15.69	264	127738	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.61	112	48592m	10.07	ng	0.02
7) Phenol-d6	6.62	99	36983m	6.23	ng	0.02
23) Nitrobenzene-d5	7.53	82	78806	14.94	ng	0.00
41) 2,4,6-Tribromophenol	10.80	330	32615	22.58	ng	0.00
44) 2-Fluorobiphenyl	9.31	172	202143	24.57	ng	0.00
78) Terphenyl-d14	13.07	244	157337	18.84	ng	0.00
Target Compounds						
31) Naphthalene	8.26	128	1138697	72.237	ng	99
51) Acenaphthene	10.03	154	28008	3.639	ng	95
54) Dibenzofuran	10.21	168	28578	2.542	ng	98
57) Fluorene	10.55	166	24098	2.599	ng	96
70) Phenanthrene	11.52	178	43265	3.862	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF040218\
Data File : BF104173.D
Acq On : 3 Apr 2018 10:32
Operator : JU/SJ
Sample : J2173-05DL 5X
Misc :
ALS Vial : 48 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
QNWP8-MW24D-GWDL

Manual Integrations
APPROVED

Sohil
4/3/2018 2:28:48 PM

Quant Time: Apr 03 12:50:09 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 02 13:40:00 2018
Response via : Initial Calibration

