


Benzaldehyde


Response Ratio



$R = -1.557e-001 A^2 + 1.146e+000 A + 1.105e-002$
Coef of Det (r^2) = 0.998992 Curve Fit: Quadratic
Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF040221.M
Calibration Table Last Updated: Fri Apr 02 01:16:47 2021

2,2'-oxybis(1-Chloropropane)

Response Ratio



$$\text{Response} = 2.119\text{e}+000 * \text{Amt} + 4.818\text{e}-001$$


Coef of Det (r^2) = 0.994997 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF040221.M

Calibration Table Last Updated: Fri Apr 02 01:16:47 2021

Carbazole

Response Ratio



$$\text{Response} = 9.105\text{e-}001 * \text{Amt} + 1.267\text{e-}001$$


Coef of Det (r^2) = 0.994123 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF040221.M

Calibration Table Last Updated: Fri Apr 02 01:16:47 2021

Di-n-butylphthalate

Response Ratio



$$\text{Response} = 1.199\text{e}+000 * \text{Amt} + 2.116\text{e}-001$$

Coef of Det (r^2) = 0.991676 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF040221.M

Calibration Table Last Updated: Fri Apr 02 01:16:47 2021