



Data Path : Z:\HPCHEM1\BNA F\DATA\BF040315\
Data File : BF078239.D
Acq On : 3 Apr 2015 13:39
Operator : TP/IZ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 06 13:02:25 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 03 14:42:25 2015
Response via : Initial Calibration



TIC: BF078239.D

(69) Pentachlorophenol (C)



12.328min (-0.000) 0.00ng

response 22265

Ion	Exp%	Act%
265.70	100	100
268.00	62.60	62.94
264.00	62.50	64.86
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040315\
Data File : BF078239.D
Acq On : 3 Apr 2015 13:39
Operator : TP/IZ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 06 13:02:25 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 03 14:42:25 2015
Response via : Initial Calibration



TIC: BF078239.D

(76) Benzidine

14.248min (+0.000) 0.00ng

response 265509

Ion	Exp%	Act%
184.00	100	100
185.00	15.20	13.58
183.00	11.90	11.45
0.00	0.00	0.00