


Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF040318\  
 Data File : BF104174.D  
 Acq On : 3 Apr 2018 11:00  
 Operator : JU/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF032818.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Tue Apr 03 15:22:43 2018



AutoFind: Scans 1637, 1638, 1639; Background Corrected with Scan 1633

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 31.4      | 18261   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.0       | 187     | PASS             |
| 69          | 198          | 0.00         | 100          | 33.7      | 19596   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 49.4      | 28784   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.2       | 118     | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 58226   | PASS             |
| 199         | 198          | 5            | 9            | 6.8       | 3955    | PASS             |
| 275         | 198          | 10           | 60           | 26.1      | 15212   | PASS             |
| 365         | 198          | 1            | 100          | 3.1       | 1815    | PASS             |
| 441         | 443          | 0.01         | 100          | 77.5      | 8007    | PASS             |
| 442         | 198          | 50           | 100          | 90.2      | 52498   | PASS             |
| 443         | 442          | 15           | 24           | 19.7      | 10331   | PASS             |