


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF040417\
 Data File : BF094168.D
 Acq On : 4 Apr 2017 15:29
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032217.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Mar 31 14:38:29 2017



AutoFind: Scans 1673, 1674, 1675; Background Corrected with Scan 1667

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.9	44290	PASS
68	69	0.00	2	1.9	817	PASS
69	198	0.00	100	36.0	43181	PASS
70	69	0.00	2	0.3	143	PASS
127	198	10	80	50.7	60872	PASS
197	198	0.00	2	0.4	482	PASS
198	198	100	100	100.0	120056	PASS
199	198	5	9	6.6	7950	PASS
275	198	10	60	22.1	26485	PASS
365	198	1	100	2.6	3179	PASS
441	442	0.01	24	15.4	17615	PASS
442	198	50	100	95.3	114376	PASS
443	442	15	24	19.2	21923	PASS