


Data Path : Z:\HPCHEM1\BNA_F\Data\BF040418\
 Data File : BF104223.D
 Acq On : 4 Apr 2018 13:51
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032818.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Apr 04 18:06:58 2018



AutoFind: Scans 1635, 1636, 1637; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.1	25319	PASS
68	69	0.00	2	1.4	386	PASS
69	198	0.00	100	34.6	27338	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.8	39296	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	78901	PASS
199	198	5	9	7.1	5585	PASS
275	198	10	60	27.0	21265	PASS
365	198	1	100	3.4	2687	PASS
441	443	0.01	100	81.3	11496	PASS
442	198	50	100	93.2	73549	PASS
443	442	15	24	19.2	14136	PASS