

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040516\
 Data File : BF086086.D
 Acq On : 5 Apr 2016 21:00
 Operator : UM/SJ
 Sample : H2028-01RE
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 W1DAY1RE

Quant Time: Apr 06 01:17:28 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 01 23:17:06 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.75	152	122154	20.00	ng	-0.02
21) Naphthalene-d8	8.04	136	488762	20.00	ng	-0.02
38) Acenaphthene-d10	9.79	164	230406	20.00	ng	-0.02
63) Phenanthrene-d10	11.28	188	427810	20.00	ng	-0.02
75) Chrysene-d12	13.91	240	287944	20.00	ng	-0.02
86) Perylene-d12	15.31	264	236419	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.36	112	31612	4.42	ng	-0.01
7) Phenol-d6	6.40	99	43051	4.74	ng	-0.02
23) Nitrobenzene-d5	7.32	82	33299	4.02	ng	-0.02
41) 2,4,6-Tribromophenol	10.58	330	4557	2.11	ng	-0.02
44) 2-Fluorobiphenyl	9.12	172	43908	3.17	ng	-0.02
78) Terphenyl-d14	12.85	244	31422	2.57	ng	-0.02
Target Compounds						
10) Phenol	6.41	94	113348	10.95	ng	Qvalue 98
32) Benzoic acid	7.81	122	30961	5.42	ng	97
49) Dimethylphthalate	9.52	163	97244	5.69	ng	99
83) Bis(2-ethylhexyl)phthalate	13.89	149	26945	2.22	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040516\
Data File : BF086086.D
Acq On : 5 Apr 2016 21:00
Operator : UM/SJ
Sample : H2028-01RE
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
W1DAY1RE

Quant Time: Apr 06 01:17:28 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 01 23:17:06 2016
Response via : Initial Calibration

