


Data Path : Z:\HPCHEM1\BNA F\DATA\BF040516\
 Data File : BF086080.D
 Acq On : 5 Apr 2016 16:58
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF040216.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Apr 01 23:17:06 2016



AutoFind: Scans 826, 827, 828; Background Corrected with Scan 822

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.8	14347	PASS
68	69	0.00	2	1.0	136	PASS
69	198	0.00	100	35.3	13752	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	53.1	20712	PASS
197	198	0.00	2	0.3	106	PASS
198	198	100	100	100.0	39000	PASS
199	198	5	9	6.8	2638	PASS
275	198	10	60	24.8	9663	PASS
365	198	1	100	3.0	1170	PASS
441	442	0.01	24	15.7	5319	PASS
442	198	50	100	86.9	33904	PASS
443	442	15	24	18.3	6204	PASS