

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040517\
 Data File : BF094203.D
 Acq On : 5 Apr 2017 21:38
 Operator : SJ/MA
 Sample : I2522-15
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 BASING-8(0-2)

Manual Integrations
 APPROVED

mohammad
 4/6/2017 3:38:36 PM

Quant Time: Apr 06 02:21:20 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 31 14:38:29 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.86	152	177989	20.00	ng	-0.04
21) Naphthalene-d8	7.36	136	714101	20.00	ng	-0.05
38) Acenaphthene-d10	9.50	164	308685	20.00	ng	-0.05
63) Phenanthrene-d10	11.33	188	502979	20.00	ng	-0.05
75) Chrysene-d12	14.57	240	344620	20.00	ng	-0.05
86) Perylene-d12	16.20	264	300609	20.00	ng	-0.05
System Monitoring Compounds						
5) 2-Fluorophenol	4.40	112	931378	88.38	ng	-0.04
7) Phenol-d6	5.47	99	1192849	91.50	ng	-0.04
23) Nitrobenzene-d5	6.51	82	731359	58.45	ng	-0.05
41) 2,4,6-Tribromophenol	10.47	330	210960	86.48	ng	-0.05
44) 2-Fluorobiphenyl	8.69	172	1240205	61.28	ng	-0.05
78) Terphenyl-d14	13.30	244	781382	50.82	ng	-0.05
Target Compounds						Qvalue
49) Dimethylphthalate	9.18	163	113270	5.16	ng	97
70) Phenanthrene	11.35	178	87567	3.13	ng	97
74) Fluoranthene	12.82	202	140003	4.89	ng	98
77) Pyrene	13.09	202	118525	4.74	ng	97
80) Benzo(a)anthracene	14.56	228	50778	2.53	ng	98
82) Chrysene	14.60	228	53769	2.88	ng	99
87) Benzo(b)fluoranthene	15.80	252	62838m	3.41	ng	
89) Benzo(a)pyrene	16.14	252	40469	2.38	ng	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040517\
Data File : BF094203.D
Acq On : 5 Apr 2017 21:38
Operator : SJ/MA
Sample : I2522-15
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
BASING-8(0-2)

Manual Integrations
APPROVED

mohammad
4/6/2017 3:38:36 PM

Quant Time: Apr 06 02:21:20 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 31 14:38:29 2017
Response via : Initial Calibration

