

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040517\
 Data File : BF094215.D
 Acq On : 6 Apr 2017 4:08
 Operator : SJ/MA
 Sample : I2536-01
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 E2-BB7-BASE

Quant Time: Apr 06 05:11:44 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 31 14:38:29 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.86	152	162395	20.00	ng	-0.04
21) Naphthalene-d8	7.37	136	618776	20.00	ng	-0.04
38) Acenaphthene-d10	9.51	164	253884	20.00	ng	-0.04
63) Phenanthrene-d10	11.33	188	384049	20.00	ng	-0.04
75) Chrysene-d12	14.59	240	326665	20.00	ng	-0.04
86) Perylene-d12	16.21	264	231540	20.00	ng	-0.04
System Monitoring Compounds						
5) 2-Fluorophenol	4.40	112	1296053	134.80	ng	-0.03
7) Phenol-d6	5.47	99	1649343	138.67	ng	-0.04
23) Nitrobenzene-d5	6.52	82	980905	90.47	ng	-0.04
41) 2,4,6-Tribromophenol	10.49	330	262330	130.76	ng	-0.04
44) 2-Fluorobiphenyl	8.69	172	1454574	87.39	ng	-0.04
78) Terphenyl-d14	13.32	244	894717	61.39	ng	-0.04
Target Compounds						Qvalue
10) Phenol	5.49	94	128970	9.53	ng	90
17) 2-Methylphenol	6.17	107	21936	2.42	ng	97
20) 3+4-Methylphenols	6.35	107	31621	2.88	ng	# 62
27) 2,4-Dimethylphenol	6.97	122	62968	7.17	ng	93
31) Naphthalene	7.39	128	206266	6.93	ng	100
37) 2-Methylnaphthalene	8.23	142	82035	4.14	ng	99
49) Dimethylphthalate	9.19	163	161392	8.94	ng	98
70) Phenanthrene	11.36	178	72084	3.37	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040517\
Data File : BF094215.D
Acq On : 6 Apr 2017 4:08
Operator : SJ/MA
Sample : I2536-01
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
E2-BB7-BASE

Quant Time: Apr 06 05:11:44 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 31 14:38:29 2017
Response via : Initial Calibration

