


Data Path : Z:\HPCHEM1\BNA F\Data\BF040517\
 Data File : BF094187.D
 Acq On : 5 Apr 2017 13:36
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032217.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Mar 31 14:38:29 2017



AutoFind: Scans 1670, 1671, 1672; Background Corrected with Scan 1664

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.8	46752	PASS
68	69	0.00	2	2.0	944	PASS
69	198	0.00	100	37.6	46554	PASS
70	69	0.00	2	0.3	129	PASS
127	198	10	80	49.7	61437	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	123720	PASS
199	198	5	9	6.4	7931	PASS
275	198	10	60	22.9	28352	PASS
365	198	1	100	2.8	3435	PASS
441	442	0.01	24	16.1	18933	PASS
442	198	50	100	95.0	117594	PASS
443	442	15	24	18.8	22166	PASS