

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF040519\
 Data File : BF113636.D
 Acq On : 6 Apr 2019 1:09
 Operator : JU/SJ
 Sample : K2240-08
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-03-(0-6)

Manual Integrations
 APPROVED

Sohil
 4/8/2019 9:48:23 AM

Quant Time: Apr 06 03:36:17 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF040319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 04 04:05:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.69	152	111582	20.00	ng	0.00
21) Naphthalene-d8	7.97	136	410263	20.00	ng	0.00
39) Acenaphthene-d10	9.72	164	183753	20.00	ng	0.00
64) Phenanthrene-d10	11.20	188	349481	20.00	ng	0.00
76) Chrysene-d12	13.83	240	228752	20.00	ng	-0.01
87) Perylene-d12	15.24	264	173931	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.30	112	629793	96.25	ng	0.00
7) Phenol-d6	6.34	99	818679	101.51	ng	0.00
23) Nitrobenzene-d5	7.25	82	424695	60.64	ng	0.00
42) 2,4,6-Tribromophenol	10.51	330	205182	92.96	ng	0.00
45) 2-Fluorobiphenyl	9.04	172	842038	73.77	ng	-0.01
79) Terphenyl-d14	12.78	244	777731	69.22	ng	-0.01
Target Compounds						
50) Dimethylphthalate	9.43	163	116058	8.568	ng	Qvalue 100
75) Fluoranthene	12.41	202	58775	3.158	ng	98
78) Pyrene	12.63	202	51291	2.947	ng	97
81) Benzo(a)anthracene	13.82	228	26602	2.016	ng	91
83) Chrysene	13.86	228	30468m	2.278	ng	
88) Benzo(b)fluoranthene	14.84	252	37847m	3.555	ng	
90) Benzo(a)pyrene	15.18	252	22015	2.327	ng	# 60

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF040519\
Data File : BF113636.D
Acq On : 6 Apr 2019 1:09
Operator : JU/SJ
Sample : K2240-08
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-03-(0-6)

Manual Integrations
APPROVED

Sohil
4/8/2019 9:48:23 AM

Quant Time: Apr 06 03:36:17 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF040319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 04 04:05:31 2019
Response via : Initial Calibration

