

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF040519\
 Data File : BF113638.D
 Acq On : 6 Apr 2019 2:06
 Operator : JU/SJ
 Sample : K2256-08 10X
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CLEARWELL

Manual Integrations
 APPROVED

Sohil
 4/8/2019 9:48:28 AM

Quant Time: Apr 06 06:55:59 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF040319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 04 04:05:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.69	152	104647	20.00	ng	0.00
21) Naphthalene-d8	7.97	136	388962	20.00	ng	0.00
39) Acenaphthene-d10	9.72	164	171896	20.00	ng	0.00
64) Phenanthrene-d10	11.20	188	316923	20.00	ng	0.00
76) Chrysene-d12	13.83	240	206109	20.00	ng	-0.01
87) Perylene-d12	15.24	264	166402	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.32	112	22308m	3.64	ng	0.02
7) Phenol-d6	6.36	99	19015	2.51	ng	0.02
23) Nitrobenzene-d5	7.25	82	45768	6.89	ng	0.00
42) 2,4,6-Tribromophenol	10.51	330	13918	6.74	ng	0.00
45) 2-Fluorobiphenyl	9.04	172	120987	11.33	ng	-0.01
79) Terphenyl-d14	12.77	244	102695	10.14	ng	-0.02
Target Compounds						
15) Benzyl Alcohol	6.80	79	64441	12.600	ng	Qvalue # 56

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF040519\
Data File : BF113638.D
Acq On : 6 Apr 2019 2:06
Operator : JU/SJ
Sample : K2256-08 10X
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CLEARWELL

Manual Integrations
APPROVED

Sohil
4/8/2019 9:48:28 AM

Quant Time: Apr 06 06:55:59 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF040319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 04 04:05:31 2019
Response via : Initial Calibration

