

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040615\
 Data File : BF078287.D
 Acq On : 7 Apr 2015 7:02
 Operator : TP/IZ
 Sample : G1760-01
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DCW-WC-065-40115

Quant Time: Apr 07 07:07:02 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 07 04:11:17 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.98	152	39101	20.00	ng	0.00
21) Naphthalene-d8	8.56	136	160775	20.00	ng	0.00
38) Acenaphthene-d10	10.72	164	81882	20.00	ng	0.00
63) Phenanthrene-d10	12.54	188	152986	20.00	ng	0.00
75) Chrysene-d12	15.81	240	168161	20.00	ng	0.00
86) Perylene-d12	17.53	264	169099	20.00	ng	0.06
System Monitoring Compounds						
5) 2-Fluorophenol	5.26	112	241171	101.69	ng	0.01
7) Phenol-d6	6.58	99	319957	107.65	ng	0.00
23) Nitrobenzene-d5	7.68	82	184221	69.45	ng	0.00
41) 2,4,6-Tribromophenol	11.70	330	74105	109.12	ng	0.00
44) 2-Fluorobiphenyl	9.90	172	403932	70.51	ng	0.00
78) Terphenyl-d14	14.55	244	462938	62.21	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	8.58	128	618046	73.86	ng	99
37) 2-Methylnaphthalene	9.44	142	140157	24.67	ng	98
45) 1,1'-Biphenyl	10.02	154	14415	2.15	ng	98
48) Acenaphthylene	10.54	152	18558	2.18	ng	# 95
49) Dimethylphthalate	10.42	163	24886	4.05	ng	# 96
51) Acenaphthene	10.75	154	31811	6.64	ng	98
54) Dibenzofuran	10.97	168	39891	5.45	ng	99
57) Fluorene	11.39	166	89469	15.08	ng	99
70) Phenanthrene	12.58	178	312424	35.30	ng	100
71) Anthracene	12.64	178	135929	15.59	ng	99
72) Carbazole	12.85	167	89695	10.98	ng	99
74) Fluoranthene	14.04	202	143760	15.40	ng	96
77) Pyrene	14.32	202	136000	12.88	ng	# 96
80) Benzo(a)anthracene	15.80	228	69553	6.95	ng	93
82) Chrysene	15.85	228	75364	8.02	ng	97
87) Benzo(b)fluoranthene	17.09	252	49248m	4.71	ng	
89) Benzo(a)pyrene	17.46	252	38520	4.22	ng	# 95
91) Benzo(g,h,i)perylene	19.19	276	19239	2.10	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040615\
Data File : BF078287.D
Acq On : 7 Apr 2015 7:02
Operator : TP/IZ
Sample : G1760-01
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DCW-WC-065-40115

Quant Time: Apr 07 07:07:02 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 07 04:11:17 2015
Response via : Initial Calibration

