

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040616\
 Data File : BF086116.D
 Acq On : 6 Apr 2016 23:53
 Operator : UM/SJ
 Sample : H2096-02 5X
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TEC-B8COMP

Manual Integrations
 APPROVED

Sohil
 4/7/2016 5:59:06 PM

Quant Time: Apr 07 15:18:17 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 01 23:17:06 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.74	152	97883	20.00	ng	-0.03
21) Naphthalene-d8	8.03	136	364593	20.00	ng	-0.03
38) Acenaphthene-d10	9.78	164	132447	20.00	ng	-0.03
63) Phenanthrene-d10	11.26	188	231137	20.00	ng	-0.03
75) Chrysene-d12	13.91	240	221052	20.00	ng	-0.02
86) Perylene-d12	15.31	264	150842	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.33	112	138444	24.18	ng	-0.03
7) Phenol-d6	6.40	99	166679	22.92	ng	-0.02
23) Nitrobenzene-d5	7.31	82	101200	16.37	ng	-0.03
41) 2,4,6-Tribromophenol	10.57	330	19143	15.40	ng	-0.03
44) 2-Fluorobiphenyl	9.10	172	162337	20.38	ng	-0.03
78) Terphenyl-d14	12.84	244	116078	12.34	ng	-0.03
Target Compounds						
74) Fluoranthene	12.48	202	40117	3.06	ng	Qvalue 93
77) Pyrene	12.71	202	45474	2.92	ng	98
80) Benzo(a)anthracene	13.89	228	34448	2.64	ng	# 83
82) Chrysene	13.93	228	36695m	2.92	ng	
87) Benzo(b)fluoranthene	14.91	252	31631m	3.33	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040616\
Data File : BF086116.D
Acq On : 6 Apr 2016 23:53
Operator : UM/SJ
Sample : H2096-02 5X
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TEC-B8COMP

Manual Integrations
APPROVED

Sohil
4/7/2016 5:59:06 PM

Quant Time: Apr 07 15:18:17 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 01 23:17:06 2016
Response via : Initial Calibration

