


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF040616\
 Data File : BF086088.D
 Acq On : 6 Apr 2016 10:24
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF040216.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Apr 01 23:17:06 2016



AutoFind: Scans 825, 826, 827; Background Corrected with Scan 822

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.8	13858	PASS
68	69	0.00	2	1.0	136	PASS
69	198	0.00	100	33.7	13816	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	51.1	20936	PASS
197	198	0.00	2	0.5	206	PASS
198	198	100	100	100.0	40976	PASS
199	198	5	9	6.6	2684	PASS
275	198	10	60	23.5	9631	PASS
365	198	1	100	2.2	909	PASS
441	442	0.01	24	15.6	4798	PASS
442	198	50	100	75.2	30804	PASS
443	442	15	24	19.4	5970	PASS