

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040617\
 Data File : BF094218.D
 Acq On : 6 Apr 2017 10:20
 Operator : SJ/MA
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 06 13:24:49 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 06 13:20:29 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.569	0.570	-0.2	94	0.00
3	Pyridine	1.550	1.553	-0.2	92	0.00
4	n-Nitrosodimethylamine	0.638	0.629	1.4	90	0.00
5 S	2-Fluorophenol	1.184	1.183	0.1	95	0.00
6	Aniline	2.038	1.900	6.8	86	0.00
7 S	Phenol-d6	1.465	1.454	0.8	94	0.00
8	2-Chlorophenol	1.302	1.337	-2.7	96	0.00
9	Benzaldehyde	0.989	0.888	10.2	85	0.00
10 C	Phenol	1.667	1.632	2.1	89	0.00
11	bis(2-Chloroethyl)ether	1.303	1.315	-0.9	95	0.00
12	1,3-Dichlorobenzene	1.460	1.492	-2.2	96	0.00
13 C	1,4-Dichlorobenzene	1.479	1.520	-2.8	97	0.00
14	1,2-Dichlorobenzene	1.362	1.405	-3.2	98	0.00
15	Benzyl Alcohol	1.072	1.038	3.2	89	0.00
16	2,2'-oxybis(1-Chloropropane	2.007	1.879	6.4	87	0.00
17	2-Methylphenol	1.115	1.115	0.0	94	0.00
18	Hexachloroethane	0.520	0.537	-3.3	95	0.00
19 P	n-Nitroso-di-n-propylamine	0.941	0.920	2.2	92	0.00
20	3+4-Methylphenols	1.350	1.419	-5.1	101	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
22	Acetophenone	0.446	0.467	-4.7	102	0.00
23 S	Nitrobenzene-d5	0.350	0.356	-1.7	95	0.00
24	Nitrobenzene	0.346	0.346	0.0	93	0.00
25	Isophorone	0.616	0.613	0.5	94	0.00
26 C	2-Nitrophenol	0.165	0.182	-10.3	98	0.00
27	2,4-Dimethylphenol	0.284	0.288	-1.4	95	0.00
28	bis(2-Chloroethoxy)methane	0.406	0.397	2.2	92	0.00
29 C	2,4-Dichlorophenol	0.266	0.271	-1.9	95	0.00
30	1,2,4-Trichlorobenzene	0.274	0.277	-1.1	95	0.00
31	Naphthalene	0.962	0.966	-0.4	96	0.00
32	Benzoic acid	0.189	0.166	12.2	73	0.00
33	4-Chloroaniline	0.390	0.394	-1.0	95	0.00
34 C	Hexachlorobutadiene	0.140	0.142	-1.4	96	0.00
35	Caprolactam	0.085	0.086	-1.2	93	0.00
36 C	4-Chloro-3-methylphenol	0.277	0.285	-2.9	96	0.00
37	2-Methylnaphthalene	0.641	0.637	0.6	94	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
39	1,2,4,5-Tetrachlorobenzene	0.547	0.573	-4.8	99	0.00
40 P	Hexachlorocyclopentadiene	0.256	0.251	2.0	85	0.00
41 S	2,4,6-Tribromophenol	0.158	0.170	-7.6	99	0.00
42 C	2,4,6-Trichlorophenol	0.387	0.405	-4.7	97	0.00
43	2,4,5-Trichlorophenol	0.419	0.437	-4.3	93	0.00
44 S	2-Fluorobiphenyl	1.311	1.351	-3.1	97	0.00

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040617\
 Data File : BF094218.D
 Acq On : 6 Apr 2017 10:20
 Operator : SJ/MA
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 06 13:24:49 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 06 13:20:29 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.613	1.621	-0.5	94	0.00
46	2-Chloronaphthalene	1.263	1.282	-1.5	96	0.00
47	2-Nitroaniline	0.375	0.395	-5.3	95	0.00
48	Acenaphthylene	2.099	2.111	-0.6	94	0.00
49	Dimethylphthalate	1.422	1.453	-2.2	96	0.00
50	2,6-Dinitrotoluene	0.326	0.345	-5.8	95	0.00
51 C	Acenaphthene	1.225	1.217	0.7	94	0.00
52	3-Nitroaniline	0.383	0.408	-6.5	98	0.00
53 P	2,4-Dinitrophenol	0.155	0.177	-14.2	101	0.00
54	Dibenzofuran	1.667	1.620	2.8	90	0.00
55 P	4-Nitrophenol	0.300	0.296	1.3	87	0.00
56	2,4-Dinitrotoluene	0.409	0.409	0.0	89	0.00
57	Fluorene	1.382	1.357	1.8	98	0.00
58	2,3,4,6-Tetrachlorophenol	0.287	0.296	-3.1	94	0.00
59	Diethylphthalate	1.405	1.399	0.4	93	0.00
60	4-Chlorophenyl-phenylether	0.589	0.570	3.2	96	0.00
61	4-Nitroaniline	0.367	0.403	-9.8	99	0.00
62	Azobenzene	1.329	1.299	2.3	90	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
64	4,6-Dinitro-2-methylphenol	0.122	0.136	-11.5	99	0.00
65 c	n-Nitrosodiphenylamine	0.692	0.693	-0.1	96	0.00
66	4-Bromophenyl-phenylether	0.197	0.200	-1.5	96	0.00
67	Hexachlorobenzene	0.199	0.202	-1.5	96	0.00
68	Atrazine	0.207	0.212	-2.4	96	0.00
69 C	Pentachlorophenol	0.124	0.114	8.1	84	0.00
70	Phenanthrene	1.113	1.101	1.1	96	0.00
71	Anthracene	1.146	1.141	0.4	95	0.00
72	Carbazole	1.175	1.169	0.5	95	0.00
73	Di-n-butylphthalate	1.222	1.221	0.1	93	0.00
74 C	Fluoranthene	1.139	1.136	0.3	96	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
76	Benzidine	0.792	0.756	4.5	88	0.00
77	Pyrene	1.451	1.453	-0.1	95	0.00
78 S	Terphenyl-d14	0.892	0.887	0.6	96	0.00
79	Butylbenzylphthalate	0.618	0.649	-5.0	95	0.00
80	Benzo(a)anthracene	1.166	1.189	-2.0	96	0.00
81	3,3'-Dichlorobenzidine	0.417	0.446	-7.0	96	0.00
82	Chrysene	1.082	1.108	-2.4	97	0.00
83	Bis(2-ethylhexyl)phthalate	0.844	0.877	-3.9	94	0.00
84 c	Di-n-octyl phthalate	1.410	1.495	-6.0	95	0.00
85	Indeno(1,2,3-cd)pyrene	0.937	0.986	-5.2	103	0.00
86 I	Perylene-d12	1.000	1.000	0.0	103	0.00
87	Benzo(b)fluoranthene	1.226	1.303	-6.3	107	0.00

Data Path : Z:\HPCHEM1\BNA F\DATA\BF040617\
Data File : BF094218.D
Acq On : 6 Apr 2017 10:20
Operator : SJ/MA
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 06 13:24:49 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 06 13:20:29 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.199	1.127	6.0	94	0.00
89 C	Benzo(a)pyrene	1.129	1.142	-1.2	101	0.00
90	Dibenzo(a,h)anthracene	0.956	0.956	0.0	103	0.00
91	Benzo(a,h,i)perylene	0.964	0.958	0.6	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0